## BRAHMAPUTRA VALLEY FERTILIZER CORPORATION LIMITED

(A GOVT. OF INDIA UNDERTAKING) NAMRUP



Please reply to:

Chief Materials Manager BVFCL, Namrup P.O. Parbatpur – 786 623 Dist. Dibrugarh (Assam),India.

CIN No: **U24123AS2002GOI006786** GST No: 18AABCB9399R1ZK

Phone: +91 [374] 2507216; FAX: +91 [374] 2500204 / 2500317; e-mail: matmgt@bvfcl.co.in,www.bvfcl.com

To,

M/s.

NOTICE INVITING TENDER

1. NIT NO

: AN/2020/47-73/130/JB/Eng-02\_14

2. DATE OF ISSUE 3. TYPE OF BID

Single Stage 0020,

4. BID OPENING EARNEST MONEY DEPOSIT

BRAHMAPUTRA VALLEY FERTILIZER CORPORATION LIMITED, NAMRUP INVITES SEALED BID IN

SI.No.	DESCRIPTION	QUANTITY
1.	Ultrasonic Thickness Gauge as per specifications at Annexure-A	01 Na
	NB:	01 No.
	<ol> <li>"Make" of the item shall be specified in the offer.</li> </ol>	
	2. Confirmation of furnishing Test and Calibration Certificate along with supply.	

## SPECIAL TERMS & CONDITIONS:

- 1. The bid will include / indicate the following.
  - a) The complete scope of supply conforming to the specifications stipulated above and at Annexure- A and 1&II.
  - b) Basic Price, Packing & Forwarding charges, if any. The applicable GST and any other applicable charges.
  - c) The mode of transportation and charges thereon for delivery of the materials at Namrup, Assam.
  - d) Acceptance of terms of payment within 30 days as per Clause No.7 of Annexure I of the NIT.
  - e) Confirmation that the firm prices have been quoted.
  - f) Confirmation of the bid validity for 90 days from the bid opening date.
  - g) Please clearly indicate PAN/ GST No. in your quotation.
  - h) Acceptance of all other terms & conditions of Annexure I &II
- 2. The parties who are not registered as pre-qualified vendor of BVFCL will submit the documents pertaining to the eligibility criteria at point no. 19 of Annexure - I.
- MSEs shall be given the benefits of issue of tender documents free of cost, exemption from submission of Earnest money deposit, price preference, etc. as per Govt. Guidelines.
- SC/ST owned enterprises shall submit relevant SC/ST certificates and registration should remain valid at the time of bidding.
- Condition of prior turnover & prior experience with respect to start-up and micro & small enterprises will be relaxed as per Govt. Policy circular No. 1(2) (1)/2016-MA Dtd. 10/03/2016 & F.20/2/2014-PPD (Pt) Dtd. 20/09/16 subject to meeting of Quality & technical specifications
- 4. Goods bearing ISI mark will be given preference.
- 5. Materials shall be packed properly to avoid any damage during transit.
- 6. BVFCL is committed to a corruption free work environment. "All the above purchases, services and commitments of BVFCL will be honoured without the citizen having to pay any bribe". In case any person demands any bribe, it is the duty of the responsible to inform the matter to vigilance office, BVFCL, Namrup, P. O. Parbatpur, Dist. Dibrugarh, Assam PIN; 786623. (Tel no: 0374-2507092 / 0374-2507167).

Enclosures: 1. Annexure- A and I&II

Yours faithfully. For and on behalf of Brahmaputra Valley Fertilizer Corporation Ltd

Asstt. Materials Officer (P)

## Scope of Supply

Ultrasonic thickness gauge for pipe/plate thickness measurement as per the following specifications:

- 1. Measuring Range: preferably 1 mm to 100 mm. Display resolution: 0.01 mm.
- Key features: Should be able to cover wide range of thickness measurement for variety
  of materials -metals & plastics including steels, SS, Aluminum, Copper, steel Castings,
  hard plastics, etc.
- 3. Physical description: portable, hand held environmentally sealed case, water resistant key pad with backlight, light weight.
- 4. Display: with backlit, LCD/LED. •English / metric selectable. •mm displayed. •Display flashes to indicate Low battery / battery replacement.
- 5. Power supply: 'AA' alkaline battery or NiCad interchangeable. In case of rechargeable Battery, charger to be supplied.
- 6. Warranty: 1 year with post warranty support.
- Calibration: One step calibration for steels. Standard calibration block with calibration certificate.
- 8. Transducers with automatic probe recognition
- 9. Instruction Manual.
- 10. Quantity: 01 no

## NOTE:

- 1. Item shall be genuine, brand new and of reputed make.
- 2. Test and calibration certificate to be provided by the party.
- 3. Material should be packed properly to avoid damage during transit.