Results Editor

Raw Data

Name	Value	Description
Task Internal Id: T919		
Sample Scope: Sample scope: Scope 1/1, Sample 1/3 (FISHER (Ref. 25°C) | 4.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	11/7/2018 10:58:41 AM	
Sample start time	11/7/2018 10:58:41 AM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	11/7/2018 10:58:41 AM	
Method Function: Method function: Stir (Stir1)		
Creation time	11/7/2018 10:58:50 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	11/7/2018 10:59:51 AM	
T	18.0 °C	
E	161.7 mV	
tMe	00:30 min	
ESTMe	161.8 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 2/3 (FISHER (Ref. 25°C) | 7.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	11/7/2018 11:00:49 AM	
Sample start time	11/7/2018 11:00:49 AM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	11/7/2018 11:00:49 AM	
Method Function: Method function: Stir (Stir1)		
Creation time	11/7/2018 11:00:49 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	11/7/2018 11:01:50 AM	

Results Editor

LabX

Name		Value	Description
	T	18.0 °C	
	E	161.8 mV	
	tMe	00:30 min	
	ESTMe	161.8 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 3/3 (FISHER (Ref. 25°C) | 10.00)			
Method Function: Method function: Sample (Calib) (C1)			
	Creation time	11/7/2018 11:02:49 AM	
	Sample start time	11/7/2018 11:02:49 AM	
	Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)			
	Creation time	11/7/2018 11:02:49 AM	
Method Function: Method function: Stir (Stir1)			
	Creation time	11/7/2018 11:02:50 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)			
	Creation time	11/7/2018 11:03:50 AM	
	T	17.9 °C	
	E	161.8 mV	
	tMe	00:30 min	
	ESTMe	161.8 mV	
Sample Scope: After Sample Scopes			
Method Function: Method function: Calibration (Calibration1)			
	Creation time	11/7/2018 11:04:41 AM	
	Successful	Yes	
	Calib. temperature	18.0 °C	
	SLOPECal1	0.03 mV/pH	
	SLOPECal2	0.02 mV/pH	
	ZEROCal1	-4905.216 pH	
	ZEROCal2	-8880.833 pH	

Results Editor

Name		Value	Description
	Segment limit 1, 2	7.032 pH	
Method Function: Method function: Zero point 1 (Cal1.Z1)			
	Creation time	11/7/2018 11:04:41 AM	
	Result	-4905.216 pH	
	Decimal places	3	
Method Function: Method function: Slope 1 (Cal1.S1)			
	Creation time	11/7/2018 11:04:41 AM	
	Result	0.03 mV/pH	
	Decimal places	2	
Method Function: Method function: Zero point 2 (Cal1.Z2)			
	Creation time	11/7/2018 11:04:41 AM	
	Result	-8880.833 pH	
	Decimal places	3	
Method Function: Method function: Slope 2 (Cal1.S2)			
	Creation time	11/7/2018 11:04:41 AM	
	Result	0.02 mV/pH	
	Decimal places	2	