

Raw Data

Name	Value	Description
Task Internal Id: T1556		
Sample Scope: Sample scope: Scope 1/1, Sample 1/3 (FISHER (Ref. 25°C) 4.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	5/29/2019 9:00:34 AM	
Sample start time	5/29/2019 9:00:34 AM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	5/29/2019 9:00:35 AM	
Method Function: Method function: Stir (Stir1)		
Creation time	5/29/2019 9:00:44 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	5/29/2019 9:01:45 AM	
T	19.7 °C	
E	170.4 mV	
tMe	00:30 min	
ESTMe	170.4 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 2/3 (FISHER (Ref. 25°C) 7.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	5/29/2019 9:02:43 AM	
Sample start time	5/29/2019 9:02:43 AM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	5/29/2019 9:02:43 AM	
Method Function: Method function: Stir (Stir1)		
Creation time	5/29/2019 9:02:43 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	5/29/2019 9:03:43 AM	

Results Editor

LabX

Name	Value	Description
T	19.7 °C	
E	-2.5 mV	
tMe	00:30 min	
ESTMe	-2.2 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 3/3 (FISHER (Ref. 25°C) 10.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	5/29/2019 9:04:43 AM	
Sample start time	5/29/2019 9:04:43 AM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	5/29/2019 9:04:43 AM	
Method Function: Method function: Stir (Stir1)		
Creation time	5/29/2019 9:04:43 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	5/29/2019 9:05:44 AM	
T	19.7 °C	
E	-175.1 mV	
tMe	00:30 min	
ESTMe	-174.4 mV	
Sample Scope: After Sample Scopes		
Method Function: Method function: Calibration (Calibration1)		
Creation time	5/29/2019 9:06:34 AM	
Successful	Yes	
Calib. temperature	19.7 °C	
SLOPECal1	-57.21 mV/pH	
SLOPECal2	-56.74 mV/pH	
ZEROCal1	6.978 pH	
ZEROCal2	6.978 pH	

Results Editor

LabX

Name	Value	Description
Segment limit 1, 2	7.022 pH	
Method Function: Method function: Zero point 1 (Cal1.Z1)		
Creation time	5/29/2019 9:06:34 AM	
Result	6.978 pH	
Decimal places	3	
Method Function: Method function: Slope 1 (Cal1.S1)		
Creation time	5/29/2019 9:06:34 AM	
Result	-57.21 mV/pH	
Decimal places	2	
Method Function: Method function: Zero point 2 (Cal1.Z2)		
Creation time	5/29/2019 9:06:34 AM	
Result	6.978 pH	
Decimal places	3	
Method Function: Method function: Slope 2 (Cal1.S2)		
Creation time	5/29/2019 9:06:34 AM	
Result	-56.74 mV/pH	
Decimal places	2	