Results Editor

Raw Data

Name	Value	Description
Task Internal Id: T1119		
Sample Scope: Sample scope: Scope 1/1, Sample 1/3 (FISHER (Ref. 25°C) | 4.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	12/12/2018 2:59:56 PM	
Sample start time	12/12/2018 2:59:56 PM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	12/12/2018 2:59:56 PM	
Method Function: Method function: Stir (Stir1)		
Creation time	12/12/2018 3:00:06 PM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	12/12/2018 3:01:06 PM	
T	17.5 °C	
E	167.5 mV	
tMe	00:30 min	
ESTMe	166.8 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 2/3 (FISHER (Ref. 25°C) | 7.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	12/12/2018 3:02:04 PM	
Sample start time	12/12/2018 3:02:04 PM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	12/12/2018 3:02:04 PM	
Method Function: Method function: Stir (Stir1)		
Creation time	12/12/2018 3:02:04 PM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	12/12/2018 3:03:05 PM	

Results Editor

Name		Value	Description
	T	17.4 °C	
	E	6.5 mV	
	tMe	00:30 min	
	ESTMe	9.6 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 3/3 (FISHER (Ref. 25°C) | 10.00)			
Method Function: Method function: Sample (Calib) (C1)			
	Creation time	12/12/2018 3:04:04 PM	
	Sample start time	12/12/2018 3:04:04 PM	
	Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)			
	Creation time	12/12/2018 3:04:04 PM	
Method Function: Method function: Stir (Stir1)			
	Creation time	12/12/2018 3:04:05 PM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)			
	Creation time	12/12/2018 3:05:05 PM	
	T	17.4 °C	
	E	-159.0 mV	
	tMe	00:30 min	
	ESTMe	-154.8 mV	
Sample Scope: After Sample Scopes			
Method Function: Method function: Calibration (Calibration1)			
	Creation time	12/12/2018 3:05:56 PM	
	Successful	Yes	
	Calib. temperature	17.4 °C	
	SLOPECal1	-52.94 mV/pH	
	SLOPECal2	-54.17 mV/pH	
	ZEROCal1	7.158 pH	
	ZEROCal2	7.156 pH	

Results Editor

Name		Value	Description
	Segment limit 1, 2	7.036 pH	
Method Function: Method function: Zero point 1 (Cal1.Z1)			
	Creation time	12/12/2018 3:05:56 PM	
	Result	7.158 pH	
	Decimal places	3	
Method Function: Method function: Slope 1 (Cal1.S1)			
	Creation time	12/12/2018 3:05:56 PM	
	Result	-52.94 mV/pH	
	Decimal places	2	
Method Function: Method function: Zero point 2 (Cal1.Z2)			
	Creation time	12/12/2018 3:05:56 PM	
	Result	7.156 pH	
	Decimal places	3	
Method Function: Method function: Slope 2 (Cal1.S2)			
	Creation time	12/12/2018 3:05:56 PM	
	Result	-54.17 mV/pH	
	Decimal places	2	