

## Raw Data

Name	Value	Description
Task Internal Id: T1095		
Sample Scope: Sample scope: Scope 1/1, Sample 1/3 (FISHER (Ref. 25°C)   4.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	12/5/2018 11:30:04 AM	
Sample start time	12/5/2018 11:30:04 AM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	12/5/2018 11:30:05 AM	
Method Function: Method function: Stir (Stir1)		
Creation time	12/5/2018 11:30:14 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	12/5/2018 11:31:15 AM	
T	13.0 °C	
E	166.3 mV	
tMe	00:30 min	
ESTMe	166.1 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 2/3 (FISHER (Ref. 25°C)   7.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	12/5/2018 11:32:13 AM	
Sample start time	12/5/2018 11:32:13 AM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	12/5/2018 11:32:13 AM	
Method Function: Method function: Stir (Stir1)		
Creation time	12/5/2018 11:32:13 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	12/5/2018 11:33:13 AM	

# Results Editor

LabX

Name	Value	Description
T	13.3 °C	
E	0.1 mV	
tMe	00:30 min	
ESTMe	2.1 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 3/3 (FISHER (Ref. 25°C)   10.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	12/5/2018 11:34:13 AM	
Sample start time	12/5/2018 11:34:13 AM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	12/5/2018 11:34:13 AM	
Method Function: Method function: Stir (Stir1)		
Creation time	12/5/2018 11:34:13 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	12/5/2018 11:35:13 AM	
T	13.4 °C	
E	-167.1 mV	
tMe	00:30 min	
ESTMe	-163.5 mV	
Sample Scope: After Sample Scopes		
Method Function: Method function: Calibration (Calibration1)		
Creation time	12/5/2018 11:36:04 AM	
Successful	Yes	
Calib. temperature	13.2 °C	
SLOPECal1	-54.31 mV/pH	
SLOPECal2	-54.16 mV/pH	
ZEROCal1	7.059 pH	
ZEROCal2	7.059 pH	

# Results Editor

LabX

Name	Value	Description
Segment limit 1, 2	7.057 pH	
Method Function: Method function: Zero point 1 (Cal1.Z1)		
Creation time	12/5/2018 11:36:04 AM	
Result	7.059 pH	
Decimal places	3	
Method Function: Method function: Slope 1 (Cal1.S1)		
Creation time	12/5/2018 11:36:04 AM	
Result	-54.31 mV/pH	
Decimal places	2	
Method Function: Method function: Zero point 2 (Cal1.Z2)		
Creation time	12/5/2018 11:36:04 AM	
Result	7.059 pH	
Decimal places	3	
Method Function: Method function: Slope 2 (Cal1.S2)		
Creation time	12/5/2018 11:36:04 AM	
Result	-54.16 mV/pH	
Decimal places	2	