Results Editor

Raw Data

Name	Value	Description
Task Internal Id: T1165		
Sample Scope: Sample scope: Scope 1/1, Sample 1/3 (FISHER (Ref. 25°C) | 4.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	1/3/2019 10:32:19 AM	
Sample start time	1/3/2019 10:32:19 AM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	1/3/2019 10:32:19 AM	
Method Function: Method function: Stir (Stir1)		
Creation time	1/3/2019 10:32:29 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	1/3/2019 10:33:29 AM	
T	17.1 °C	
E	165.6 mV	
tMe	00:30 min	
ESTMe	165.5 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 2/3 (FISHER (Ref. 25°C) | 7.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	1/3/2019 10:34:27 AM	
Sample start time	1/3/2019 10:34:27 AM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	1/3/2019 10:34:28 AM	
Method Function: Method function: Stir (Stir1)		
Creation time	1/3/2019 10:34:28 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	1/3/2019 10:35:28 AM	

Results Editor

Name		Value	Description
	T	17.1 °C	
	E	-6.3 mV	
	tMe	00:30 min	
	ESTMe	-6.4 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 3/3 (FISHER (Ref. 25°C) | 10.00)			
Method Function: Method function: Sample (Calib) (C1)			
	Creation time	1/3/2019 10:36:28 AM	
	Sample start time	1/3/2019 10:36:28 AM	
	Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)			
	Creation time	1/3/2019 10:36:28 AM	
Method Function: Method function: Stir (Stir1)			
	Creation time	1/3/2019 10:36:28 AM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)			
	Creation time	1/3/2019 10:37:28 AM	
	T	17.2 °C	
	E	-178.7 mV	
	tMe	00:30 min	
	ESTMe	-178.8 mV	
Sample Scope: After Sample Scopes			
Method Function: Method function: Calibration (Calibration1)			
	Creation time	1/3/2019 10:38:19 AM	
	Successful	Yes	
	Calib. temperature	17.1 °C	
	SLOPECal1	-56.49 mV/pH	
	SLOPECal2	-56.40 mV/pH	
	ZEROCal1	6.926 pH	
	ZEROCal2	6.926 pH	

Results Editor

Name		Value	Description
	Segment limit 1, 2	7.037 pH	
Method Function: Method function: Zero point 1 (Cal1.Z1)			
	Creation time	1/3/2019 10:38:19 AM	
	Result	6.926 pH	
	Decimal places	3	
Method Function: Method function: Slope 1 (Cal1.S1)			
	Creation time	1/3/2019 10:38:19 AM	
	Result	-56.49 mV/pH	
	Decimal places	2	
Method Function: Method function: Zero point 2 (Cal1.Z2)			
	Creation time	1/3/2019 10:38:19 AM	
	Result	6.926 pH	
	Decimal places	3	
Method Function: Method function: Slope 2 (Cal1.S2)			
	Creation time	1/3/2019 10:38:19 AM	
	Result	-56.40 mV/pH	
	Decimal places	2	