Results Editor

Raw Data

Name	Value	Description
Task Internal Id: T1118		
Sample Scope: Sample scope: Scope 1/1, Sample 1/3 (FISHER (Ref. 25°C)	4.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	12/12/2018 1:59:36 PM	
Sample start time	12/12/2018 1:59:36 PM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	12/12/2018 1:59:36 PM	
Method Function: Method function: Stir (Stir1)		
Creation time	12/12/2018 1:59:45 PM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	12/12/2018 2:00:46 PM	
T	17.4 °C	
E	167.1 mV	
tMe	00:30 min	
ESTMe	166.1 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 2/3 (FISHER (Ref. 25°C)	7.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	12/12/2018 2:01:44 PM	
Sample start time	12/12/2018 2:01:44 PM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	12/12/2018 2:01:44 PM	
Method Function: Method function: Stir (Stir1)		
Creation time	12/12/2018 2:01:44 PM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	12/12/2018 2:02:44 PM	

Results Editor

Name		Value	Description
	T	17.3 °C	
	E	8.4 mV	
	tMe	00:30 min	
	ESTMe	12.6 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 3/3 (FISHER (Ref. 25°C)	10.00)			
Method Function: Method function: Sample (Calib) (C1)			
	Creation time	12/12/2018 2:03:44 PM	
	Sample start time	12/12/2018 2:03:44 PM	
	Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)			
	Creation time	12/12/2018 2:03:44 PM	
Method Function: Method function: Stir (Stir1)			
	Creation time	12/12/2018 2:03:44 PM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)			
	Creation time	12/12/2018 2:04:45 PM	
	T	17.3 °C	
	E	-159.6 mV	
	tMe	00:30 min	
	ESTMe	-154.9 mV	
Sample Scope: After Sample Scopes			
Method Function: Method function: Calibration (Calibration1)			
	Creation time	12/12/2018 2:05:35 PM	
	Successful	Yes	
	Calib. temperature	17.3 °C	
	SLOPECal1	-52.17 mV/pH	
	SLOPECal2	-54.98 mV/pH	
	ZEROCal1	7.197 pH	
	ZEROCal2	7.189 pH	

Results Editor

Name		Value	Description
	Segment limit 1, 2	7.036 pH	
Method Function: Method function: Zero point 1 (Cal1.Z1)			
	Creation time	12/12/2018 2:05:35 PM	
	Result	7.197 pH	
	Decimal places	3	
Method Function: Method function: Slope 1 (Cal1.S1)			
	Creation time	12/12/2018 2:05:35 PM	
	Result	-52.17 mV/pH	
	Decimal places	2	
Method Function: Method function: Zero point 2 (Cal1.Z2)			
	Creation time	12/12/2018 2:05:35 PM	
	Result	7.189 pH	
	Decimal places	3	
Method Function: Method function: Slope 2 (Cal1.S2)			
	Creation time	12/12/2018 2:05:35 PM	
	Result	-54.98 mV/pH	
	Decimal places	2	