Results Editor

Raw Data

Name	Value	Description
Task Internal Id: T1955		
Sample Scope: Sample scope: Scope 1/1, Sample 1/3 (FISHER (Ref. 25°C) | 4.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	9/19/2019 1:13:09 PM	
Sample start time	9/19/2019 1:13:09 PM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	9/19/2019 1:13:09 PM	
Method Function: Method function: Stir (Stir1)		
Creation time	9/19/2019 1:13:19 PM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	9/19/2019 1:14:19 PM	
T	19.8 °C	
E	172.3 mV	
tMe	00:30 min	
ESTMe	172.2 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 2/3 (FISHER (Ref. 25°C) | 7.00)		
Method Function: Method function: Sample (Calib) (C1)		
Creation time	9/19/2019 1:15:18 PM	
Sample start time	9/19/2019 1:15:18 PM	
Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)		
Creation time	9/19/2019 1:15:18 PM	
Method Function: Method function: Stir (Stir1)		
Creation time	9/19/2019 1:15:18 PM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)		
Creation time	9/19/2019 1:16:18 PM	

Results Editor

Name		Value	Description
	T	19.6 °C	
	E	2.9 mV	
	tMe	00:30 min	
	ESTMe	4.2 mV	
Sample Scope: Sample scope: Scope 1/1, Sample 3/3 (FISHER (Ref. 25°C)	10.00)			
Method Function: Method function: Sample (Calib) (C1)			
	Creation time	9/19/2019 1:17:18 PM	
	Sample start time	9/19/2019 1:17:18 PM	
	Temperature	25.0 °C	
Method Function: Method function: Titration stand (Titrationstand1)			
	Creation time	9/19/2019 1:17:18 PM	
Method Function: Method function: Stir (Stir1)			
	Creation time	9/19/2019 1:17:18 PM	
Method Function: Method function: Measure (normal) [1] (Measurenormal1)			
	Creation time	9/19/2019 1:18:18 PM	
	T	19.7 °C	
	E	-161.4 mV	
	tMe	00:30 min	
	ESTMe	-157.9 mV	
Sample Scope: After Sample Scopes			
Method Function: Method function: Calibration (Calibration1)			
	Creation time	9/19/2019 1:19:09 PM	
	Successful	Yes	
	Calib. temperature	19.7 °C	
	SLOPECal1	-56.02 mV/pH	
	SLOPECal2	-54.03 mV/pH	
	ZEROCal1	7.074 pH	
	ZEROCal2	7.076 pH	

Results Editor

Name		Value	Description
	Segment limit 1, 2	7.022 pH	
Method Function: Method function: Zero point 1 (Cal1.Z1)			
	Creation time	9/19/2019 1:19:09 PM	
	Result	7.074 pH	
	Decimal places	3	
Method Function: Method function: Slope 1 (Cal1.S1)			
	Creation time	9/19/2019 1:19:09 PM	
	Result	-56.02 mV/pH	
	Decimal places	2	
Method Function: Method function: Zero point 2 (Cal1.Z2)			
	Creation time	9/19/2019 1:19:09 PM	
	Result	7.076 pH	
	Decimal places	3	
Method Function: Method function: Slope 2 (Cal1.S2)			
	Creation time	9/19/2019 1:19:09 PM	
	Result	-54.03 mV/pH	
	Decimal places	2	