

TRAINING EXTRACT - paraphrased from Baxter manufacturer notice FA-2025-039, dated 2025-08-29. This is a historical demonstration, not a current safety notice. Consult the full original notice for every affected product and instruction. Only two catalog/lot pairs are included here.

Manufacturer: Baxter

Product: CLEARLINK IV extension and solution sets

Catalog: 2C8632; Lots: R25C31031

Catalog: 2R8858; Lots: R25A13024

Reason (paraphrase): A potential leak can compromise delivery and introduce risks during use.

Manufacturer wording: "Immediately locate, isolate, and cease all use of the affected product."

Response (paraphrase): Identify stock, follow the supplier-specific return or replacement route, and acknowledge the notice. Direct-purchase and distributor-purchase response routes differ.

Source: <https://www.medline.com/media/assets/pdf/vendor-list/FA-2025-039-Customer-Letter-Final-Combined.pdf>