



उत्पाद मैनुअल
एकल-उपयोग रोगाणुमुक्त रबड़ सर्जिकल
दस्ताने — विशिष्ट
आईएस 13422: 2024 / आईएसओ 10282: 2023 के अनुसार

PRODUCT MANUAL FOR
Single-Use Sterile Rubber Surgical
Gloves — Specification
as per IS 13422: 2024 / ISO 10282: 2023

विभिन्न उत्पादों के लिए भारतीय मानक ब्यूरो) अनुरूपता मूल्यांकन (विनियम, 2018 की योजना -I के तहत प्रमाणन के संचालन में एकरूपता और पारदर्शिता के लिए इस उत्पाद मैनुअल का उपयोग सभी क्षेत्रीय / शाखा कार्यालयों और लाइसेंसधारियों द्वारा संदर्भ सामग्री के रूप में किया जाएगा। दस्तावेज़ का उपयोग बीआईएस प्रमाणन प्राप्त करने के इच्छुक संभावित आवेदकों द्वारा भी किया जा सकता है।

This Product Manual shall be used as reference material by all Regional/Branch Offices & licensees to ensure uniformity of practice and transparency in operation of certification under Scheme-I of Bureau of Indian Standards (Conformity Assessment) Regulations, 2018 for various products. The document may also be used by prospective applicants desirous of obtaining BIS certification.

1.	मानक संख्या IS No.	:	IS 13422: 2024 / ISO 10282: 2023
	शीर्षक Title	:	Single-Use Sterile Rubber Surgical Gloves — Specification
	संशोधनों की संख्या No. of amendments	:	None
2.	नमूना दिशानिर्देश Sampling Guidelines		
a)	कच्चा माल Raw material	:	To be declared by the manufacturer as per Cl. 5.1 of IS 13422
b)	समूहीकरण दिशानिर्देश Grouping Guidelines	:	Please refer ANNEX – A
c)	नमूने का परिमाण Sample Quantity	:	100 Nos
d)	परीक्षण अनुरोध में घोषित किए जाने वाले पैरामीटर Parameters to be	:	As per Scope of the license (See Sl. No. 6)

	Declared in Test Request		
3.	परीक्षण उपकरणों की सूची List of Test Equipment	:	Please refer ANNEX – B
4.	निरीक्षण और परीक्षण की स्कीम Scheme of Inspection and Testing	:	Please refer ANNEX – C
5.	एक दिन में संभावित परीक्षण Possible tests in a day	:	All tests
6.	लाइसेंस का दायरा /Scope of the Licence:		
“License is granted to use Standard Mark as per <i>IS 13422: 2024 / ISO 10282: 2023</i> with the following scope:			
Name of the product		Single-Use Sterile Rubber Surgical Gloves — Specification	
Material		To be declared by the manufacturer as per Cl. 5.1 of IS 13422	
Type		Type – 1 / Type – 2	
Design		Gloves with straight fingers / Gloves with fingers curved in the palmar direction	
Cuff Termination		Cut / Rolled Rim	
Finish		Textured surface over part or the entire glove / Smooth surface / Powdered surface / Powder-free surface	
Sizes		5 / 5.5 / 6 / 6.5 / 7 / 7.5 / 8 / 8.5 / 9 / 9.5	

ANNEX -A**Grouping Guidelines**

1. Single-Use Sterile Rubber Surgical Gloves as per IS 13422 are classified based on the following:

Material	To be declared by the manufacturer as per Cl. 5.1 of IS 13422
Type	Type – 1 / Type – 2
Design	Gloves with straight fingers / Gloves with fingers curved in the palmar direction
Cuff Termination	Cut / Rolled Rim
Finish	Textured surface over part or the entire glove / Smooth surface / Powdered surface / Powder-free surface
Sizes	5 / 5.5 / 6 / 6.5 / 7 / 7.5 / 8 / 8.5 / 9 / 9.5

2. For considering GoL/ CSoL, Single-Use Sterile Rubber Surgical Gloves of lowest and highest size from each type of raw material, type of gloves, design, cuff termination and finish shall be tested to cover all other sizes of Single-Use Sterile Rubber Surgical Gloves in the scope of licence.
3. Firm shall declare the varieties of Single-Use Sterile Rubber Surgical Gloves that they intend to be covered in the Licence. The Scope of Licence may be restricted based on the Manufacturing and Testing capabilities of the Manufacturer.
4. During the operation of the Licence, BO shall ensure that all the gloves covered in the Scope of the Licence are tested in rotation to the extent possible.

Note: The firm shall declare the following:

- a) Type of surface treatment it intends to carry out
- b) Process of Sterility

ANNEX- B
LIST OF TEST EQUIPMENTS
 (INDICATIVE LIST, FOR GUIDANCE ONLY)

Sl. No.	Tests used in with Clause Reference	Test Equipment
1.	Dimensions as per Cl. 7.1 and Table 2	Thickness gauge Scale
2.	Watertightness as per Cl. 7.2	Test set up as per Annex B Volume / flow meter Stop watch Temperature Sensor
3.	Tensile Strength and Elongation at Break as per Cl. 7.3.2 and Cl. 7.3.3	Thickness gauge Scale Universal Testing Machine Dumb bell Die cutter
4.	Accelerated Ageing	Hot Air Oven with timer

ANNEX C
SCHEME OF INSPECTION AND TESTING

1. QUALITY ASSURANCE PLAN

- 1.1. It is expected that manufacturers (licensees/applicants) will implement a Quality Assurance Plan i.e. a plan of regular testing and in-process controls, designed to ensure that the product bearing the Standard Mark conforms to all requirements of the Indian Standard.
- 1.2 The manufacturers shall define a Quality Assurance Plan defining the control unit (i.e. lot/batch etc.) and the levels of control (i.e. the frequency and number of samples for conducting the different tests as per the Indian Standard) and submit the same to BIS Branch Office for information. The manufacturer shall comply with the same and maintain test records in accordance with para 2.4.

1.3 RECOMMENDED LEVELS OF CONTROL/CONTROL UNIT:

- 1.3.1 For the guidance of manufacturers, the recommended definition of control unit is: *Single-Use Sterile Rubber Surgical Gloves of same type of raw material, type of gloves, design, cuff termination and finish manufactured in a day under similar conditions* shall constitute a control unit.
- 1.3.2 For the guidance of manufacturers in preparing the Quality Assurance Plan, recommended levels of control are given in Table 1.
- 1.3.3 The manufacturer shall ensure inspection and testing as per the Quality Assurance Plan submitted by them on the whole production of the factory which is covered by this plan. Alternatively, the manufacturer has the option of adherence to the quality plan as per levels of control recommended in column 3 of Table 1.
- 1.4 However, all manufacturers shall ensure compliance of their products to all the requirements of the Indian Standard.

2. **ENSURING COMPLIANCE THROUGH TESTING-** It is expected that manufacturers (licensees/applicants) will establish a suitably equipped and staffed in house laboratory (In house testing facility) for testing at least those parameters of the Indian Standard which require routine testing for ensuring quality of the product. This includes in-process controls as may be defined and put in place by the manufacturer and testing parameters/requirements which can only be performed in the factory.

- 2.1 For the guidance of manufacturers, Table 1 giving the recommended levels of control is given below. Column 2 of Table 1 indicates routine tests where test equipment is required in house as “R” or other tests which can be subcontracted as “S”. Subcontracting is permitted to BIS recognized/empanelled laboratory or any other laboratory having valid NABL accreditation as per IS/ISO/IEC 17025.

- 2.2 **For MSME manufacturers, the requirement of maintaining a laboratory/in-house testing facility for routine tests (indicated as “R” in Column 2 of Table 1) is also optional.**

- 2.2.1 MSME manufacturers may utilize common cluster based facilities as per guidelines for the utilization of cluster based test facilities by MSMEs or the provisions of Sharing of testing facilities or get testing done

from BIS recognized/empaneled laboratory or any other laboratory having valid NABL accreditation as per IS/ISO/IEC 17025.

2.3 Large Scale manufacturers shall maintain an in-house laboratory equipped at least with test facilities for routine tests (indicated as “R” in Column 2 of Table 1), where different tests given in the specification shall be carried out in accordance with the method given in the specification. They shall also implement a calibration plan for the in-house test equipment.

2.3.1 Alternatively, in lieu of an in-house laboratory, large scale manufacturers can also utilize the provisions of Sharing of testing facilities as per the Guidelines for Grant of Licence available on BIS website www.bis.gov.in. (Under Conformity Assessment>Product Certification Process). Even for subcontracted tests, provisions for sharing of testing facilities can be utilized.

2.4 TEST RECORDS- The manufacturers maintaining an in-house laboratory or utilizing common cluster based facilities or shared test facilities shall maintain test records for the tests carried out to establish conformity. For the tests being subcontracted to BIS recognized/empanelled laboratory or any other laboratory having valid NABL accreditation as per IS/ISO/IEC 17025, test reports issued by the laboratories shall be available for inspection by BIS.

3. PACKING AND MARKING - The Standard Mark as given in the Schedule of the licence shall be incorporated legibly and indelibly on the packing of Single-Use Sterile Rubber Surgical Gloves provided always that the gloves so marked conforms to each requirement of the specification.

3.1. Marking on Single-Use Sterile Rubber Surgical Gloves shall be done as per designation specified at Cl. 9 of IS 13422: 2024 / ISO 10282: 2023.

3.2 Additional Marking requirements: The packaging of Single-Use Sterile Rubber Surgical Gloves shall also be marked with the following additional requirement on each packing:

a) “For BIS certification details please visit www.bis.gov.in”

3.3 All the production which conforms to the Indian Standard and covered under the scope of this licence shall be marked with the Standard Mark.

4. HYGIENIC CONDITIONS (if applicable) – NA.

5. REJECTION - Disposal of non-conforming product shall be done in such a way so as to ensure that there is no violation of provisions of BIS Act, 2016.

TABLE-1

(1)				(2)	(3)		
Test Details				Test equipment requirement R: required (or) S:Sub-contracting permitted	Levels of Control		
Cl.	Requirement	Test Methods			No. of Sample	Frequency	Remarks
		Clause	Reference				
5	Material	5.1 to 5.3	IS 13422	S	One	Each Consignment of raw material	No further testing is required if accompanied with TC or ISI marked
5.4	Bio-compatibility		ISO 10993	S	One	Once in a month	
7.1 Table 2	Dimensions	Annex A	IS 13422	R	Firm to have adequate in-process controls to check compliance of this parameter as per the tolerances given in the Indian Standard. However, appropriate records shall be maintained by the manufacturer for evidence of conformity		
7.2	Watertightness	Annex B	IS 13422	R	One	Each Control Unit	
7.3	Tensile Properties						
7.3.2 Table 3	<i>Force at break and elongation at break before accelerated ageing</i>		ISO 37 IS 3400 (Part 1)	R	One	Each Control Unit	
	Min Force at break						
	Min Elongation at break						
	Maximum force required to produce 300 % elongation						
	<i>Force at break and elongation at break after accelerated ageing</i>		ISO 37	R	One	Once in a month	
Min Force at break							
	Min Elongation at break						