

Hoffmann Messtechnik GmbH

Optical gas analysis and spectrometer service . Heidelberg, Germany . service@hoffmann-mt.de

Calibration certificate

Certificate number: HMT-CAL-2025-11842
Instrument: DOAS UV spectrometer, scanning
Serial number: DOAS-EM27-2208
Customer: Martha Glass, Íslenska jarðeðlisfræðistofnunin (ÍJE)
Calibration date: 18 November 2025
Next calibration due: 18 November 2026
Procedure: ISO 17025 traceable, SO2 reference cells

Parameter	Specification	Measured	Result
Wavelength range	290 to 420 nm	290 to 420 nm	Pass
Wavelength accuracy	+/- 0,10 nm	0,04 nm	Pass
Spectral resolution	0,55 nm	0,52 nm	Pass
SO2 cell 1 (98 ppm m)	+/- 5%	+2,1%	Pass
SO2 cell 2 (485 ppm m)	+/- 5%	-1,4%	Pass
Dark current stability	stable	stable	Pass
Detector linearity	R2 > 0,999	0,9997	Pass

The instrument meets all manufacturer specifications and is suitable for field measurement of volcanic sulphur dioxide flux. Reference cells are traceable to national standards. Calibration performed at 21 C and 45% relative humidity.

Calibrated by: T. Hoffmann, senior optical engineer. This certificate is valid for the instrument identified above only.