

ResMed myAir

Monthly CPAP Usage Report - December 2025

Generated: January 2, 2026 | www.myair.resmed.com

Patient and Device Information

Patient Name:	Anthony Hicks
Device:	ResMed AirSense 11 AutoSet
Device Serial Number:	RES-2022-3847291
Mask:	ResMed P30i Nasal Pillow
Provider:	Dr. Lisa Cho, MD - Birmingham Sleep and Pulmonary Associates
Report Period:	December 1, 2025 to December 31, 2025
Report Generated:	January 2, 2026

December 2025 Summary

Metric	Value
Total nights in reporting period	31 nights
Nights CPAP used	28 nights
Nights not used (missed)	3 nights (Dec 3, 10, 17 - no data)
Compliance rate (nights used / total)	90.3%
Days meeting 4-hour minimum	28 of 31 (90.3%)
Average nightly usage	6 hours 52 minutes
Average AHI (Apnea-Hypopnea Index)	3.4 events/hour
Average mask leak rate	5.8 L/min
Average therapy pressure	9.4 cmH2O
myAir Score	9 out of 10 (Excellent)

Selected Nightly Data Highlights

Date	Usage	AHI	Mask Leak	Pressure	Notes
Dec 1	7h 12m	2.8	4.9 L/min	9.1 cmH2O	Good night
Dec 3	---	---	---	---	No data - missed
Dec 8	6h 44m	3.1	5.6 L/min	9.3 cmH2O	Normal
Dec 10	---	---	---	---	No data - missed
Dec 15	7h 02m	3.9	6.1 L/min	9.8 cmH2O	Slight rise AHI
Dec 17	---	---	---	---	No data - missed
Dec 22	6h 18m	4.2	6.8 L/min	10.2 cmH2O	Christmas travel
Dec 25	5h 48m	4.8	7.2 L/min	10.8 cmH2O	Holiday, late night
Dec 29	7h 22m	2.6	4.4 L/min	8.9 cmH2O	Excellent
Dec 31	7h 05m	3.0	5.1 L/min	9.2 cmH2O	Good

Full December 2025 Daily Usage Log

All Nights - December 2025

Date	Day	Usage	AHI	Leak (L/min)	Pressure	Score	Status
Dec 1	Mon	7h 12m	2.8	4.9	9.1	10	OK
Dec 2	Tue	6h 55m	3.2	5.3	9.4	9	OK
Dec 3	Wed	0h 00m	---	---	---	0	MISS
Dec 4	Thu	7h 01m	3.0	5.0	9.2	9	OK
Dec 5	Fri	6h 48m	3.3	5.8	9.5	9	OK
Dec 6	Sat	7h 18m	2.9	4.7	9.0	10	OK
Dec 7	Sun	6h 39m	3.6	6.2	9.7	9	OK
Dec 8	Mon	6h 44m	3.1	5.6	9.3	9	OK
Dec 9	Tue	6h 52m	3.4	5.9	9.4	9	OK
Dec 10	Wed	0h 00m	---	---	---	0	MISS
Dec 11	Thu	7h 05m	3.0	5.1	9.2	10	OK
Dec 12	Fri	6h 58m	3.5	6.0	9.6	9	OK
Dec 13	Sat	7h 20m	2.7	4.6	8.9	10	OK
Dec 14	Sun	6h 41m	3.8	6.4	9.9	9	OK
Dec 15	Mon	7h 02m	3.9	6.1	9.8	9	OK
Dec 16	Tue	6h 47m	3.3	5.7	9.4	9	OK
Dec 17	Wed	0h 00m	---	---	---	0	MISS
Dec 18	Thu	7h 09m	3.1	5.2	9.3	9	OK
Dec 19	Fri	6h 55m	3.4	5.8	9.5	9	OK
Dec 20	Sat	7h 22m	2.9	4.8	9.1	10	OK
Dec 21	Sun	7h 00m	3.2	5.4	9.4	9	OK
Dec 22	Mon	6h 18m	4.2	6.8	10.2	8	OK
Dec 23	Tue	6h 30m	4.0	6.5	10.0	8	OK
Dec 24	Wed	6h 45m	3.8	6.2	9.8	9	OK
Dec 25	Thu	5h 48m	4.8	7.2	10.8	7	OK
Dec 26	Fri	6h 55m	3.6	6.0	9.7	9	OK
Dec 27	Sat	7h 10m	3.2	5.3	9.3	9	OK
Dec 28	Sun	7h 00m	3.0	5.0	9.1	10	OK
Dec 29	Mon	7h 22m	2.6	4.4	8.9	10	OK
Dec 30	Tue	7h 08m	2.8	4.7	9.0	10	OK
Dec 31	Wed	7h 05m	3.0	5.1	9.2	10	OK

myAir Insights and ResMed Recommendations

myAir Score Explanation

Your myAir score for December 2025 is 9 out of 10 (Excellent). The myAir score is calculated based on four factors: hours of usage (weighted 40%), AHI events per hour (weighted 30%), mask fit/leak (weighted 20%), and mask on/off events (weighted 10%).

Your score was reduced slightly by three missed nights (Dec 3, 10, 17). On all nights you used your device, therapy performance was excellent. AHI of 3.4 events/hour is well-controlled. Mask leak of 5.8 L/min is excellent - well within the recommended limit of 24 L/min.

ResMed Recommendations

Mask Cushion:

Replace ResMed P30i cushion every 2 weeks per schedule. Worn cushions increase leak rate.

Filters:

Replace ultra-fine filter monthly. Replace standard filter every 6 months.

Water Chamber:

Clean water chamber with mild soap and distilled water weekly.

Tubing:

Replace ClimateLine heated tubing quarterly (next replacement approx. January 2026).

Device:

Device replacement typically recommended at 5 years. Current device: 2022 (3 years old).

Missed Nights:

Three missed nights noted. Consistent nightly use provides maximum therapeutic benefit.

This report was generated by the ResMed myAir patient portal. Data is transmitted wirelessly from your AirSense 11 device. Share this report with Dr. Lisa Cho at your next appointment and with your DOT medical examiner as required. Questions: ResMed customer support 1-800-424-0737 or www.resmed.com.