

METROVERIFY 360

Directorate of Legal Metrology • Government of India
Statutory NAWI Metrological Test Dossier (OIML R-76)



Application Reference: APP-2026-0840

Certificate: CERT-2026-917002

1. INSTRUMENT OWNER & CUSTODY DETAILS

Owner Name: Rajesh Sharma **Mobile / Email:** +91 98765 43210 | applicant@metroverify.gov.in
Organization: Apex Agro & Logistics Hub Ltd **Installation Site:** Island #3, Highway Retail Fuel Station, NH-48 Express Corridor

2. INSTRUMENT TECHNICAL SPECIFICATIONS

Instrument: Multi-Product Electronic Fuel Dispenser **Classification:** Fuel Dispenser (MPD)
Make & Model: Gilbarco Veeder-Root India / Horizon 4- **Serial Number:** GVR-MPD-2024-8871
Capacity: 50 L/min (Accuracy +/- 0.3%) **Accuracy Class:** Class III (Medium)

3. FIELD INSPECTION & METROLOGICAL TEST FINDINGS

Inspection Date: 19-Sept-2026 **Verifying Officer:** Inspector S. K. Verma (LMI-DL-2018-044)
Standard Weights: Conical Metal Prover Measures 5L & 10L (RRSL/FL/2026/012) **Zero Load Error:** 0.00 Litres at cut-off valve
Observed Reading: Delivery of 5.004 L on 5.000 L preset (+0.08%) **Max Permissible Error:** +/- 0.3% (+/- 15 mL on 5 Litres)
Repeatability: Satisfactory (Range < 1 e) **Statutory Seal Tag:** DL-LM-SEAL-88402

Final Metrological Result: PASS

Officer Remarks: Totalizer reading 1,421,908 Litres. Security seal wire intact. Delivery error +0.08% well within permissible limits.

4. STATUTORY VERIFICATION NOTICE

Notice: This NAWI report documents physical and accuracy test readings performed under Section 24 of The Legal Metrology Act, 2009. Automated generation of this test dossier records compliance measurements and does not replace the statutory verification seal plate affixed to the instrument receiver.