

Accelerated Magnetic Resonance Imaging by Adversarial Neural Network

Ohad Shitrit, Tammy Riklin Raviv

¹*Department of Electrical Engineering,* ²*The Zlotowski Center for Neuroscience Ben-Gurion University of the Negev, Israel*

Abstract

A main challenge in Magnetic Resonance Imaging (MRI) for clinical applications is speeding up scan time. Beyond the improvement of patient experience and the reduction of operational costs, faster scans are essential for time-sensitive imaging, where target movement is unavoidable, yet must be significantly lessened, e.g., fetal MRI, cardiac cine, and lungs imaging. Moreover, short scan time can enhance temporal resolution in dynamic scans, such as functional MRI or dynamic contrast enhanced MRI. Current imaging methods facilitate MRI acquisition at the price of lower spatial resolution and costly hardware solutions.

We introduce a practical, software-only framework, based on deep learning, for accelerating MRI scan time allows maintaining good quality imaging. This is accomplished by partial MRI sampling, while using an adversarial neural network to estimate the missing samples. The inter-play between the generator and the discriminator networks enables the introduction of an adversarial cost in addition to a fidelity loss used for optimizing the peak signal-to-noise ratio (PSNR). Promising image reconstruction results are obtained for 3T and 1.5T MRI where only 40%, 25% and 16.6% of the original data are used. Extended performance assessment was conducted in order to emphasize the advantage of proposed method.

Keywords: MRI, GAN, Accelerated, K-space

1. Introduction

Magnetic Resonance Imaging (MRI) is a non-ionizing imaging modality, and is therefore widely used in diagnostic medicine and biomedical research. The physical principles of MRI are based on a strong magnetic field and pulses of radio frequency (RF) electromagnetic

radiation. Images are produced when hydrogen atoms, which are prevalent in living organisms, emit the absorbed RF energy that is then received by antennas in close proximity to the anatomy being examined. Spatial localization of the detected MRI signals is obtained by varying the magnetic field gradients. The discretized RF output is presented in a Fourier space (called K-space), where the x-axis refers to the frequency and the y-axis to the phase. An inverse fast Fourier transform (IFFT) of the K-space is then used for generating anatomically meaningful MRI scans. Figure 1 presents K-space traversal patterns used in conventional imaging. Each row of the k-space is acquired after one RF excitation pulse. The number of rows multiplied by the number of slices (z-axis) determines the total scan time.

The duration of standard single structural MRI acquisition is approximately 5 minutes. Usually, several scans of different modalities or a sequence of scans are acquired such that the overall scan time is much longer. Lengthy imaging process reduces patient comfort and is more vulnerable to motion artifacts. In cases where motion is inevitable, e.g., fetal MRI, cardiac cine, and lungs imaging, scan time must be significantly shortened, otherwise the produced images might be useless. Moreover, in dynamic MRI sequences, acquisition must be brief such that the temporal resolution of the sequence would allow capturing significant temporal changes, e.g., instantaneous increment of the contrast-enhanced material concentration in DCE-MRI or differences in hemodynamic response expressed in fMRI [1].

A straight forward reduction of the scan time can be obtained by sampling fewer slices, thus reducing the spatial resolution in the z-axis. Spatial distances between adjacent slices of fetal MRI or fMRI, for example, are often as high as 0.5 centimeters. Therefore, a significant portion of the potential input is not conveyed through imaging. On the other hand, under-sampling in the x-y domain leads to aliasing, as predicted by the Nyquist sampling theorem.

Numerous research groups as well as leading MRI scanner manufacturers make significant efforts to accelerate the MRI acquisition process. Hardware solutions allow parallel imaging by using multiple coils [2] to sample k-space data. There exist two major approaches [3] that are currently implemented in commercial MRI machines. Both reconstruct an image from the under-sampled k-space data provided by each of the coils. The sensitivity encoder (SENSE) transforms the partial k-spaces into images, then merges the resulting aliased images into one coherent image [4]. The GeneRalized Autocalibrating Partial Parallel Acquisition (GRAPPA) techniques [5] operate on signal data within the complex frequency

domain before the IFFT.

The compressed sensing (CS) technique [6] allows efficient acquisition and reconstruction of a signal with fewer samples than the Nyquist-Shannon sampling theorem requires, if
40 the signal has sparse representation in a known transform domain. Using CS for MRI reconstruction by sampling a small subset of the k-space grid had been proposed in [7]. The underlying assumption is that the undersampling is random, such that the zero-filled Fourier reconstruction exhibits incoherent artifacts that behave similarly to additive random noise. This, however, makes the algorithm less sensitive to fine details and noise like texture,
45 results in smooth reconstruction with loss of diagnostically information, as we will see in section 3.

Recently machine learning techniques based on manifold learning [8, 9] and dictionary learning [10, 11] were suggested for MRI reconstruction. MRI reconstruction using convolutional neural networks (CNN) was introduced in [12]. The network learns the mapping
50 between zero-filled and fully-sampled MR images. In [13], residual network was proposed for MRI super-resolution. Their model is able to receive multiple inputs acquired from different viewing planes for better image reconstruction. In [14, 15] a multi-scale residual network, also known as U-net [16] is used and provides good reconstruction MR images. Deep convolutional networks have been successfully applied to acceleration of dynamic MRI in [17].
55 All these works address the reconstruction problem in the image domain rather than the k-space domain.

The proposed framework utilizes recent advances in deep learning, while similarly to the CS methods addresses MRI reconstruction directly from the k-space. Specifically, we use generative adversarial networks (GAN) [18, 19, 20]. GANs are based on the inter-
60 play between two networks: a generator and a discriminator. The generator is capable of learning the distribution over a data-base, and sample realizations of this distribution. The discriminator is trained to distinguish between ‘generated’ samples and real ones. This powerful combination has been used for MRI acceleration in [21, 22], and also for generating Computed Tomography (CT)-like images from MRIs [23]. Here, the generator is used for
65 reconstruction of the entire k-space grid from under-sampled data. Its loss is a combination of an adversarial loss, based on the discriminator output and a fidelity loss with respect to the fully sampled MRI. Promising results are obtained for brain MRI reconstruction using only 40%, 25% and 16.6% of the data.

The paper is organized as follows. Section 2 presents some theoretical foundation and
 70 our method. Section 3 describes the experimental results. Conclusions and future directions
 are describes in section 4.

2. Method

2.1. *K-space*

Let \mathbf{u} denote the desired signal, a 2D MR image, obtained by the IFFT of the complex
 k-space signal s_0 . Let M_F denote a full sampling mask such that the reconstructed MR
 image is:

$$\mathbf{u} = F^H M_F \odot s_0 \quad (1)$$

where H is the Hermitian transpose operation, \odot denotes element-wise multiplication, and
 75 F^H is an orthonormal 2D IFFT operator, such that $F^H F = I$. While sampling part of the k-
 space, using M_p as a sampling mask, the reconstructed MR image suffers from artifacts and
 aliasing. An example artifacts caused by under-sampling in the phase axis using Gaussian
 mask (40%) is shown in Figure 1.




Figure 1: Under-sampling artifacts: the arrows illustrate the sampling methodology

2.2. *Objective*

Let $s_p = M_p \odot s_0$ denote the under-sampled k-space. Given a sampling mask and a
 model f , defined by the set of parameters Θ , our goal is to estimate the missing k-space
 samples such that:

$$\Theta = \arg \min_{\Theta} L(F^H f(s_p; \Theta), \mathbf{u}) \quad (2)$$

80 where $L(\cdot)$ is the loss function. While choosing the loss to be L2 norm is reasonable for
 natural images, for the k-space, which has different spatial features, this may not be enough.

As mentioned in [20], L2 minimization provides a blurry solution, results in a loss of fine details. Averaging the high frequency details in the k-space domain results in very poor reconstruction. In order to address this problem, we used the adversarial loss, based on
85 GAN.

We trained our model using the adversarial strategy, as described in [18, 19]. This method is based on a generator G , which takes noise z with uniform distribution $p_u(z)$ as input and generates samples from the data distribution. A discriminator D is trained to distinguish between “real” examples from the data and generated (“fake”) examples from G .
90 During the training process, we optimize G to maximize the discriminator’s probability of error. Simultaneously, D is getting better and provides more accurate predictions.

Let s_0 denote a “real” k-space sample from the distribution $p_r(s_0)$. The following optimization process can be described by two-players min-max game:

$$\min_G \max_D \mathbb{E}_{s_0 \sim p_r(s_0)} \log [D(x)] + \mathbb{E}_{z \sim p_u(z)} \log [1 - D(G(z))] \quad (3)$$

In equilibrium, the generator G is able to generate samples that look like the real data. In
95 our case, G estimates the missing k-space samples from a linear combination of the sampled data and a uniform noise with distribution $p_u(z)$. An L2 fidelity constraint is added to the adverbial loss of the generator, as follows:

$$L_G = \alpha \cdot \mathbb{E}_{z \sim p_u(z)} \log [1 - D(F^{-1}(\hat{s}_0))] + \beta \cdot \|(1 - M_p) \odot (\hat{s}_0 - s_0)\|_2^2 \quad (4)$$

where \hat{s}_0 is the estimated k-space and $\alpha = 1$, $\beta = 1$ are hyperparameters tuned by a cross-validation process. The discriminator’s input is the reconstructed MR image, i.e., after
100 IFFT. By that, we are integrating the reconstruction phase in our optimization. In this work, we use the Wasserstein distance as the GAN loss function [24] which helps us to improve the training stability and provides better reconstruction than the classic GAN.

2.3. Network Architecture

The generator input is a two-channel image representing the real and the imaginary parts of the partially sampled k-space image, s_p . Each missing sample is initialized by uniform i.i.d. noise. The pixel (i, j) in the generator input image is:

$$G_{in}(i, j) = s_{p_{i,j}} + (1 - M_p)_{i,j} z_{i,j} \quad (5)$$

Due to the combination of the adversarial and the fidelity loss, G produces reasonable k-space samples from a given samples and noise distribution $p_u(z)$. In order to use the sampled data, s_p , and estimate only the missing samples we used a residual network [25] as used in [13], such that:

$$\hat{s}_0 = s_p + (1 - M_p) \odot G_{out} \quad (6)$$

where G_{out} is the generator output. Figure 2 describes our framework:




Figure 2: Framework architecture: G and D are the generator and discriminator networks, respectively. F is a 2D IFFT operator.

105 A common architecture is used for the discriminator, composed of convolutional layers, batch normalization, and leaky-ReLU as suggested in [19]. For the generator, we compose a dedicated architecture based on multi-channel input for representing the real and imaginary components. Both architectures are shown in Figure 3. The training methodology is doing k_g generator update steps for each discriminator single step.

110 3. Experiments

The training data consists of 500 3D brain MRI (T1) scans of different patients from the IXI dataset¹. The data has been acquired by three MR machines, Philips 1.5T, 3T and GE 1.5. All images padded to resolution of 256×256 pixels. From each 3D volume we extract 93 2D sagittal slices. We used 37.2k (80%) 2D slices for training and 9.2k (20%) for testing
 115 (100 3D volumes). In order to create k-space images for training, inverse orthonormal 2D FFT is applied to the fully-sampled MR images. We sample the k-space using 2D Gaussian mask with sampling factor of 2.5, 4 and 6. Data augmentation is created by random offsets

¹<http://brain-development.org/ixi-dataset/>




Figure 3: Networks architecture. The generator input is a two-channel signal, real and imaginary. For each layer, k is the kernel size and C is the number of output channels.

of the proposed mask and image flipping. This leads us to reconstruction of the MR image from 40%, 25% and 16.6% (Figure 4) of the original k-space data.

120 The generator is composed of 5 blocks of CONV-BatchNorm-ReLU, with output channels 16, 32, 64, 32, 8, respectively. The last layer is CONV with two outputs channels (for real and imaginary parts). The discriminator is composed of 4 blocks of CONV-Pool-BatchNorm-LReLU with output channels 8, 16, 32, 16 and one fully-connected layer. All CONV layers kernel size is 3×3 . All weights was initialized by Xavier [26]. We used RMSprop solver

125 with fixed learning rate of $5e-6$ and set k_d to 1.

We compare the proposed method to reconstruction results obtained by using a conventional compressed sensing method CS-MRI [7] and Zero-filling. In addition, we optimized two networks: 1. CNN-L2 - a k-space generator (with same architecture as the proposed method) using only L2 loss. 2. IM-CNN-L2 - a network which performs on the image-space

130 and optimized to remove the artifacts caused by under-sampling and zero padding. We used the U-net [16] architecture and the same reconstruction method as presented in [15]. Specifically, we apply an FFT on the U-net image to get an estimated k-space. The fully k-space is then recovered by integrating the sampled k-space according to the sampling mask. The same sampling masks was used for all cases.

135 A common metric used to quantify the reconstructed image quality is the PSNR. In order to provide a reliable and robust evaluation of the reconstruction quality, which will

be used for medical diagnostic, we suggest the following test: PSNR, brain’s shape measure and tissue segmentation.




Figure 4: 2D Gaussian sampling masks

3.1. Peak signal-to-noise ratio (PSNR)

PSNR measures the Mean Squared Error (MSE) between the fully-sampled MR image and the reconstructed image. Therefore, this measure is not a sufficient metric for edges and general shape. For example, a model can provides a good reconstruction in the sense of PSNR but with blurry edges, which results in a poor performance of algorithms that use them, for example segmentation. However, PSNR is still a good indication for the image quality, especially if an expert should view it. Quantitative evaluation is presented in Table 1 and a graphical visualization in Figure 3.1. The PSNR calculated on the whole image without masking. Note that the proposed method outperforms the other.

PSNR Method	Factor 2.5	Factor 4	Factor 6
Zero-filled	32.044 ± 2.616	26.447 ± 1.997	16.470 ± 2.181
CS-MRI	39.053 ± 2.479	33.273 ± 2.452	26.951 ± 3.380
IM-CNN-L2	35.281 ± 2.210	31.577 ± 2.109	23.417 ± 2.676
CNN-L2	38.978 ± 2.454	33.405 ± 2.232	31.010 ± 2.299
Proposed	39.802 ± 2.489	34.595 ± 2.519	31.555 ± 2.487

Table 1: Error in PSNR, without masking




Figure 5: PSNR without masking error-bar

An extended validation of our model done by calculating the PSNR measure on different brain tissues. We applied an image segmentation algorithm, FAST [27], on the original
 150 fully-sampled MR image and then use it for calculating the masked-PSNR on gray matters, white matters and the Cerebrospinal Fluid (CSF). Results are presented in Table 2.

PSNR Method	Factor 2.5			Factor 4			Factor 6		
	White	Gray	CSF	White	Gray	CSF	White	Gray	CSF
Zero-filled	41.494±3.48	36.736±3.32	38.546±2.45	37.279±2.18	34.379±2.52	36.369±1.87	24.887±1.62	24.987±1.91	29.337±2.08
CS-MRI	44.374±4.20	40.042±3.84	40.744±3.16	41.655±3.12	37.247±2.99	39.119±2.43	35.595±3.15	33.111±2.91	36.648±2.26
IM-CNN-L2	42.992±3.81	38.827±3.69	39.922±3.11	40.574±2.66	37.102±2.90	38.473±2.31	34.742±2.89	31.496±2.43	33.710±1.80
CNN-L2	45.727±4.70	41.593±4.16	41.689±3.69	43.265±2.95	40.151±3.15	40.247±2.47	41.119±2.48	39.277±2.65	39.102±1.93
Proposed	46.206±4.77	41.942±4.34	41.616±2.45	43.857±3.16	40.109±3.25	40.496±2.57	41.919±2.77	39.014±2.88	39.366±2.17

Table 2: Error in PSNR, with masking

ONLY 2 DIGITS - WHICH IS BETTER??

PSNR Method	Factor 2.5			Factor 4			Factor 6		
	White	Gray	CSF	White	Gray	CSF	White	Gray	CSF
Zero-filled	41.49±3.48	36.73±3.32	38.54±2.45	37.27±2.18	34.37±2.52	36.36±1.87	24.88±1.62	24.98±1.91	29.33±2.08
CS-MRI	44.37±4.20	40.04±3.84	40.74±3.16	41.65±3.12	37.24±2.99	39.11±2.43	35.59±3.15	33.11±2.91	36.64±2.26
IM-CNN-L2	42.99±3.81	38.82±3.69	39.92±3.11	40.57±2.66	37.10±2.90	38.47±2.31	34.74±2.89	31.49±2.43	33.71±1.80
CNN-L2	45.72±4.70	41.59±4.16	41.68±3.69	43.26±2.95	40.15±3.15	40.24±2.47	41.11±2.48	39.27±2.65	39.10±1.93
Proposed	46.20±4.77	41.94±4.34	41.61±2.45	43.85±3.16	40.11±3.25	40.49±2.57	41.91±2.77	39.01±2.88	39.36±2.17

Table 3: Error in PSNR, with masking




Figure 6: Examples of reconstructed MR images from under-sampled k-space.

3.2. Brain Extraction - Skull Stripping

Brain extraction (Skull stripping) is an algorithm that delineates the brain boundary
 (Figure 7). It is necessary for almost every brain analysis algorithm. In tissue segmentation
 155 for example, skull stripping is a pre-processing step which affects directly on the segments
 partition. We examine the different reconstruction methods by applying the Brain Extrac-
 tion Tool (BET) [28] on each different MR reconstructed image. Then, we compared the
 skull stripping results to the fully-sampled result using the Modified Hausdorff Distance
 160 (MHD) [29].

Let $C_1, C_2 \in R^2$ denotes the brain contours extract from skull stripping algorithm on
 two different reconstruction methods respectively. MHD measures the distance between the

contours such that:

$$\begin{aligned}
MHD(C_1, C_2) &= \max \left\{ \frac{1}{|C_1|} \sum_{\sigma_1 \in C_1} d(\sigma_1, C_2), \frac{1}{|C_2|} \sum_{\sigma_2 \in C_2} d(\sigma_2, C_1) \right\} \\
d(\sigma_1, C_2) &= \min_{\sigma_2 \in C_2} \|\sigma_1 - \sigma_2\| \\
d(\sigma_2, C_1) &= \min_{\sigma_1 \in C_1} \|\sigma_2 - \sigma_1\|
\end{aligned} \tag{7}$$

where σ_1, σ_2 are points on the contours C_1, C_2 respectively (Figure 7.c). MHD values for
165 different sampling ratios are presented in Table 4.

MHD Method	Factor 2.5	Factor 4	Factor 6
Zero-filled	1.111 ± 0.563	2.617 ± 1.214	3.121 ± 1.279
CS-MRI	0.701 ± 0.511	1.447 ± 1.027	3.114 ± 1.617
IM-CNN-L2	0.654 ± 0.414	0.878 ± 0.548	2.760 ± 1.748
CNN-L2	0.420 ± 0.270	0.715 ± 0.561	1.083 ± 1.052
Proposed	0.391 ± 0.250	0.617 ± 0.306	1.050 ± 1.033

Table 4: MHD - brain extraction




Figure 7: Example of brain extraction

3.3. Brain Segmentation

Brain segmentation is the task of partitioning the brain image (MRI in our case), into the following segments: background, white matter, gray matter and CSF. This process is of a great importance for brain MR image analysis, especially for therapy planning and
170 also for clinical research. Sampling artifacts, results in poor signal to noise and blurry

edges, may hurts the algorithm performance. We suggest a way to evaluate and compare different MRI reconstruction methods based on segmentation results. We create a reference by extracting the brain (using BET [28]) from the original fully-sampled image and run a brain segmentation algorithm (FAST [27]). Then, we did the same process for all the
175 reconstruction methods and compared the segmentation results to the reference.

The measure used for segmentations comparison is the Dice score [30] between the original and the reconstructed image segments. Dice score is a measure of sets similarity which gets the values between 0 (no similarity) to 1.0 (max similarity). Denotes by $S_{i,j}$ the set of pixels in segment j for reconstruction method i , the Dice score between this method to the reference is:

$$DICE(S_{ref,j}, S_{i,j}) = \frac{2 \cdot |S_{ref,j} \cap S_{i,j}|}{|S_{ref,j}| + |S_{i,j}|} \quad (8)$$

Dice values for different sampling ratios are presented in Table 5. Our methods achieved the best Dice score in all brain tissues.

DICE Method	Factor 2.5			Factor 4			Factor 6		
	White	Gray	CSF	White	Gray	CSF	White	Gray	CSF
Zero-filled	0.882±0.08	0.826±0.05	0.796±0.05	0.718±0.10	0.644±0.05	0.627±0.05	0.599±0.12	0.466±0.07	0.050±0.06
CS-MRI	0.942±0.05	0.910±0.03	0.868±0.03	0.871±0.10	0.805±0.08	0.770±0.05	0.708±0.14	0.639±0.08	0.621±0.07
IM-CNN-L2	0.908±0.04	0.863±0.03	0.839±0.03	0.832±0.05	0.754±0.03	0.754±0.03	0.695±0.11	0.590±0.06	0.560±0.06
CNN-L2	0.948±0.02	0.919±0.02	0.891±0.02	0.888±0.06	0.836±0.04	0.813±0.03	0.836±0.09	0.770±0.06	0.747±0.06
Proposed	0.954±0.01	0.928±0.02	0.900±0.002	0.903±0.06	0.858±0.04	0.833±0.03	0.851±0.09	0.789±0.06	0.767±0.06

Table 5: Segmentation Dice score for different sampling ratios

4. Discussion

We proposed a software-only framework, using GANs for accelerating MRI acquisition.
180 Specifically, high-quality MRI reconstruction using only 40%, 25% and 16.6% of the original k-space data is demonstrated. The key idea is based on utilizing an adversarial loss in addition to L2 loss. Extensive evaluation of the reconstruction quality was conducted in order to emphasize the advantage of proposed method. Future work will concentrate on generation of MRI in the presence of pathologies.

4.1. Acknowledgment

This research is partially supported by the Israel Science Foundation (T.R.R. 1638/16) and the IDF Medical Corps (T.R.R.)




Figure 8: Examples of reconstructed MR images from under-sampled k-space.

References

- [1] S. Moeller, E. Yacoub, C. A. Olman, E. Auerbach, J. Strupp, N. Harel, K. Ugurbil,
190 Multiband multislice ge-epi at 7 tesla, with 16-fold acceleration using partial parallel imaging with application to high spatial and temporal whole-brain fmri, *Magnetic Resonance in Medicine* 63 (5) (2010) 1144–1153.
- [2] P. B. Roemer, W. A. Edelstein, C. E. Hayes, S. P. Souza, O. Mueller, The nmr phased array, *Magnetic resonance in medicine* 16 (2) (1990) 192–225.
- [3] A. Deshmane, V. Gulani, M. A. Griswold, N. Seiberlich, Parallel mr imaging, *Journal of Magnetic Resonance Imaging* 36 (1) (2012) 55–72.
195
- [4] K. P. Pruessmann, M. Weiger, M. B. Scheidegger, P. Boesiger, et al., Sense: sensitivity encoding for fast mri, *Magnetic resonance in medicine* 42 (5) (1999) 952–962.
- [5] M. A. Griswold, P. M. Jakob, R. M. Heidemann, M. Nittka, V. Jellus, J. Wang,
200 B. Kiefer, A. Haase, Generalized autocalibrating partially parallel acquisitions (grappa), *Magnetic resonance in medicine* 47 (6) (2002) 1202–1210.
- [6] D. L. Donoho, Compressed sensing, *Information Theory, IEEE Transactions on* 52 (4) (2006) 1289–1306.
- [7] M. Lustig, D. Donoho, J. M. Pauly, Sparse mri: The application of compressed sensing
205 for rapid mr imaging, *Magnetic resonance in medicine* 58 (6) (2007) 1182–1195.
- [8] M. Usman, G. Vaillant, D. Atkinson, T. Schaeffter, C. Prieto, Compressive manifold learning: Estimating one-dimensional respiratory motion directly from undersampled k-space data, *Magnetic Resonance in Medicine* 72 (4) (2014) 1130–1140.
- [9] K. K. Bhatia, J. Caballero, A. N. Price, Y. Sun, J. V. Hajnal, D. Rueckert, Fast
210 reconstruction of accelerated dynamic mri using manifold kernel regression, in: *Medical Image Computing and Computer-Assisted Intervention–MICCAI 2015*, Springer, 2015, pp. 510–518.
- [10] S. Ravishankar, Y. Bresler, Mr image reconstruction from highly undersampled k-space data by dictionary learning, *Medical Imaging, IEEE Transactions on* 30 (5) (2011)
215 1028–1041.

- [11] J. Caballero, A. N. Price, D. Rueckert, J. V. Hajnal, Dictionary learning and time sparsity for dynamic mr data reconstruction, *Medical Imaging, IEEE Transactions on* 33 (4) (2014) 979–994.
- [12] S. Wang, Z. Su, L. Ying, X. Peng, S. Zhu, F. Liang, D. Feng, D. Liang, Accelerating magnetic resonance imaging via deep learning, in: *Biomedical Imaging (ISBI), 2016 IEEE 13th International Symposium on*, IEEE, 2016, pp. 514–517.
- [13] O. Oktay, W. Bai, M. Lee, R. Guerrero, K. Kamnitsas, J. Caballero, A. de Marvao, S. Cook, D. OARegan, D. Regan, D. Rueckert, Multi-input cardiac image super-resolution using convolutional neural networks, in: *International Conference on Medical Image Computing and Computer-Assisted Intervention*, Springer, 2016, pp. 246–254.
- [14] D. Lee, J. Yoo, J. C. Ye, Deep residual learning for compressed sensing mri, in: *Biomedical Imaging (ISBI 2017), 2017 IEEE 14th International Symposium on*, IEEE, 2017, pp. 15–18.
- [15] C. M. Hyun, H. P. Kim, S. M. Lee, S. Lee, J. K. Seo, Deep learning for undersampled mri reconstruction, *arXiv preprint arXiv:1709.02576*.
- [16] O. Ronneberger, P. Fischer, T. Brox, U-net: Convolutional networks for biomedical image segmentation, in: *International Conference on Medical Image Computing and Computer-Assisted Intervention*, Springer, 2015, pp. 234–241.
- [17] C. M. Sandino, N. Dixit, J. Y. Cheng, S. S. Vasanawala, Deep convolutional neural networks for accelerated dynamic magnetic resonance imaging, preprint.
- [18] I. Goodfellow, J. Pouget-Abadie, M. Mirza, B. Xu, D. Warde-Farley, S. Ozair, A. Courville, Y. Bengio, Generative adversarial nets, in: *Advances in neural information processing systems*, 2014, pp. 2672–2680.
- [19] A. Radford, L. Metz, S. Chintala, Unsupervised representation learning with deep convolutional generative adversarial networks, *arXiv preprint arXiv:1511.06434*.
- [20] D. Pathak, P. Krahenbuhl, J. Donahue, T. Darrell, A. A. Efros, Context encoders: Feature learning by inpainting, in: *Proceedings of the IEEE Conference on Computer Vision and Pattern Recognition*, 2016, pp. 2536–2544.

- [21] S. Yu, H. Dong, G. Yang, G. Slabaugh, P. L. Dragotti, X. Ye, F. Liu, S. Arridge, J. Keegan, D. Firmin, et al., Deep de-aliasing for fast compressive sensing mri, arXiv preprint arXiv:1705.07137.
- [22] M. Mardani, E. Gong, J. Y. Cheng, S. Vasanawala, G. Zaharchuk, M. Alley, N. Thakur, S. Han, W. Dally, J. M. Pauly, et al., Deep generative adversarial networks for compressed sensing automates mri, arXiv preprint arXiv:1706.00051.
- [23] D. Nie, R. Trullo, C. Petitjean, S. Ruan, D. Shen, Medical image synthesis with context-aware generative adversarial networks, arXiv preprint arXiv:1612.05362.
- [24] M. Arjovsky, S. Chintala, L. Bottou, Wasserstein gan, arXiv preprint arXiv:1701.07875.
- [25] K. He, X. Zhang, S. Ren, J. Sun, Deep residual learning for image recognition, in: Proceedings of the IEEE Conference on Computer Vision and Pattern Recognition, 2016, pp. 770–778.
- [26] X. Glorot, Y. Bengio, Understanding the difficulty of training deep feedforward neural networks., in: Aistats, Vol. 9, 2010, pp. 249–256.
- [27] Y. Zhang, M. Brady, S. Smith, Segmentation of brain mr images through a hidden markov random field model and the expectation-maximization algorithm, IEEE transactions on medical imaging 20 (1) (2001) 45–57.
- [28] S. M. Smith, Fast robust automated brain extraction, Human brain mapping 17 (3) (2002) 143–155.
- [29] M.-P. Dubuisson, A. K. Jain, A modified hausdorff distance for object matching, in: Pattern Recognition, 1994. Vol. 1-Conference A: Computer Vision & Image Processing., Proceedings of the 12th IAPR International Conference on, Vol. 1, IEEE, 1994, pp. 566–568.
- [30] L. R. Dice, Measures of the amount of ecologic association between species, Ecology 26 (3) (1945) 297–302.