

AIRDATA MULTIMETER CERTIFICATE OF RECALIBRATION

S/N: M86253
Order #: R19171

LOW VELOCITY CONFIRMATION (FPM)

TEST METER TOLERANCE = $\pm 3.0\% \pm 7$ FPM AS-RECEIVED TEST WITHIN SPEC YES NO N/A See Notes

Vel Eqv Trans Std: S/N: M02009	As-Rcvd	<u>Test 2</u>	Test 3	Vel Eqv Trans Std: S/N: M10840	As-Rcvd	Test 2	Test 3
Vel Eqv Trans Std: S/N: M02803	As-Rcvd	<u>Test 2</u>	Test 3	Vel Eqv Trans Std: S/N: M10897	As-Rcvd	Test 2	Test 3
Vel Eqv Trans Std: S/N: M02903	As-Rcvd	Test 2	Test 3	Vel Eqv Trans Std: S/N: M10901	<u>As-Rcvd</u>	Test 2	Test 3
Vel Eqv Trans Std: S/N: M10839	As-Rcvd	Test 2	Test 3	Vel Eqv Trans Std: S/N: M13492	As-Rcvd	Test 2	Test 3

Approx Set Point	Standard	Test Meter	Diff	Standard	Test Meter	Diff	Standard	Test Meter	Diff
100	107	107	0	110	110	0	N/A		
500	512	513	1	518	515	-3			

ADM-880C, ADM-870/870C and ADM-860/860C models are read in AirFoil Mode. ADM-850/850L models are read in Pitot Tube Mode.

TEMPERATURE TEST - AIRDATA MULTIMETER (° F)

TEST METER TOLERANCE = $\pm 0.2^\circ$ F AS-RECEIVED TEST WITHIN SPEC YES NO N/A See Notes

RTD Simulator: S/N 249	<u>As-Rcvd</u>	<u>Test 2</u>	Test 3	Set Point: <u>35.6° F</u>	95° F	154.4° F
RTD Simulator: S/N 250	<u>As-Rcvd</u>	<u>Test 2</u>	Test 3	Set Point: 35.6° F	<u>95° F</u>	154.4° F
RTD Simulator: S/N 253	<u>As-Rcvd</u>	<u>Test 2</u>	Test 3	Set Point: 35.6° F	95° F	<u>154.4° F</u>
RTD Simulator: S/N 254	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F
RTD Simulator: S/N 256	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F
RTD Simulator: S/N 257	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F
RTD Simulator: S/N 292	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F
RTD Simulator: S/N 293	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F
RTD Simulator: S/N 294	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F
RTD Simulator: S/N 313	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F
RTD Simulator: S/N 314	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F
RTD Simulator: S/N 315	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F
RTD Simulator: S/N 316	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F
RTD Simulator: S/N 317	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F
RTD Simulator: S/N 318	As-Rcvd	Test 2	Test 3	Set Point: 35.6° F	95° F	154.4° F

RTD Simulator Temperature Equivalent Set Point	Test Meter	Difference	Test Meter	Difference	Test Meter	Difference
35.60	35.6	0	35.7	.1	N/A	
95.00	95.0	0	95.1	.1		
154.40	154.5	.1	154.5	.1		

NOTES: _____

The enclosed ADM Calibration Standards for Pressure and Temperature form(s) is/are an integral part of this calibration and must remain with this Certificate of Calibration. Note: There may be more than one such form included that pertains to this calibration.

Shortridge Instruments, Inc.
7855 East Redfield Road Scottsdale, Arizona 85260
(480) 991-6744 • Fax (480) 443-1267 • www.shortridge.com