

External Validation Performance

Disease	External Dataset	Dataset Distribution	Optimized Threshold	AUC	Accuracy (%)	Sensitivity (%)	Specificity (%)
DR	Messidor-2	1017 / 727	0.5	0.8364	78.3	70.3	84.1
Glaucoma	REFUGE2	360 / 40	0.98	0.8995	89.0	80.0	90.0
AMD	OCT2017	250 / 500	0.5	1.0	99.9	99.8	100.0
DED	Roboflow Dry Eye Dataset	50 / 89	0.539	0.7126	63.3	56.2	76.0