

5(a)	pulses of ultrasound	B1
	ultrasound incident on quartz crystal	B1
	waves make crystal oscillate	B1
	oscillations (of crystal) generates an e.m.f. (across the crystal)	B1
5(b)	specific acoustic impedances of air and skin are very different	B1
	intensity reflection coefficient depends on difference between acoustic impedance	B1
	most ultrasound reflected so little transmission	B1

Question	Answer	Marks
-----------------	---------------	--------------