

Section B

- 9 (a) (i) edges can be (clearly) distinguished B1 [1]
- (ii) e.g. size of X-ray source / anode / target / aperture
 scattering of X-ray beam
 pixel size
(any two, 1 each) B2
 further detail e.g. use of lead grid B1 [3]
- (b) X-ray image involves a single exposure B1
 CT scan: exposure of a slice from many different angles M1
 repeated for different slices A1
 CT scan involves a (much) greater exposure B1 [4]

- 10 (a) e.g. infinite input impedance / resistance