ISO 17025 CALIBRATION CERTIFICATE

Certificate Number: MNT-6506

This Calibration Certificate shall not be reproduced, except in its full entirety, without the written permission of Sensor Calibrations.

Customer

Monnit Corporation
3400 South, West Temple
South Salt Lake, Utah 84115

Device

Manufacturer: Monnit Corp
Model: MNS2-9-W2-TS-ST-L03
Identifier/Serial: 707250
Condition: Functional

Calibration Summary

Calibration Date: 08/30/2024
Calibration Procedure: 4
Result: PASS

Next Calibration Date*: 10/3/2026
Lab Conditions: 31.62% RH @ 21.15°C

Calibration Data

Set Point	Reference	As Found	Adjustment	As Left	Tolerance	Uncertainty†	Status
28°C	28.0°C	28.0°C	0.0°C	28.0°C	1.0°C	0.011°C	PASS
-25°C	-24.9°C	-24.8°C	N/A	-24.8°C	1.0°C	0.089°C	PASS
90°C	89.9°C	89.7°C	N/A	89.7°C	1.0°C	0.022°C	PASS

† Uncertainty data presented above is Total Expanded Uncertainty (TEU). The coverage factor of K = 2 results in a coverage probability of approximately 95%.

Reference Standards

Asset	Description	Cal-Date	Due-Date
S000115	Fluke 5615 PRT	08/23/2024	04/23/2026
S000116	Fluke 5615 PRT	07/25/2024	01/25/2026
S000128	Fluke 5615 PRT	08/16/2023	02/14/2025
S000134	Fluke 1586A "SUPER-DAQ"	03/25/2024	09/25/2025

NIST Traceability and Disclaimers

All reference standards used in this calibration are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) or an equivalent National Measurement Institute that is a signatory to CIPM MRA.

The decision rule is Simple Acceptance (SA) with a Test Uncertainty Ratio (TUR) greater than or equal to 4:1 .

Details of our scope of accreditation are available at www.A2LA.org. The parameters of this calibration are covered under Sensor Calibrations' current scope of accreditation.

The results in this certificate apply only to the calibration of the device identified above.

*The listed Next Calibration Date is provided by the customer and does not imply continuing conformance during the resultant calibration interval between the Calibration Date and the Next Calibration Date.

Calibration Technician
Colton Robinson

Authorizer
Kelly S Lewis

Date Report Issued: 9/3/2024

Certified at Sensor Calibrations LLC