

Sea & Sun Technology GmbH

Document of Calibration

Manufacturer: Sea & Sun Technology GmbH

Date: 22.05.2023

Type / No.: CTD60M / 180

Sensor: Conductivity / Range : 0..70 mS/cm

Standard Deviation: 0,000832

Calibration Coefficients:

Polynomial:

$$\text{Cond [mS/cm]} = \sum_{i=0}^5 A[i] \cdot X^i, X = \text{raw value}$$

A[0] = -6,10007E-02
 A[1] = 1,04493E-03
 A[2] = 2,07684E-11
 A[3] = 0,00000E+00
 A[4] = 0,00000E+00
 A[5] = 0,00000E+00

Rawdata / digits	Reference / mS/cm	Calculated / mS/cm	Residual
59	0,00000	0,00065	-0,000650
10383	10,79100	10,79077	0,000228
10432	10,84200	10,84200	0,000005
20626	21,50100	21,50061	0,000388
20736	21,61600	21,61565	0,000351
30560	31,89200	31,89153	0,000468
30665	32,00200	32,00138	0,000617
41146	42,96900	42,96895	0,000048
41214	43,04000	43,04012	-0,000123
50339	52,59100	52,59248	-0,001483
50535	52,79600	52,79770	-0,001700
60845	63,59600	63,59480	0,001197
61025	63,78400	63,78335	0,000654

Remarks:

Name: Jolanta Rerich

Signum:





Sea & Sun Technology GmbH

Document of Calibration

Manufacturer: Sea & Sun Technology GmbH

Date: 23.05.2023

Type / No.: CTD60M / 180

Sensor: Pressure / Range : 0..6000 dbar

Standard Deviation: 0,224792

Polynomial:

$$\text{Press [dbar]} = \left(\sum_{i=0}^4 A[i] * X^i \right) + A[5], X = \text{raw value}$$

Calibration Coefficients:

A[0] = -3,55019E+02

A[1] = 1,04987E-01

A[2] = -5,02180E-08

A[3] = -1,21191E-13

A[4] = 0,00000E+00

A[5] = 0,00000E+00

Rawdata / digits	Reference / dbar	Calculated / dbar	Residual
3388	0,14000	0,09581	0,044189
9140	600,14000	600,27460	-0,134599
14922	1200,14000	1200,01300	0,127212
20744	1800,14000	1800,14000	-0,000474
26602	2400,14000	2400,02700	0,113307
32507	3000,14000	3000,56600	-0,425554
38444	3600,14000	3599,99700	0,143008
44430	4200,14000	4199,79400	0,346376
50470	4800,14000	4800,18000	-0,039536
56560	5400,14000	5400,47000	-0,329877
62696	6000,14000	5999,98400	0,155948

Remarks:

Name: André Schröder

Signum:

Sea & Sun Technology GmbH

Constants Protocol

Probe Number: CTM180

Date: 25.05.2023

COM (Baud rate, Parity, Databits, Stopbit): 1200,o,8,1

Vbatt	Dimension: Volt	A[0]: -6.11455E+00
Address: 1	Type: 1 D-cell, Alkaline	A[1]: 1.74586E-04
Parameter: Battery voltage		A[2]: 0.00000E+00
Range: 1..5		A[3]: 0.00000E+00
Calibration Range: 1..5	Calibration Date: 02.11.2006	A[4]: 0.00000E+00
		A[5]: 0.00000E+00
Press	Dimension: dbar	A[0]: -3.55019E+02
Address: 2	Type: PA8-600	A[1]: 1.04987E-01
Parameter: Pressure		A[2]: -5.02180E-08
Range: 0..6000		A[3]: -1.21191E-13
Calibration Range: 0..6000	Calibration Date: 23.05.2023	A[4]: 0.00000E+00
		A[5]: 0.00000E+00
Temp	Dimension: °C	A[0]: -2.23481E+00
Address: 3	Type: PT100	A[1]: 5.90924E-04
Parameter: Temperature		A[2]: 0.00000E+00
Range: -2..36		A[3]: 0.00000E+00
Calibration Range: 0..30	Calibration Date: 22.05.2023	A[4]: 0.00000E+00
		A[5]: 0.00000E+00
Cond	Dimension: mS/cm	A[0]: -6.10007E-02
Address: 4	Type: small	A[1]: 1.04493E-03
Parameter: Conductivity		A[2]: 2.07684E-11
Range: 0..70		A[3]: 0.00000E+00
Calibration Range: 0..60	Calibration Date: 22.05.2023	A[4]: 0.00000E+00
		A[5]: 0.00000E+00

Remarks:

Signum:

