

Patient Treatment Preparation Procedure Sequence	(300A,0790)	2	Procedures for preparing for the treatment of the patient. Zero or more Items shall be included in this Sequence.
>Patient Treatment Preparation Procedure Index	(300A,0795)	1	Index of the Patient Treatment Preparation Procedure in this Sequence. The value shall start at 1 and increase monotonically by 1.
>Patient Treatment Preparation Procedure Code Sequence	(300A,0791)	1	Procedure performed to prepare the patient for treatment. Only a single Item shall be present in this Sequence.
>>Include <a href="#">Table 8.8-1 "Code Sequence Macro Attributes"</a>			<a href="#">BCID 9577 "Patient Treatment Preparation Procedure"</a> .
>Patient Treatment Preparation Device Sequence	(300A,078F)	3	Devices used for preparing the patient for treatment. Only a single Item is permitted in this Sequence.
>>Include <a href="#">Table C.36.2.2.3-1 "RT Accessory Device Identification Macro Attributes"</a>			<a href="#">BCID 9573 "Patient Treatment Preparation Device"</a> .
>Patient Treatment Preparation Procedure Parameter Description	(300A,078E)	2	User-defined description of patient treatment preparation procedure parameters.
>Patient Treatment Preparation Procedure Parameter Sequence	(300A,0793)	2	Parameters for the Patient Treatment Preparation Procedure. Zero or more Items shall be included in this Sequence.
>>Include <a href="#">Table 10.2.1-1 "Content Item with Modifiers Macro Attributes"</a>			See <a href="#">C.36.25.1.2</a> for Baseline TID.

er(s).