, ,			3		J	DATE: (01-06-2013
SL NO	Description	Make	Range	Least Count	Calibration Frequency	Acceptable Error	Remarks
DEW-3	Dial Caliper	Mitutoyo	0- 300 mm	0.02	12 Months	0.02 mm	
DEW-4	Micro Meter	Mitutoyo	0 -25 mm	0.01	12 Months	0.02 mm	
DEW 5	Micro Meter	Mitutoyo	25-50mm	0.01	12 Months	0.02 mm	
DEW 6	Micro Meter	Mitutoyo	50-75mm	0.01	12 Months	0.02 mm	
DEW 9	Micro Meter	Mitutoyo	75-100mm	0.01	12 Months	0.02 mm	
DEW-34	Digital Vernier Caliper	Mitutoyo	0 -600 mm	0.01	12 Months	0.02 mm	
DEW 35	Digital Height gage	TESA	0-600mm	0.001	24 Months	0.01	
DEW 27	Slip Gages	Peacock	1to100mm	0.0002	60 Months	0.002mm	
DEW-38	Pre-setter & measuring machin	e Zoller	400mm	0.0001	18 Months	0.001mm	
DEW-8	Vernier Caliper	Mitutoyo	0 -600 mm	0.02	12 Months	0.02 mm	
DEW 39	Tool makers microscope	Radical	25x25mm	0.01	12 Months	0.02 mm	
DEW 30	Dial vernier Caliper	Mitutoyo	0-150mm	0.02	12 Months	0.02 mm	
DEW 40.1	Digital Vernier Caliper	Mitutoyo	0-150mm	0.01	12 Months	0.02 mm	
DEW 40.2	Digital Vernier Caliper	Mitutoyo	0-150mm	0.01	12 Months	0.02 mm	
DEW 4.1	Digital micrometer	Mitutoyo	0-25mm	0.01	12 Months	0.02 mm	
DEW 32	Dial Bore gage	Mitutoyo	10-18.5mm	0.01	12 Months	0.02 mm	
DEW 4.6	Disc micrometer	Mitutoyo	0-25mm	0.01	12 Months	0.02 mm	
DEW 31	Dial Bore gage	Mitutoyo	18-35mm	0.01	12 Months	0.02 mm	
DEW 16	Dial Bore gage	Mitutoyo	35-60mm	0.01	12 Months	0.02 mm	
DEW 17	Dial Bore gage	Mitutoyo	50-100mm	0.01	12 Months	0.02 mm	
DEW 41	СММ	Carl-Zeiss	700x1000x6 00mm	0.0002	18 Months	0.0018mm	
DEW 36	Dial vernier Caliper	Mitutoyo	0-200mm	0.01	12 Months	0.02 mm	
DEW 36.1	Dial vernier Caliper	Mitutoyo	0-200mm	0.01	12 Months	0.02 mm	
DEW 40.3	Digital Vernier Caliper	Mitutoyo	0-150mm	0.01	12 Months	0.02 mm	
DEW 40.4	Digital Vernier Caliper	Mitutoyo	0-150mm	0.01	12 Months	0.02 mm	
DEW 40.5	Digital Vernier Caliper	Mitutoyo	0-150mm	0.01	12 Months	0.02 mm	
DEW 40.6	Digital Vernier Caliper	Mitutoyo	0-150mm	0.01	12 Months	0.02 mm	
DEW 4.2	Digital micrometer	Mitutoyo	0 -25 mm	0.01	12 Months	0.02 mm	
DEW 4.3	Digital micrometer	Mitutoyo	0-25mm	0.01	12 Months	0.02 mm	

List Of Measuring Instruments And Gages

RECORD REF: DEW/QA/D/01

Approved By:

Divya Engineering Works (P) Ltd

Divya Engineering Works (P) Ltd		List Of Measuring Instruments And Gages					ORD REF: DEW/QA/D/01
				J		E: 01-06-2013	
							JE NO: 06
DEW 4.4	Digital micrometer	Mitutoyo	0-25mm	0.01	12 Months	0.02 mm	
DEW 4.5	Digital micrometer	Mitutoyo	0-25mm	0.01	12 Months	0.02 mm	
DEW 5.1	Micro Meter	Mitutoyo	25-50mm	0.01	12 Months	0.02 mm	
DEW 33	Micro Meter	Mitutoyo	100-200mm	0.01	12 Months	0.02 mm	
DEW 15	Dial Bore gage	Mitutoyo	18-35mm	0.01	12 Months	0.02 mm	
DEW 16.1	Dial Bore gage	Mitutoyo	35-60mm	0.01	12 Months	0.02 mm	
DEW 31.1	Dial Bore gage	Mitutoyo	18-35mm	0.01	12 Months	0.02 mm	
DEW 32.1	Dial Bore gage	Mahr	4-7mm	0.001	12 Months	0.01mm	
DEW 7	Dial Gage	Mitutoyo	0-1mm	0.01	12 Months	0.02 mm	
DEW 7.1	Dial Gage	Mitutoyo	0-10mm	0.01	12 Months	0.02 mm	
DEW 7.2	Dial Gage	Mitutoyo	0-0.6mm	0.01	12 Months	0.02 mm	
DEW 7.3	Dial Gage	Mitutoyo	0-0.6mm	0.01	12 Months	0.02 mm	
DEW 7.4	Dial Gage	Mitutoyo	0-0.8mm	0.01	12 Months	0.02 mm	
DEW 21	Dial Gage	Mitutoyo	0-1mm	0.001	12 Months	0.01 mm	
DEW 22	Dial Gage	Mitutoyo	0-10mm	0.01	12 Months	0.02 mm	
DEW 7.5	Dial Gage	Mahr	-0.1 to 0.1mm	0.002	12 Months	0.02 mm	
DEW 7.6	Millimess	Mahr	0-0.1mm	0.001	12 Months	0.01 mm	
DEW 35.1	Digital Height gage	Mahr	0-600mm	0.001	24 Months	0.01 mm	
DEW 24	Surface plate	VTG	1000x1000 mm	0.007	60 Months	0.01 mm	
DEW 42	Surface plate	MMT	1600x1000 mm	0.006	36 Months	0.008mm	
DEW 43	Hygro Thermometer	Vartech	0to50deg,30 to90	0.1	12 Months	1deg C	
101-1	Plain plug gage	KCP	6H7	0	12 Months	0.0025	
101-3	Plain plug gage	KCP	12H7	0	12 Months	0.003	
101-4	Plain plug gage	KCP	10H7	0	12 Months	0.0025	
101-5	Plain plug gage	KCP	5H8	0	12 Months	0.0025	
101-6	Plain plug gage	KCP	4H7	0	12 Months	0.0025	
501-2	Thread plug gage	TGP	2-56-UNC 2B	0	12 Months	0.005	
501-3	Thread plug gage	THM	M4x0.7-6H	0	12 Months	0.009	
501-4	Thread plug gage	GRAPHICA	M8x1.25-6H	0	12 Months	0.010	
501-5	Thread plug gage	HIP	M8-6H	0	12 Months	0.010	

Approved By:

Divya Engineering Works (P) Ltd		List Of Measuring Instruments And Gages					RECORD REF: DEW/QA/D/01
, g (. , 		3 11 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1					DATE: 01-06-2013
			•				ISSUE NO: 06
501-6	Thread plug gage	SIZE CONTROL	M6x1.0-6H	0	12 Months	0.010)
501-7	Thread plug gage	QUALITY MASTER	M5x0.8-6H	0	12 Months	0.009)
501-8	Thread plug gage	TGP	M10x1.5-6H	0	12 Months	0.010)
501-9	Thread plug gage	QUALITY MASTER	M6x1.0-6H	0	12 Months	0.010)
501-10	Thread plug gage	THM	M3x0.5-6H	0	12 Months	0.009)
501-11	Thread plug gage	SIZE CONTROL	M5x0.8-6H	0	12 Months	0.009)
501-12	Thread plug gage	THM	M2.5x0.45-6 H	0	12 Months	0.009)
501-13	Thread plug gage	THM	4-40-UNC 2B	0	12 Months	0.005	5
501-14	Thread plug gage	Micro	M8x1.25-6H- Heli	0	12 Months	0.009	
501-15	Thread plug gage	Baker	10-32-UNF2 B	0	12 Months	0.007	76
501-16	Thread plug gage	Micro	10-32-UNF 3B	0	12 Months	0.007	76
501-17	Thread plug gage	TGP	M10x1.25-6 H	0	12 Months	0.010)
501-18	Thread plug gage	Master Metrolog	2-56-UNC 2B	0	12 Months	0.005	5
501-19	Thread plug gage	SIZE CONTROL	M4x0.7-6H	0	12 Months	0.009)
501-20	Thread plug gage	SIZE CONTROL	M10x1.5-6H	0	12 Months	0.010)
501-21	Thread plug gage	Micro	6-32-UNC 2B	0	12 Months	0.007	76
501-22	Thread plug gage	Micro	8-32-UNC 2B	0	12 Months	0.007	76
501-23	Thread plug gage	THM	M12x1.75-6 H	0	12 Months	0.010)
501-24	Thread plug gage	HIP	M12-6H	0	12 Months	0.010)
501-25	Thread plug gage	QUALITY MASTER	M16x2-6H	0	12 Months	0.014	ļ.
501-26	Thread plug gage	HIP	M2.5- 6H	0	12 Months	0.009)

Approved By: