

BME6013C Lab#11

Slade Brooks
M13801712

12.02.25

1 Introduction

As generative AI grows in popularity and capability, there are many opportunities for it to become a useful tool across many different disciplines. One possible application is medical imaging and data processing. Generative AI has the ability to generate convincing recreations of many types of images. It may even be capable of generating anatomically correct medical imagery that contains the relevant biological features for analyzing medical scans or developing tools for processing medical imagery. Furthermore, language models have shown significant use in generating code. This skill may prove useful for assisting engineers in developing programs and methodology for effectively processing images.

This lab sought to explore the limitations of generative AI for medical imagery applications. The lab process employed a generative AI model, Microsoft Copilot, to generate some type of medical image. Multiple image processing and analysis techniques were then employed in an attempt to make the AI-generated image resemble a real medical image of the same region more closely. Attempting this process helps to better understand the current capability of a state-of-the-art generative AI to create accurate medical imagery and make changes to the images based on multiple sequential prompts, as well as to use AI as a coding assistant to help generate sample code for processing and analysis of the images.

2 Methods

2.1 AI Image Generation

The first major technique employed for this lab was employing Microsoft Copilot generative AI and its image generation capabilities. It was decided to use a side-view x-ray of a human skull as the medical image to compare to. Figure 1 shows the real image that was selected to attempt to generate with AI. First, a very simple prompt was used, giving no additional



Figure 1: Real medical image of a side-view x-ray of a human skull.

information or instruction past “generate an image of human skull x-ray side view”. Since the resulting image was fairly similar to the real image, focus was shifted to making small adjustments to better match the real image. The generated image appeared to be largely anatomically correct with some minor errors. The first important change was to attempt to match the head tilt angle so the images could be a better 1:1 comparison. As seen in Appendix 5.1, multiple prompts were used trying to make small adjustments to the chin tilt angle, but the AI would overcorrect and oscillate between a chin position too high and too low. These attempts were abandoned after a few tries. However, during this process, the AI decided to add labels to various features, totally unprompted. The labels were also incorrect. The next prompt removed the labels. Lastly, an attempt was made to match the cropping of the real image. All of the AI-generated images contained the shoulders and upper chest, whereas the real image only covered four vertebra. However, the AI was unable to crop the image. It appeared to want to generate an image in the same aspect ratio every time, and

would have been unable to crop to the desired size without changing this. This attempt was also abandoned after one try, as it was clear the AI would be unable to achieve this. Figure 2 shows the final AI image to be used for the analysis and processing.



Figure 2: AI-generated image of a side-view x-ray of a human skull.

2.2 Image Analysis Techniques

The first main image analysis technique employed was histogram analysis. Histogram analysis allows for a visualization of the color distribution of the image. This can show important features of the images and the differences between them. Histogram analysis can make it clear where major differences are occurring by understanding the distribution and amount of

each color in an image. Furthermore, histograms are valuable for image segmentation. While segmenting an image can be done with guess-and-check methods, a histogram can help inform the guesses so less time is required to determine how to segment the image. Matlab makes histogram analysis of an image incredibly simple. Matlab's `histogram` function accepts an image input and multiple options to generate a histogram of an image. For analysis in this lab, a normalized probability distribution histogram with 256 levels was used. This setting matches the 256-level grayscale colormap used for both images and allows for them to be directly compared since both are normalized. The resulting histogram plots were analyzed to understand the color content of both images, but primarily informed segmentation values. The histogram provided information about what values to select as background, tissue, or other features and what value to set them to during the processing steps.

The other analysis technique employed was Fourier analysis. Taking a 2-D digital Fourier transform (DFT) or a 1-D DFT on a slice of the image provides valuable information about the differences between them. For example, if a DFT shows more high-frequency content in one image compared to the other, it becomes clear that a low-pass filter will make the images more similar. In order to compare two different-sized images, a size-normalization technique was used. The generative AI was queried about methods for comparing two images with vastly different resolution. The most promising technique was to normalize the dimensions to a known feature and take separate DFTs on each image based on the resulting spatial frequencies. Appendix 5.1 shows sample code that was generated by AI, then modified and implemented in the Matlab processing script to compute 1- and 2-D DFTs of the images that could be directly compared. This method normalizes the pixel size of both images to an average skull width, and scales by the size step for plotting so all DFTs are displayed along the same range of normalized spatial frequencies. The 1-D DFTs used were computed on slices of the image. This was used in conjunction with the 2-D DFTs to provide additional information. While the 2-D DFT shows more information, it can be unintuitive to analyze to make processing decisions. Thus, 1-D DFT "slices" were used at some points to better visualize the differences between the images and make filter selections.

2.3 Image Processing Techniques

The first image processing technique used was image segmentation. As discussed previously, the histogram analysis provided incredibly useful for segmenting the AI image. By analyzing the histogram and the image visually, it was possible to determine what values correspond to which anatomical parts of the image. First, the tissue was removed and set to be equal to the background color since there was no tissue visible in the real image. Then, a spot correction was applied to a section where a nonexistent bone was seen on the AI image. This section of pixels was also set to the background color. All segmentation was done with

Matlab boolean matrices and operations (mainly and, not, and or).

The other processing technique used was a simple Gaussian filter. As seen in Appendix 5.1, Copilot was referenced for advice on implementing a Gaussian filter. The sample code provided was not used, but it did lead to a documentation exploration that uncovered a convenient implementation. Matlab has a built-in Gaussian filter function for images, `imgaussfilt`, that accepts a filter strength input. This strength was manually tuned to make a 1-D DFT slice match as close as possible between the AI image and original image. The filtering was applied to the image after segmentation. Some more complex filtering operations were attempted, but the best results came from just a simple Gaussian low-pass filter applied to the whole image, not from applying different filters to different segmented image parts.

3 Results

3.1 Histogram Analysis

The first major analysis completed was a histogram analysis of both images. Figure 3 shows the normalized probability histogram of the real medical image and the original AI image. It




Figure 3: Normalized probability histogram of the real and AI-generated images.

was clear from visual inspection that there was significant difference between the two images in terms of the color distribution. The AI image seemed to be much brighter and have a much more gradual change in its colors. This was confirmed in the histogram. Firstly, it can be seen that there is a much higher level of mid-low dark (0.05–0.3) in the AI image versus the real image. The AI image also appears to have slightly more content in the brighter color range. Overall, the AI image appears to have a relatively flat color content across the majority of the spectrum. This is not surprising as visual inspection of the image shows a large amount of noise, and perfect white noise should have equal content at all colors. Furthermore, a large difference in the background can be observed. The real image has a very concentrated background color of almost 0, making up over 40% of the histogram and all concentrated at one color. The AI image, conversely, has a broad range of background values that spans a significant portion of the dark colors. This is not surprising as the AI image contains tissue that is close to the background color and fades into the background smoothly,

whereas the real image has no tissue and a sharp change between the actual content of the x-ray and the background.

3.2 Fourier Analysis

Both 1-D and 2-D DFT analysis was performed on the real and AI image. First, Fig. 4 shows the 1-D DFT in the x-direction across the skull just barely above the eyes. It can be




Figure 4: 1-D DFT in the x-direction above the eye socket in the real image and AI image.

seen that the AI image has a Fourier transform similar to that of Gaussian white noise. This makes sense as visual inspection of the image shows that it clearly has very high noise levels across every feature. This is in stark contrast to the real image, which has no visible noise and shows a power spectrum that is a fairly constant decrease across the spatial frequency range. Further analysis was completed via a 2-D DFT, shown in Fig. 5. There are several clear differences between the two images. Firstly, the total frequency content. It is clear that the real image has very little frequency content above a normalized spatial frequency of 0.4. The AI image, however, has frequency content all the way to the edge of the normalized spatial frequency range. Again, this follows with the visible noise in the image. The central frequency content is fairly similar. Both images have some interesting lobes with similar magnitude and shape at the center. However, the real image has much large lobes in a slightly different pattern that cover a larger portion of the spatial frequency spectrum. The AI image has some lobes, but they are much smaller and it does not have all the visible lobes that the real image does. This will partially be due to the orientation of the skull in the image, so it is not surprising that there is some difference in the lobe directions. However,




Figure 5: 2-D DFT of the (left) real image and (right) AI image.

significant differences are caused by differences in the anatomical structure in each image. Finally, the AI image has a large blank ring at the mid-frequencies. While the real image has very high frequency content in the center that somewhat linearly falls off as you move further out, the AI image has very high content in the center, then very low content at mid-frequencies, then very high content at high-frequencies.

3.3 Image Processing

The first major processing technique was image segmentation. First, the histogram was used to determine the background and tissue. These were combined and set to the same value as the background had in the real image. The edges were also selected, but were not processed separately. During this process, a spot correction was also applied as described in the Image Processing Techniques. Figure 6 shows the real image, the original AI image after manual cropping, and the segmented AI image. The edges were set to the maximum value to be as bright as possible.

Next, a Gaussian low-pass filter was applied to the segmented image using the Matlab `imgaussfilt`. The value of σ was manually tuned to have the 1-D DFT of the AI image match the 1-D DFT of the real image as closely as possible. Figure 7 shows the result of the best σ value determined, $\sigma = 0.85$ for the same slice previously analyzed with Fourier analysis.

It can be seen that the low-pass filter had a significant impact on the similarity of the spectra. The filtered AI image now has a much more similar rolloff at higher spatial frequency,




Figure 6: 2-D DFT of the (left) real image, (middle) cropped AI image, and (right) segmented AI image.




Figure 7: 1-D DFT in the x-direction above the eye socket in the real image, original AI image, and the filtered AI image.

and appears less similar to Gaussian white noise. This is a significant improvement. Further analysis can be done with the 2D DFTs shown in Fig. 8. It is clear that the Gaussian




Figure 8: 2-D DFT of the (left) real image, (middle) original AI image, and (right) filtered AI image.

low-pass filter significantly improved the overall similarity between the DFTs of the real image and the AI image. The original AI image had lots of high-frequency content and a gap in the mid-frequencies. The filtered image has this high-frequency content removed and its overall magnitude matches the real image 2-D DFT much closer. However, this did not fix the differences in the lobes. The lobes are still somewhat visible, but have become mostly blurred into the surrounding frequency content. Overall, the low-pass filter has made a significant improvement in the similarity of the images. Figure 9 shows the real medical image, the original AI image after being cropped, and the final AI image with segmenting and filtering.




Figure 9: The (left) real image, (middle) original AI image, and (right) final AI image.

Figure 10 shows the final histogram for each of the images. Clearly, the changes had a major impact on the histogram similarity between the real image and the final AI image. The segmenting successfully fixed the spread-out background values that the original AI image




Figure 10: Normalized probability histogram of the real, AI-generated, and filtered AI images.

had by fixing the background values and removing the tissue. It appears that the filtering helped adjust the total color content to better match the real image as well. The shape of the color content throughout the majority of the histogram matches closely between the final image and the real image, and no longer resembles the expected white noise histogram. The final histogram comparison and 2-D DFT shows that while some differences in anatomical structure are present, the filtering and segmenting largely corrected the differences between the AI image and the real image.

The final AI-generated image is fairly similar to the real medical image. Firstly, the bulk of the skull is missing detail that is present in the real image, but appears to be a good approximation of the general content and shape there. Most of the large features are preserved. The mess near the ear, the thickness and shape of the edge of the skull, the tooth and jaw shape, and vertebra shape all appear to match very closely, specially after the processing. The largest anatomical difference is behind the eyes. The real image shows much information in that region, but it became mostly empty after the filtering on the AI image. This may be because the AI is trying to represent a different section of the head, but is likely just incorrect anatomical generation. This was unable to be remedied with the filtering. Overall, the final AI image appears to be a very good anatomical approximation of the real medical image, showing almost all of the same features with very similar shape,

color magnitude, and proportions.

4 Discussion/Conclusion

Both image analysis techniques used provided incredibly useful. The histogram analysis provided helpful information for informing the segmenting process, leading to much less guess-and-check. Matlab's implementation was incredibly easy to use. The Fourier analysis was also an invaluable tool. Understanding the frequency content of both images and how they differed allowed for specifically targeting problem areas and making effective decisions to adjust the image. Taking a 1-D slice instead of a full 2-D transform also allowed for some ease of analysis when tuning the filter, and proved to be accurate enough for the entire 2-D spectra.

Image segmentation was a very useful technique. It can be seen that the segmenting to remove the tissue, fix the background color, and spot-correct an artifact were all imperative to making the image match more closely. We can see in the original histogram that there was significant difference between the AI and real image, but largely due to the segmenting changes, these were remedied in the final version of the image. The combination of the histogram analysis and the image segmenting was a valuable piece in the processing technique. The low-pass filter was also vital. The AI image clearly has extremely high noise levels. While doing a median filter may address the problems, it was clear that a tunable low-pass filter would be able to be tuned to match the spectrum more accurately. Matlab's built-in functionality made this process easy, and manually tuning the filter while using the Fourier analysis to examine the DFT of a slice of the image allowed for quickly determining a good filter strength. It was clear from the final 2-D DFT of the image that the low-pass filter, after being applied to the segmented image, significantly improved the similarity of the images.

Some clear limitations are found in employing these techniques. Firstly, since the AI image has so much noise, segmenting it can be difficult. While clear segments do exist, it is nearly impossible to segment the original image perfectly. This may be fixed by applying the low-pass filter first, then segmenting, then applying more filters, but that does lose some detail and may remove features or make segmenting harder in some cases. Furthermore, while the Fourier analysis can provide useful information, it is not possible to exactly match the DFT of the real image through only processing. As seen in the results, there are many lobes in the real image DFT that likely represent various anatomical features in the image. These lobes are not all present in the AI image. Without employing the AI to generate new information into the image, there is little that can be done through filtering to add in features where they should be. Some erosion or filling may be useful to help replace some missing information, but it is not possible to add anatomically-correct features with the processing, only approximations of them.

Generative AI clearly has some capabilities and limitations in biomedical image processing. Firstly, it is quite capable of generating reasonably accurate medical images. While

some prompts of more difficult images do not function well, basic x-ray images can be generated fairly accurately with minimally-descriptive prompts. One major limitation of the generation fo the images is the stubbornness of the AI. Once it has generated an image, it is quite difficult to get it to make major changes or add real features. It also tends to add features that were not requested, such as the labels it added in one of the prompts. This can be seen in Appendix 5.1, where one of the prompts resulted in a labeled image despite containing no instruction to add such features. The AI showed much capability in terms of processing images. As expected, it is quite competent at code generation. The sample code it generated was able to be implemented almost drag-and-drop from minimally descriptive prompts. The final Matlab code, shown in Appendix 5.2, contains snippets from the AI that were adapted to work with human-writted code. It also provided valuable brainstorming support for methods to compare images.

Overall, generative AI shows much promise in biomedical signal and image processing. While its limitations can make it clunky to use and prevent it from generating truly accurate medical imagery, it seems to generate images that are “close enough” to real medical imagery for various applications. Learning signal and image processing could be done with AI-generated images. There also may be applications for developing processing tools. While the AI appears capable of generating processing code on its own, it also may have an application in using AI-generated medical images to assist humans in developing code or processes for processing real medical imagery without needing real imagery to work on such codes. AI will likely serve as a valuable support tool for biomedical engineers attempting to process medical imagery in the coming years.

5 Appendix

5.1 Copilot Prompts

November 19

generate an image of human skull x-ray side view

 Copilot ⓘ

Here's the generated image of a human skull X-ray in a side view. Would you like me to **add labels for the major bones** (e.g., mandible, maxilla, frontal bone) or **highlight specific regions** like the jaw or cervical vertebrae?



15

that is a good start. can you tilt the chin upwards slightly so that the jaw is just under parallel to the horizontal?

 Copilot 

Here's the updated X-ray image with the chin tilted upward so the jaw is nearly parallel to the horizontal.

Would you like me to **add anatomical labels** (e.g., frontal bone, mandible, cervical vertebrae) or **highlight specific regions** like the jaw joint or sinus cavities?



this is pretty good! let's try moving the chin downward a tiny tiny bit
- I'd like it to be about halfway between the first image and the
most recent one

 Copilot 

Here's the updated X-ray with the chin moved slightly downward—about halfway between the previous two positions.

Would you like me to **fine-tune it even more, or add anatomical labels for clarity?**




ok, let's move the chin back up slightly. almost to the point it was at in the image before this most recent one. also, let's crop the image so it only shows down to the 4th vertebra

 Copilot 

Here's the updated image with the chin moved slightly back up and cropped to show only down to the 4th cervical vertebra.

Would you like me to **keep the labels**, remove them, or add more detailed annotations (e.g., C1–C4 vertebra names)?



[remove the labels](#)

 Copilot [?](#)

Here's the updated image with all labels removed and cropped to show only down to the 4th cervical vertebra.

Would you like me to [adjust brightness/contrast for better clarity](#), or keep it as is?




perfect, looks great. now, try to maintain the image but crop it so that nothing below the bottom of the chin is visible. I would like the image to span from slightly above the top of the head to slightly below the bottom of the chin. no shoulder should be visible

 Copilot 

Here's the cropped version, spanning from slightly above the top of the head to just below the chin, with no shoulder visible.

Would you like me to **enhance the contrast for sharper bone detail**, or keep the current look?



I am using some image processing in matlab to make your generated image look more like a real x-ray. However, the x-ray I have is very low resolution, and that is making the fourier transforms of both images difficult to compare in a 1:1 way. What techniques can I use to adjust the images to make them more compatible for directly comparing the spatial frequency content of both images?

 Copilot ⓘ

To make the two images more compatible for comparing spatial frequency content via Fourier transforms, you need to address differences in resolution, scale, and sampling. Here are some effective techniques:

1. Resample to a Common Resolution

- Use **interpolation** (bilinear, bicubic, or Lanczos) to upscale or downscale the low-resolution image so both images have the same pixel dimensions.
 - This ensures that the frequency domain representation aligns properly.
-

2. Normalize Physical Scale

- If the images represent different physical sizes, apply **geometric scaling** so anatomical features (e.g., skull size) match proportionally.
 - You can estimate scale by measuring a known feature (like skull width) and apply a uniform scaling factor.
-

3. Apply Windowing Before FFT

- Use a **2D window function** (Hann or Hamming) to reduce edge artifacts caused by cropping.
 - This improves the accuracy of frequency comparison by minimizing discontinuities at image borders.
-

4. Convert to Grayscale and Normalize Intensity

- Ensure both images are single-channel grayscale.
- Normalize intensity ranges (e.g., min-max normalization or histogram matching) so contrast differences don't skew frequency analysis.

5. Match Sampling Frequency

- If one image has a different pixel pitch, adjust by resampling so both have equivalent spatial sampling rates.
- This avoids aliasing and ensures comparable frequency content.

6. Apply Low-Pass Filtering to High-Resolution Image

- If the high-resolution image contains fine details absent in the low-resolution one, apply a Gaussian blur or low-pass filter to approximate the frequency cutoff of the lower-resolution image.
- This makes the frequency spectra more comparable.

7. Zero-Pad for FFT

- Pad both images to the same size before computing the FFT.
- Zero-padding improves frequency resolution and ensures alignment in the frequency domain.

Would you like me to write MATLAB code that:

- Resamples both images to the same resolution,
- Normalizes intensity,
- Applies windowing,
- And computes their Fourier transforms for comparison?



let's explore option 2 more. how would I implement the physical scale into the image? Do i just measure the width of the skull in pixels, then using some width for the skull, set my spatial frequency to be a multiple of that pixel width scaled for both images?

 Copilot i

Great question! Yes, the idea is to establish a **physical scale** so that the spatial frequency content corresponds to real-world dimensions rather than arbitrary pixel units. Here's how you can do it:

Step 1: Measure Skull Width in Pixels

- For each image, measure the width of the skull (or another consistent anatomical feature) in pixels.

```
1 skull_width_px = <number of pixels across skull>;
```

Step 2: Assign a Physical Width

- Decide on a real-world width for the skull (e.g., 18 cm for an adult human skull).

```
1 skull_width_cm = 18; % example
```

Step 3: Compute Pixel Size

- Convert pixel dimensions to physical size:

```
1 pixel_size_cm = skull_width_cm / skull_width_px;
```

- This gives you the physical size per pixel for each image.

Step 4: Rescale Spatial Frequencies

- When you compute the Fourier transform, the frequency axis should be expressed in **cycles per cm** instead of cycles per pixel:

≡ Q, ⌂, ⌂, </> MATLAB

```
1 % Example for frequency axis:  
2 freq_axis = (0:N-1) * (1/(N * pixel_size_cm)); % N = number of pixels
```

- This ensures both images share the same physical frequency units.

Step 5: Optional – Resample to Match Physical Scale

- If the pixel sizes differ significantly, resample one image so both have the same pixel size in cm.

≡ Q, ⌂, ⌂, </> MATLAB

```
1 % Use imresize with scale factor:  
2 scale_factor = pixel_size_img1 / pixel_size_img2;  
3 img2_rescaled = imresize(img2, scale_factor);
```

Why This Works

- By normalizing to a physical dimension, you eliminate discrepancies caused by different resolutions.
- The Fourier transform then reflects **true spatial frequencies** (e.g., anatomical detail per cm), making comparisons meaningful.

Do you want me to write a **MATLAB snippet** that:

- Measures skull width in pixels,
- Computes physical scaling,
- Resamples one image,
- And sets up the frequency axis for FFT comparison?

```
% set up frequency vectors
dx = 0.01;
dy = 0.01;
umax = 1/(2*dx);
vmax = 1/(2*dy);
du = 1/(NAIx*dx);
dv = 1/(NAIy*dy);
u = -umax:du:umax-du;
v = -umax:dv:vmax-dv;give me some sample code following this
syntax and variable naming, but add the skull width and
corresponding pixel size for each image (I and AI) as variables that I
will fill in values for. write only a sample code. for example, show me
how you would apply the correction and compute the 0-centered
frequency vector for the AI image
```

 Copilot ⓘ

Here's a sample MATLAB code snippet following your syntax and incorporating skull width and pixel size for physical scaling. It shows how to compute the zero-centered frequency vector for the AI image:

```
1 % --- User-defined variables ---
2 skullWidth_cm_AI = 18;           % Physical skull width in cm (example)
3 skullWidth_px_AI = 1200;          % Skull width in pixels for AI image
4 NAIx = 1024;                     % Image width in pixels
5 NAIy = 1024;                     % Image height in pixels
6
7 % Compute pixel size for AI image (in cm)
8 pixelSize_cm_AI = skullWidth_cm_AI / skullWidth_px_AI;
9
10 % Set up spatial sampling intervals (dx, dy) in cm
11 dx = pixelSize_cm_AI;            % Sampling interval in x-direction
12 dy = pixelSize_cm_AI;            % Assuming square pixels for simplicity
13
14 % Compute frequency limits
15 umax = 1/(2*dx);               % Max frequency in x-direction
16 vmax = 1/(2*dy);               % Max frequency in y-direction
```

≡ Show more lines

```
17
18 % Compute frequency increments
19 du = 1 / (NAIx * dx);
20 dv = 1 / (NAIy * dy);
21
22 % Create zero-centered frequency vectors for AI image
23 u = -umax : du : umax - du;
24 v = -vmax : dv : vmax - dv;
25
26 % Display info
27 fprintf('Pixel size (cm): %.6f\n', pixelSize_cm_AI);
28 fprintf('Frequency range (u): %.2f to %.2f cycles/cm\n', min(u), max(u));
```

Show less

This code:

- Calculates **pixel size** based on skull width and pixel width.
- Converts your frequency axis to **cycles per cm**.
- Creates zero-centered frequency vectors for the AI image.

Would you like me to **extend this snippet** to:

- Include the same calculation for the real X-ray image,
- And show how to **compare their Fourier spectra on a common physical frequency axis?**

□ ⌂ ⌚ ↻ ...

Yesterday

give me some sample code for a simple gaussian filter with sigma as a tunable variable to apply to an image

give me some sample code for a simple gaussian filter with sigma as a tunable variable to apply to an image

Copilot ①

Here's a simple MATLAB example for applying a Gaussian filter with a tunable sigma:

```
1 % --- Sample Code: Gaussian Filter ---
2 % Read image (replace 'image.png' with your image)
3 I = imread('image.png');
4 I_gray = rgb2gray(I); % Convert to grayscale if needed
5
6 % Define sigma (standard deviation)
7 sigma = 2; % You can tune this value
8
9 % Create Gaussian filter
10 filterSize = 2 * ceil(3 * sigma) + 1; % Ensure filter covers ~3*sigma
11 h = fspecial('gaussian', filterSize, sigma);
12
13 % Apply filter
14 I_filtered = imfilter(I_gray, h, 'replicate');
15
16 % Display results
17 figure;
18 subplot(1,2,1); imshow(I_gray); title('Original Image');
19 subplot(1,2,2); imshow(I_filtered); title(['Gaussian Filtered (sigma = ', num2str(sigma), ')']);

```

Show less

Key points:

- `sigma` controls the amount of blur (higher `sigma` = more smoothing).
- `filterSize` is chosen to cover the Gaussian kernel adequately (about $3 \times \sigma$).
- '`replicate`' avoids edge artifacts by replicating border pixels.

Would you like me to also show how to implement the Gaussian filter manually using `meshgrid` and `exp()` (instead of `fspecial`)?

5.2 Matlab Code

```
%%  
% Slade Brooks  
% brooksl@mail.uc.edu  
% 11.19.25  
% BME6013C  
% Lab 11  
  
clear variables  
close all  
  
%% read in both images  
% set image url  
imageURL = "https://media.istockphoto.com/id/465885010/photo/human-skull-x-ray-image.jpg"  
  
% read in image, convert to grayscale  
I = im2double(max(webread(imageURL), [], 3));  
AI = im2double(max(imread("jpeg.jpg"), [], 3));  
  
% plot image to verify  
figure();  
subplot(1, 2, 1)  
imshow(I); axis image;  
xlabel("x (px)"); ylabel("y (px)"); colorbar;  
subplot(1, 2, 2); imshow(AI); axis image;  
xlabel("x (px)"); ylabel("y (px)"); colorbar;  
  
% fix cropping of images to be more similar  
I = I(10:500, 25:end-25);  
AI = AI(50:1059,:);  
figure();  
subplot(1, 2, 1)  
imshow(I); axis image;  
xlabel("x (px)"); ylabel("y (px)"); colorbar;  
subplot(1, 2, 2); imshow(AI); axis image;  
xlabel("x (px)"); ylabel("y (px)"); colorbar;  
  
%% Part 3
```

```

%%% compute histograms of both images
figure()
hold on
histAI = histogram(AI, 256, "Normalization","probability");
histI = histogram(I, 256, "Normalization","probability");
hold off
legend(["AI Image" "Real Image"])
xlim([0 1]); grid on; %ylim([0 1]);
colormap(gray);
c = colorbar('southoutside'); c.TickLabels = []; c.Ticks = [];

% plot zoomed in version
figure()
hold on
histAI = histogram(AI, 256, "Normalization","probability");
histI = histogram(I, 256, "Normalization","probability");
hold off
legend(["AI Image" "Real Image"])
xlim([0 1]); grid on; ylim([0 0.05]);
colormap(gray);
c = colorbar('southoutside'); c.TickLabels = []; c.Ticks = [];

%%% compute spatial freq content of each image
% get size of images
NIx = size(I, 2);
NIy = size(I, 1);
NAIx = size(AI, 2);
NAIy = size(AI, 1);

% set physical dimension and correct
skW_cm = 18;           % Physical skull width in cm (example)
skW_px_AI = 914 - 16;    % Skull width in pixels for AI image
skW_px_I = 458 - 33;
pixelSize_cm_AI = skW_cm/skW_px_AI;
pixelSize_cm_I = skW_cm/skW_px_I;
dxAI = pixelSize_cm_AI;
dyAI = pixelSize_cm_AI;
dxI = pixelSize_cm_I;

```

```

dyI = pixelSize_cm_I;

% set up frequency vectors based on spatial
umaxAI = 1/(2*dxAI);
vmaxAI = 1/(2*dyAI);
umaxI = 1/(2*dxI);
vmaxI = 1/(2*dyI);
duAI = 1/(NAIx*dxAI);
dvAI = 1/(NAIy*dyAI);
duI = 1/(NIx*dxI);
dvI = 1/(NIy*dyI);
uAI = -umaxAI:duAI:umaxAI - duAI;
vAI = -vmaxAI:dvAI:vmaxAI - dvAI;
uI = -umaxI:duI:umaxI - duI;
vI = -vmaxI:dvI:vmaxI - dvI;

% now do the dft of a similar slice through the skull and the normalized 2D
% fft of each image
F_I = fftshift(abs(fft(I(125,:))))/max(abs(fft(I(125,:))), [], "all");
F_AI = fftshift(abs(fft(AI(245,:))))/max(abs(fft(AI(245,:))), [], "all");
F2_I = fftshift(abs(fft2(I)))/max(abs(fft2(I)), [], "all");
F2_AI = fftshift(abs(fft2(AI)))/max(abs(fft2(AI)), [], "all");

% plot FFTs
figure()
hold on
plot(uI*dxI, 20*log10(F_I));
plot(uAI*dxAI, 20*log10(F_AI)); grid on;
hold off
xlim([-0.5 0.5]); ylim([-80 0]);
xlabel("Spatial Frequency in x (u, cycles/cm)"); ylabel("Amplitude (dB)");
legend(["Real Image" "AI Image"]);

% plot 2D ffts
figure()
subplot(1, 2, 1); imagesc(uI*dxI, vI*dyI, 20*log10(F2_I)); axis image; colormap gray; c
 xlabel("Spatial Frequency in x * Pixel Size (u*dx)"); ylabel("Spatial Frequency in y * P
 subplot(1, 2, 2); imagesc(uAI*dxAI, vAI*dyAI, 20*log10(F2_AI)); axis image; colormap gra

```

```

xlabel("Spatial Frequency in x * Pixel Size (u*dx)"); ylabel("Spatial Frequency in y * P

%% Part 4
%%% first segment out tissue and remove
% tune to find tissue and set it as background
background = AI<20/255;
tissue = ~background & AI<125/255;
background = background | tissue;
edges = ~tissue & ~background & AI>225/255;
AI1 = AI;
AI1(background) = (0.997 + 1.994)/2/255;
AI1(edges) = 255/255;
rest = ~background & ~edges;

% spot correct weird artifact
AI1(909:1000,522:583) = 0;

% plot result
figure();
subplot(1, 3, 1)
imshow(I); axis image;
xlabel("x (px)"); ylabel("y (px)"); colorbar;
subplot(1, 3, 2)
imshow(AI); axis image;
xlabel("x (px)"); ylabel("y (px)"); colorbar;
subplot(1, 3, 3); imshow(AI1); axis image;
xlabel("x (px)"); ylabel("y (px)"); colorbar;

%%% gaussian low pass filter
% use matlab built-in 2d gaussian filter
sigma = 0.85;
Alfilt = imgaussfilt(AI1, sigma);

% plot result
figure();
subplot(1, 3, 1)
imshow(I); axis image;
xlabel("x (px)"); ylabel("y (px)"); colorbar;

```

```

subplot(1, 3, 2)
imshow(AI); axis image;
xlabel("x (px)"); ylabel("y (px)"); colorbar;
subplot(1, 3, 3); imshow(AIfilt); axis image;
xlabel("x (px)"); ylabel("y (px)"); colorbar;

% plot FFTs
figure()
hold on
F_AIfilt = fftshift(abs(fft(AIfilt(245,:))))/max(abs(fft(AIfilt(245,:))),[],"all");
plot(uI*dxI, 20.*log10(F_I));
plot(uAI*dxAI, 20.*log10(F_AIfilt));
plot(uAI*dxAI, 20.*log10(F_AI));
grid on;
hold off
xlim([-0.5 0.5]); ylim([-80 0]);
xlabel("Spatial Frequency in x (u, cycles/cm)"); ylabel("Amplitude (dB)");
legend(["Real Image" "Filtered AI Image" "Original AI Image"]);

% plot 2D ffts
F2_AIfilt = fftshift(abs(fft2(AIfilt)))/max(abs(fft2(AIfilt)),[],"all");
figure()
subplot(1, 3, 1); imagesc(uI*dxI, vI*dyI, 20*log10(F2_I)); axis image; colormap gray; c
 xlabel("Spatial Frequency in x * Pixel Size (u*dx)"); ylabel("Spatial Frequency in y * P
 subplot(1, 3, 2); imagesc(uAI*dxAI, vAI*dyAI, 20*log10(F2_AI)); axis image; colormap gra
 xlabel("Spatial Frequency in x * Pixel Size (u*dx)"); ylabel("Spatial Frequency in y * P
 subplot(1, 3, 3); imagesc(uAI*dxAI, vAI*dyAI, 20*log10(F2_AIfilt)); axis image; colormap
 xlabel("Spatial Frequency in x * Pixel Size (u*dx)"); ylabel("Spatial Frequency in y * P

% plot zoomed in version histogram
figure()
hold on
histAI = histogram(AI, 256, "Normalization","probability");
histI = histogram(I, 256, "Normalization","probability");
histAIfilt = histogram(AIfilt, 256, "Normalization","probability");
hold off
legend(["AI Image" "Real Image" "Filtered AI Image"])
xlim([0 1]); grid on; ylim([0 0.05]);

```

```
colormap(gray);
c = colorbar('southoutside'); c.TickLabels = [] ; c.Ticks = [] ;
```