PRACTITIONER PROTOCOL

LED Light Therapy

Your device maker's directions for distance, time and settings always come first. Times are typical.

CONSULT
Intake and consent signed. Screen for seizures, light-sensitizing meds, eye conditions, pregnancy and cancer.

CHOOSE THE LIGHT
Blue: surface, for breakouts. Red: calming, tone. Near-infrared (invisible): deepest. Combine per device.

PLAN THE SERIES
Typically [1–3] sessions a week for [4–6] weeks, then every [2–4] weeks. Use the device's cleared settings.

SET UP FIRST ☐ Opaque eye shields for client ☐ Goggles for you ☐ Panel cleaned ☐ Cords and arm checked ☐ Timer

#	STEP	TIME	HOW
1	**Review & inspect**	2–3 min	Ask about new medicines and supplements since last visit. Stop for open lesions, active cold sore or sunburn.
2	**Cleanse**	2–3 min	Remove all makeup and SPF so nothing blocks or reflects the light. Pat dry.
3	**Prep (optional)**	Varies	Any facial steps before the light. No acids, retinoids or fragranced products under the light unless the maker allows them.
4	**Protect eyes**	1 min	Opaque shields on the client, eyes closed. Put your own goggles on before switching on.
5	**Position**	1 min	Set the panel at the maker's distance, evenly over the area. Don't touch the skin unless designed to.
6	**Treat**	Per device	Usually [10–20] min. Use the preset program. Don't stack extra cycles to speed results.
7	**Monitor**	Throughout	Stay in the room. Check comfort and warmth. Stop for heat, dizziness, headache or any seizure sign.
8	**Finish**	2 min	Switch off before removing eye shields. Check the skin for redness.
9	**Calm & protect**	2 min	Hydrating serum and moisturizer, then broad-spectrum SPF 30+.
10	**Clean & rebook**	3 min	Disinfect panel and shields per maker. Give aftercare. Book the next session in the series.

IF SOMETHING GOES WRONG

· Too warm or red: stop, cool compress, note it.
· Seizure: switch off, keep the client safe on their side, call 911.
· Eye pain or vision change: stop; refer to an eye doctor.
· Rash or hives: stop. Call 911 if breathing is affected.
· Always write an incident report and check in with the client within 24–48 hours.

RECORD FOR THIS TREATMENT

Device · program	Light color(s)	Distance · minutes	Session # of series
Eye protection used	Meds changed? Y / N	Skin after · redness 0–3	Reaction · initials