PRACTITIONER PROTOCOL

Microcurrent Facial

Your device maker's directions for settings, tips and timing always come first. Times are typical.

HARD STOPS
Pacemaker or implanted device, epilepsy, pregnancy, active cancer. Metal in the face: follow device maker.

INJECTABLES
Wait [__] weeks after neurotoxin and [__] weeks after filler, or as the client's injector advises.

SERIES
Typically [1–2] a week for [6–12] sessions, then maintenance every [2–4] weeks.

SET UP FIRST ☐ Conductive gel (plenty) ☐ Clean probes or tips ☐ Device charged, cords checked ☐ Jewelry off ☐ Mirror for before/after

#	STEP	TIME	HOW
1	**Review & photo**	3 min	Confirm no new implants, injectables or pregnancy. Take a before photo if consented.
2	**Cleanse**	2–3 min	Remove all makeup and oils, which block current. Light exfoliation only if planned.
3	**Apply gel**	1 min	Even, generous layer on one section at a time. Re-wet as it dries; dry skin stings.
4	**Set & test**	1 min	Start at the lowest setting. Test on the jaw and adjust to comfort; more isn't better.
5	**Work the face**	20–40 min	Both probes on the skin. Glide or hold per maker, moving upward and outward, one side then match the other.
6	**Keep clear of**	Throughout	Eyes and eyelids, temples, front of the neck and thyroid, over the heart. Stay off broken skin.
7	**Monitor**	Throughout	Ask about tingling, taste or flashes. Lower the setting or add gel as needed.
8	**Remove & finish**	3 min	Wipe off gel. Serum, moisturizer, broad-spectrum SPF. After photo if consented.
9	**Clean & rebook**	3 min	Clean and disinfect probes per maker. Give aftercare. Book the next session in the series.

IF SOMETHING GOES WRONG

· Stinging or a hot spot: lift probes, add gel, lower setting.
· Dizziness, faintness or palpitations: stop, recline, monitor.
· Seizure: stop, keep the client safe on their side, call 911.
· Rash or welts from gel: remove, cool compress, note product.
· Always write an incident report and check in with the client within 24–48 hours.

RECORD FOR THIS TREATMENT

Device · program	Intensity setting(s)	Areas · minutes	Session # of series
Gel used	Last injectable · date	Photos taken Y / N	Reaction · initials