PRACTITIONER PROTOCOL

Hydrodermabrasion Facial

Your device manufacturer's settings, solutions and training always come first. Times are typical.

CONSULT
Intake and consent signed. Device use within your state's esthetics scope and your training.

CHOOSE SOLUTIONS
Skip the acid step for salicylate allergy, sensitivity, rosacea flare or recent retinoid; use a gentle solution.

TIP HYGIENE
New single-use tip per client, or clean and disinfect reusable parts exactly per the manual. Never reuse disposables.

SET UP FIRST ☐ Gloves ☐ New tips ☐ Solutions in date ☐ Waste jar empty ☐ Suction tested on forearm ☐ Eye pads

#	STEP	TIME	HOW
1	**Review & inspect**	3 min	Confirm changes. Look under magnification. Stop for cold sore, open lesions, sunburn or cystic acne.
2	**Cleanse**	2 min	Remove makeup and SPF. Pat dry. Eye pads on.
3	**Cleanse & exfoliate**	5–7 min	Exfoliating tip with cleansing solution. Short overlapping passes, skin held taut, one to two passes per zone.
4	**Acid step**	3–5 min	Mild AHA/BHA solution per manufacturer. Watch for even pinkness; stop at hot spots or blanching.
5	**Vortex extraction**	5–7 min	Suction with extraction solution. Keep the tip moving; never park it. Lowest setting over capillaries, thin or bony areas.
6	**Infuse**	5 min	Hydrating or antioxidant serum through the infusion tip.
7	**LED, optional**	Per device	Blue for breakouts, red for calming. Opaque goggles.
8	**Finish**	2 min	Moisturizer and mineral SPF 30+. Show the client the waste jar if they'd like.
9	**Clean down**	5 min	Discard tips. Empty and clean the waste jar. Flush lines and disinfect handpiece per the manual.

IF SOMETHING GOES WRONG

· Burning: stop the acid, rinse with cool water, cool compress.

· Suction mark or bruise: lower suction, skip the area.

· Hives or swelling: remove product. Call 911 if breathing is affected.

· Solution in the eye: flush with saline; see a doctor if irritation lasts.

· Device fault: stop, log it, and contact the manufacturer. Always write an incident report.

RECORD FOR THIS TREATMENT

Device · tip type	Suction setting	Acid solution · lot	Infusion serum
LED · color · minutes	Skin after · redness 0–3	Tips replaced · Y/N	Reaction · initials