PRACTITIONER PROTOCOL

# LED Light Scalp Therapy

Follow your device maker's settings, distance and times. Check your state board rules.

**SCREEN**

Intake cleared. No light-triggered epilepsy or migraine, lupus or light sensitivity.

**CHECK**

Medicines: some antibiotics, retinoids, St John's wort. Suspicious spot: refer to a dermatologist.

**PRE-SERVICE**

Clean scalp, no heavy oils or styling product. Pregnancy per [your policy].

**SET UP FIRST** ☐ Hands washed ☐ Device and cords checked ☐ Panel or cap disinfected ☐ Goggles for client and you ☐ Wavelength and timer set ☐ Fresh towel, neck support

| # | STEP | TIME | HOW |
|---|---|---|---|
| 1 | **Consult & screen** | 5 min | Confirm health changes and every new medicine or supplement. |
| 2 | **Consent & photos** | 3 min | Consent signed. Before photos of the same zones. Note session # in series. |
| 3 | **Scalp check** | 3 min | Part hair in good light. Look for sores, sunburn or a new or changing spot. |
| 4 | **Cleanse** | [ ] min | Clean, product-free scalp. Towel or blow dry per your license. |
| 5 | **Part & position** | 2 min | Part hair in sections to expose the scalp. Neck supported and neutral. |
| 6 | **Eye protection** | 1 min | Goggles on the client before the light goes on. Yours too. |
| 7 | **Place device** | 1 min | Panel or cap at the maker's distance. Red ~630–660 nm, blue ~415 nm. |
| 8 | **Light session** | 10–20 min | Start the timer. Stay nearby and check comfort and warmth. |
| 9 | **Off & inspect** | 2 min | Light off before goggles come off. Check the scalp for redness. |
| 10 | **Finish** | 3 min | Light serum if planned. Sit client up slowly. |
| 11 | **Aftercare & rebook** | 5 min | Review sheet and card. Book the next session in the series. |

## ENDPOINT AND SAFETY

- Full timed session, client comfortable, no burning
- Never promise regrowth; results vary by person
- Eyes protected the whole time the light is on
- Disinfect panel, cap and goggles after every client

## IF SOMETHING GOES WRONG

- Too warm: increase distance or stop; cool compress.
- Rash or hives: stop, ask about new medicines.
- Device fault or flicker: unplug, tag out of use.
- Headache or eye strain: light off, rest, note it.
- Dizzy or unwell: light off, sit up slowly, water.
- Seizure: protect the head, time it, call 911.

**Reaction that spreads or doesn't settle: refer to a physician. Write an incident report.**

## RECORD FOR THIS SESSION

| Device · panel or cap | Wavelength (nm) | Setting · distance | Time (min) |
|---|---|---|---|
| | | | |
| **Session # of series** | **Medicines checked** | **Scalp before · after** | **Next visit · initials** |
| | | | |