

TECHNICAL SPECIFICATIONS FOR MULTIMETER & CLAMPON METERS

Eligibility Criteria for Participation [Qualification Requirement (QR)]

1. Participation in the Tender

- Original manufacturers of Instruments in India will be eligible in the tender.
- For any non-indigenous Manufacturer having either manufacturing or assembling unit in India in collaboration with any registered Indian firm, are also eligible to quote.
- Distributor or Dealer or Agent or Channel Partner will be eligible for submission of quotation.
- Bidder will be OEM/OES in nature.
- Instruments should be calibrated from any NABL accredited Laboratory. Calibration certificate should bear NABL logo in each page. Calibration results shall comply with the specifications.
- Bids will be accepted only after meeting the technical requirement as per specification and on site-demonstration.

2. Experience: Experience in the field of manufacturing of Testing Instruments for last 5 (five) years. (Documents to be submitted in support of claims)

3. Warranty:

- a) Manufacturer shall invariably give unconditional Warranty for minimum 05 years from the date of demonstration of the materials on delivery. The Warranty shall include calibration of all the Instruments including transportation during Warranty period. The terms and condition of the Warranty shall be attached along with the Technical Bid. The bidder shall quote separately for extended Warranty, if they wish. Regarding non indigenous Manufacturers having either manufacturing or assembling unit in India in collaboration with any registered Indian Company, the Warranty should be provided by the Parent Company.
- b) All the Instruments supplied against this specification shall be Warranted for a defect liability of at least 60 months from the date of receipt at store duly passed acceptance testing at after delivery.
- c) Any Engineering error, omission, wrong provision, Instrument failure etc., the Instrument shall be attended by the bidder free of cost during the Warranty period
- d) For any defect in the kit during the Warranty period, repairing has to be completed within 15 days. Otherwise substitute set of same type has to be provided within 15 days from the date of reporting. The substitute set will be handed back to supplier on receipt of original set [duly repaired].
- e) If any defect observed during Warranty period, it is bidder's responsibility to collect kit from site /consignee to attend the same.
- f) On each failure of Test set, the supplier shall submit detailed report of failure analysis to Distribution Testing Department, WBSEDCL
- g) Necessary Life time licensed copy of interfacing and analysis software with test kit should be supplied free of cost with test Instrument.

4. Demonstration: During Technical scrutiny of the bid, the purchaser will ask to demonstrate the offered Instrument and requisite accessories at their own cost at any specified location given by the Purchaser. The Bidder shall have to demonstrate the test Instrument within 15 days of intimation from the purchaser. It is the responsibility of the bidder to bring all the Instruments with necessary software and accessories required for Demonstration of the features available in the test set as per Technical Specification at date and time communicated to them in advance. The bidders are also required to provide the operating manual /required literature and sent competent Engineers at that. The purchaser will evaluate the suitability of product of their application. The purchaser's decision will be final for acceptance/ non-acceptance of product.


13-06-15
12/6/15
13-6-15
13-6-15
13-6-15
13-6-15

5. **Performance Certificate:** Bidder shall submit at least two Performance certificate from any power utility/ Govt. Department for Instruments supplied during last 5 (five) years from date of Tender to be submitted with the Offer. The date of performance certificate should be at least after one year from the date of delivery. Documentary evidence for the supply & acceptance of quantity by the Consignee Authority to be submitted with the offer.

6. **Maintenance and support:** Bidders must give necessary service and spare part support for the entire life time of the Testing Instruments offered after completion of the Warranty period as and when required.

7. **On-site installation, commissioning and demonstration:**

The supplier shall be responsible for the on-site installation, commissioning, and demonstration of the electrical instruments at the delivery location specified by the purchaser.

The scope of the demonstration shall include:

- Unpacking and proper installation as per manufacturer's specifications.
- Electrical safety checks and power connection under supervision of the Purchaser's technical team.
- Demonstration of all operating functions and settings.
- Verification of performance as per Technical Specifications in the P.O.
- Basic training to end-users/maintenance personnel, including do's and don'ts, safety procedures, and routine maintenance tips.

The supplier must obtain a **Demonstration Completion Certificate**, signed by the purchaser's authorized representative.

No additional charges shall be claimed for demonstration, travel, or manpower.

OL
13.6.20

by
13/6

Signature 13/6/25

Signature
13.06.25

Signature
13/6/25

Signature

Signature
13/6/25

Technical Specification for Digital Multimeter:

Sl. No.	Specification	Description
1	Type	True RMS
2	Basic DC Accuracy	$\pm 0.2\%$
3	Measurement of DC Voltage, Accuracy	Upto 1000V, $\pm 0.5\%$ (Auto Range)
4	Measurement of AC Voltage, Accuracy	upto 1000V, $\pm 1.0\%$ (Auto Range)
5	Measurement of DC Current, Accuracy	Upto 10 A, $\pm 1.0\%$ (Auto Range)
6	Measurement of AC Current, Accuracy	Upto 10 A, $\pm 1.5\%$ (Auto Range)
7	Measurement of Resistance, Accuracy	0 Ω -50M Ω , $\pm 0.5\%$ (Auto Range)
8	Measurement of Frequency, Accuracy	Upto 99 kHz, $\pm 1.5\%$
9	Measurement of Capacitance, Accuracy	Upto 1000 μ F, $\pm 2\%$ (Auto Range)
10	Diode/Continuity check	Yes
11	Update Rate	3 times /sec
12	Battery	9 V/AA/ AAA Dry Cell.
13	Display	LCD & 3 and $\frac{3}{4}$ digit display
14	Backlit	Yes
15	Probe compatibility	Yes [02 set]
16	Safety Category	CAT IV - 600V
17	Transient Protection	8kV
19	Overload Protection Voltage range	1050V DC / AC
20	Overload/Fuse Protection	Yes
21	Minimum 1m Test Lead (coloured and paired) should be provided along with additional spare Test lead	Yes


13.07.25


21/6/25




13.6.25


13/6/25


12/6


31/06/25

Technical specification for Clamp on Meter: For Measurement of Leakage Current & Load Current

Sl. No.	Specification	Description	
1	Ranges	0-30mA/300mA	0-100A
2	Resolution	0.01/0.1mA	0.01/0.1A
3	Accuracy	$\pm 1\%$ of Reading	$\pm 2\%$ of Reading for 0-200A
4	Jaw Size	35 mm and above.	
5	System Voltage Limitation	At least up to 300V	
6	Power Supply	Internally mounted Removable batteries	
7	Battery Life	Approximately 40 hours	
8	Low battery & Over Range Indication	Yes	
9	Auto Power Off	After 10 min	
10	Withstanding Voltage	Min 3kV	
11	Protection	Overload Protection	
12	Data Hold Indication	Yes	
13	Auto/Manual Ranging	Both	
14	Display	LCD and 3 and $\frac{1}{2}$ digit Display	
15	Operating Temp & humidity	0 °C to 40°C and RH < 80%	
16	Weight	Light Weight	

[Signature]
13.06.24

[Signature]
12/6/24

[Signature]

[Signature]
13.6.24


[Signature]
13/6/24

[Signature]
13/6

[Signature]
13/6/24

Technical Specification for Clamp on Meter: For Measurement of Load Current

Sl. No.	Specification	Description
1	Measuring Function	Load Current
2	Display	3.5 digit LCD/ LED
3	Ranges	0-999A
4	Min. Resolution	0.01 / 0.1A
5	Accuracy	$\pm 2.0\%$ of Reading for 0-999A
6	AC Voltage Range	0-600V
7	Backlit Display	Yes
8	Maximum Reading	999
9	Safety rating	CAT III 600 V
10	Data Hold Indication	Yes
11	Operating temperature	0 - 40°C $\leq 90\%$ RH
13	Jaw Size	50mm and above
14.	Lead	2 set 1 Mtr. each
15.	Probe Compatibility	Yes


13-06-11



13/6/11




12/6/11


13-6-11


13/6


13/06/11