

BSN medical GmbH, P.O. Box 570239, 22771 Hamburg

Certificate of Analysis

Printdate
2021/05/04
Purchase order item/date

Delivery item/date

Order item/date

Customer number

Material: 72646-00003-04
CUTIMED SORBACT HYDROACTIVE STERILE 7X8.5CM 10 NL EN FR DE IT ES

Batch: 939469 Expiry Date: 2022/08/01
Manufacturing Date: 2019/08/17

Legal Manufacturer: BSN medical GmbH HH
Country of Origin: United Kingdom
CE-Class: IIb

Characteristic	Result	Requirement
Sterilization	Pass	Gamma-steril. ISO 11137-1:2015
Bioburden	Pass	0 - 10 CFU/g
Appearance	Pass	Transparent gel w. green mesh
Miscellaneous	Pass	Gel coat weight: 1,7-2,1 kg/m2 Gel saline uptake: 1000 -2000%

Analysis of the above mentioned product demonstrated conformance with the valid Product Specification agreed on with supplier.

This analysis is performed in the course of activities for monitoring and measurement of products according to EN ISO 13485 and and supplements the quality measures defined in the Quality Management Agreement.

This document is valid without signature.

BSN medical GmbH
Quality Management

