

PROSTHETIC BIONIC PALM

A PROJECT REPORT

Submitted by

ANURAG C NAIR (ASI22EBE008)

AUFFIN RASHEED (ASI22EBE010)

EBY MATHEW (ASI22EBE012)

to

the APJ Abdul Kalam Technological University

in partial fulfillment of the requirements for the award of the Degree

of

Bachelor of Technology

in

Electronics and Biomedical engineering



Department of Electronics and Biomedical Engineering

**ADISHANKARA INSTITUTE OF ENGINEERING AND
TECHNOLOGY**

KALADY, ERNAKULAM - 683574

DECLARATION

We, undersigned, hereby declare that the project report “**Prosthetic Bionic palm**”, submitted for partial fulfillment of the requirements for the award of degree of Bachelor of Technology of the APJ Abdul Kalam Technological University, Kerala is a Bonafide work done by us under the guidance of **Dr.Lakshmi M Hari**, Department of Electronics and biomedical engineering. This submission represents our ideas in our own words and where ideas or words of others have been included; We have adequately and accurately cited and referenced the original sources. We also declare that we have adhered to the ethics of academic honesty and integrity and have not misrepresented or fabricated any data or idea or fact or source in our submission. We understand that any violation of the above will be a cause for disciplinary action by the institute and/or the University and can also evoke penal action from the sources which have thus not been properly cited or from whom proper permission has not been obtained. This report has not been previously formed the basis for the award of any degree, diploma or similar title of any other University.

**ANURAG C NAIR
AUFFIN RASHEED
EBY MATHEW**

**KALADY
05-04-2025**

**DEPARTMENT OF ELECTRONICS &
BIOMEDICAL ENGINEERING
ADISHANKARA INSTITUTE OF ENGINEERING & TECHNOLOGY**



BONAFIDE CERTIFICATE

This is to certify that the project report entitled **PROSTHETIC BIONIC PALM** submitted by **ANURAG C NAIR(ASI22EBE008), AUFFIN RASHEED(ASI22EBE010), EBY MATHEW (ASI22EBE012)** to the APJ Abdul Kalam Technological University in partial fulfillment of the requirements for the award of the Degree of Bachelor of Technology in Electronics and Biomedical Engineering is a Bonafide record of the project work carried out by him under our guidance and supervision. This report in any form has not been submitted to any other University or Institute for any other purpose.

GUIDE

Dr Lakshmi M Hari
Assistant Professor

CO-GUIDE

Mr. Shinu MM
Assistant Professor

PROJECT-COORDINATOR

Ms. Krishna S Nair
Assistant Professor

HEAD OF THE DEPARTMENT

Dr. REMYA GEORGE
Associate Professor

Kalady

05-04-2025

ACKNOWLEDGMENT

We would like to express our immense gratitude and profound thanks to all those who helped us to make this project a great success. We express our gratitude to the almighty God for all the blessings endowed on us.

We express our sincere thanks to our Principal **Dr. Murali M.S** for providing us with such a great opportunity.

We also convey our gratitude to our Head of the Department **Dr. Remya George** for having given us a constant inspiration and suggestion. We extend our deep sense of gratitude to our project coordinator **Ms.Krishna S Nair**, Assistant Professor of Electronics and Biomedical Engineering Department for providing enlightening guidance through the project. We can hardly find words to express our deep appreciation for the help and warm encouragement that we have received from our project guide **Dr Lakshmi M Hari**, Assistant Professor of Electronics & Biomedical Engineering Department for his whole-hearted support.

It was their encouragement that helped us to complete the project. We can hardly find words to express our deep appreciation of the help and warm encouragement that we received from our parents. We are extremely thankful and indebted to our friends who supported us in all aspects of the project work.

ANURAG C NAIR

AUFFIN RASHEED

& EBY MATHEW

INSTITUTIONAL VISION

To emerge as a Center of Excellence in Engineering, Technology and Management by imparting quality education, focusing on empowerment and innovation.

INSTITUTIONAL MISSION

Impart quality professional education for total upliftment of the society.

Create congenial academic ambience that kindles innovative thinking and research. Mold competent professionals who are socially committed and responsible citizens.

QUALITY POLICY

We are committed to the total upliftment of the society by imparting quality professional education.

We aim at modeling totally competent professionals with ingenuity, adaptability, social commitment and ethical and spiritual values by creating a congenial academic ambience that kindles innovative thinking.

We continually upgrade the Quality Management System through empowerment and involvement.

DEPARTMENTAL VISION

Evolve as a premier center in Electronics and Biomedical Engineering to meet the ever-increasing needs for affordable and accessible health care technology.

DEPARTMENT MISSION

- Provide quality professional education at par with global standards in the field of Biomedical Engineering with excellent faculty and infrastructure.
- Foster a culture of multidisciplinary research, comprehensive practical learning, and generating pioneering innovations by collaborating with industrial and clinical experts.
- Inspire biomedical engineers' graduates to be responsible to address critical healthcare challenges with empathy and dedication for the betterment of mankind.

PROGRAMME EDUCATIONAL OBJECTIVES (PEOS)

PEO1	Function as creative professionals who excel in conducting re- search, designing, manufacturing, and testing of biomedical devices with an unwavering focus on quality and patient-safety.
PEO2	Engage in continuous learning to overcome the ever-evolving challenges of biomedical technology to support healthcare industries, allied healthcare sectors, and academia.
PEO3	Evolve as successful engineers, entrepreneurs and healthcare technology leaders through professional development and collaborations.

PROGRAMME SPECIFIC OBJECTIVES (PSOS)

PSO1	To nurture students with technically inquisitive attitude so that any real-world problem could be tackled with a problem-solving perspective, finding a suitable mathematical model with strong fundamental technological concepts to solve and apply to rapid growing arena of computer technology.
PSO2	To develop professionals with excellent exposure to the latest technologies to design high quality products unique in innovation, technology, software, security, hardware and usefulness; making high impact on society, business and technology.
PSO3	To enhance knowledge in practical implementation of technology with regard to parallelism, virtualization of networks, scientific analysis and modeling, visualization, natural language processing, digital synthesis of data and its manipulation, wireless and mobile communication, storage and retrieval of huge amount of data etc.

PROGRAMME OUTCOMES (POS)

PO1	Engineering knowledge: Apply the knowledge of mathematics, science, engineering fundamentals, and an engineering specialization to the solution of complex engineering problems.
PO2	Problem analysis: Identify, formulate, review research literature, and analyze complex engineering problems reaching substantiated conclusions using first principles of mathematics, natural sciences, and engineering sciences.

PO3	Design/development of solutions: Design solutions for complex engineering problems and design system components or processes that meet the specified needs with appropriate consideration for the public health and safety, and the cultural, societal, and environmental considerations.
PO4	Conduct investigations of complex problems: Use research-based knowledge and research methods including design of experiments, analysis and interpretation of data, and synthesis of the information to provide valid conclusions.
PO5	Modern tool usage: Create, select, and apply appropriate techniques, resources, and modern engineering and IT tools including prediction and modeling to complex engineering activities with an understanding of the limitations.
PO6	The engineer and society: Apply reasoning informed by the contextual knowledge to assess societal, health, safety, legal and cultural issues and the consequent responsibilities relevant to the professional engineering practice.
PO7	Environment and sustainability: Understand the impact of the professional engineering solutions in societal and environmental contexts, and demonstrate the knowledge of, and need for sustainable development.
PO8	Ethics: Apply ethical principles and commit to professional ethics and responsibilities and norms of the engineering practice.
PO9	Individual and team work: Function effectively as an individual, and as a member or leader in diverse teams, and in multidisciplinary settings.
PO10	Communication: Communicate effectively on complex engineering activities with the engineering community and with society at large, such as, being able to comprehend and write effective reports and design documentation, make effective presentations, and give and receive clear instructions.
PO11	Project management and finance: Demonstrate knowledge and understanding of the engineering and management principles and apply these to one's own work, as a member and leader in a team, to manage projects and in multidisciplinary environments.

PO12	Life-long learning: Recognize the need for, and have the preparation and ability to engage in independent and life-long learning in the broadest context of technological change.
------	---

COURSE OBJECTIVES

To develop skills in doing literature survey, technical presentation and report preparation.
--

COURSE OUTCOMES

The student will be able to

CO1	Envisage applications for societal needs.
CO2	Develop skills for analysis and synthesis of practical systems.
CO3	Learn to use new tools effectively and creatively.
CO4	Learns to carry out analysis and cost-effective, environmental friendly designs of engineering systems.
CO5	Develops the ability to write Technical / Project reports and oral presentation of the work done to an audience.

ABSTRACT

The field of prosthetics has seen remarkable advancements in recent years, driven by developments in biomechanics, materials science, and robotics. Losing a hand or part of it due to injury, illness, or congenital conditions can drastically affect an individual's quality of life. Advanced bionic prosthetic devices remain expensive, which makes them inaccessible to many who could benefit from them, particularly in low- income regions or developing countries. This project presents the development of a bionic prosthetic palm integrated with an ESP-32 microcontroller, Electromyography (EMG) sensors, and a temperature sensor to enhance functionality and adapt- ability. The EMG sensor detects muscle signals from the residual limb, allowing the user to control the prosthetic palm with natural muscle contractions. The ESP- 32 microcontroller, with its high processing capability and wireless communication features, processes the EMG signals and translates them into finger and wrist movements. Additionally, a temperature sensor is incorporated to monitor environmental and body temperature, ensuring user comfort and detecting potential risks such as overheating or excessive cold exposure. The system is designed to be lightweight, energy-efficient, and cost-effective, making it accessible to a broader population.

CONTENTS

ACKNOWLEDGMENT	i
INSTITUTIONAL VISION, MISSION AND QUALITY POLICY . .	ii
DEPARTMENTAL VISION, MISSION, PEOs ,PO AND PSOs	iii
ABSTRACT	vii
LIST OF FIGURES	x
LIST OF ABBREVIATIONS	xi
1 INTRODUCTION	1
1.1 GENERAL BACKGROUND	2
1.2 PROBLEM DEFINITION	2
1.3 MOTIVATION	3
1.4 OBJECTIVES	3
2 LITERATURE SURVEY	4
3 METHODOLOGY	5
3.1 PROPOSED SYSTEM	5
3.2 HARDWARE REQUIREMENTS	5
3.3 SOFTWARE REQUIREMENTS	8
4 DESIGN AND METHODOLOGY	10
4.1 ARCHITECTURE DIAGRAM	10
4.2 PROCESS FLOW DIAGRAM	11
4.3 FUNCTIONAL BLOCK DIAGRAM	13
5 RESULTS AND DISCUSSIONS	14
5.1 HARDWARE DEVELOPED	14
5.2 SOFTWARE DEVELOPED	15
6 CONCLUSION AND FUTURE SCOPE	17
6.1 FUTURE SCOPE	17
6.2 CONCLUSIONS	17
APPENDICES	xiv

I	Paper Publications and Achievements	xiv
II	Code	xxiii

LIST OF FIGURES

3.1	ESP32 Microcontroller	6
3.2	EMG sensor with electrodes	6
3.3	Servo Motor	7
3.4	NTC Thermocouple temperature sensor	7
3.5	OLED Display	7
3.6	9V Battery	8
4.1	Architecture diagram.....	10
4.2	Process flow diagram	12
4.3	Block diagram	13
5.1	Final Assembly.....	16
5.2	Temperature sensing.....	16

LIST OF ABBREVIATIONS

ABBREVIATION	DESCRIPTION
EMG	Electro Myography
NTC	Negative Temperature Coefficient
ESP	Expressive systems
ADC	Analog to Digital Converter
IoT	Internet of Things

CHAPTER 1

INTRODUCTION

The need for affordable and highly functional prosthetic devices has become increasingly evident, as many existing solutions are either prohibitively expensive or fail to provide a natural and intuitive user experience. Traditional prosthetic hands often lack the dexterity and responsiveness required for seamless daily functionality, limiting their effectiveness in restoring lost capabilities. As a result, millions of amputees worldwide struggle with prosthetic solutions that do not adequately meet their needs.

These challenges arise due to the absence of cost-effective yet technologically advanced solutions in the prosthetics industry. The development of prosthetic hands involves intricate engineering, integrating muscle signal processing and sensory feedback mechanisms. However, conventional approaches often compromise either affordability or functionality. Striking the right balance between these factors is essential to ensure that prosthetic solutions are both accessible and effective for users.

To address these issues, our project focuses on the development of a bionic palm that leverages advanced electronics alongside cost-efficient 3D printing technology. The proposed prosthetic hand incorporates an ESP32 microcontroller to process electromyography (EMG) signals from the Carpi Radialis and Carpi Ulnaris muscles, enabling precise and responsive control of hand movements. Furthermore, embedded tactile and temperature sensors in the fingertips provide critical sensory feedback, enhancing the user's interaction with their environment. The integration of 3D printing allows for customization and affordability, ensuring that the device is accessible to a wider range of users.

By combining real-time muscle signal processing with enhanced sensory capabilities, our project seeks to bridge the gap between cost and functionality in prosthetic technology. The proposed bionic palm not only improves dexterity and control but also fosters a more natural and immersive experience for amputees. This innovative approach has the potential to redefine prosthetic solutions, making them more inclusive, functional, and user-friendly.

1.1 GENERAL BACKGROUND

Bionic prosthetic palms have significantly advanced, particularly through the integration of EMG signals for intuitive control. EMG measures the electrical activity produced by skeletal muscles, enabling prosthetic devices to interpret a user's muscle movements and translate them into corresponding hand actions. This technology allows for more natural and responsive prosthetic control, enhancing the user's ability to perform daily tasks. [1] The system comprises both mechanical and electronic components, integrating artificial intelligence to process EMG signals. The researchers developed an experimental methodology to record muscle activity associated with specific tasks using three surface EMG sensors. By leveraging 3D printing technologies, such prostheses can be produced more affordably and customized to individual users, making advanced prosthetic solutions more accessible. [2]

1.2 PROBLEM DEFINITION

EMG sensors are used to capture electrical signals generated by muscle contractions, but these signals are inherently weak, typically in the microvolt (V) range, making them challenging to work with. The signals are easily susceptible to noise and interference from various sources, such as power lines (50/60 Hz), electromagnetic interference from nearby electronics, and static electricity. These sources can distort the EMG signal, making it difficult to separate the useful muscle activity from external noise. Additionally, the interface between the electrodes and the skin introduces variability in the signal quality, as skin resistance can fluctuate due to moisture, cleanliness, or even electrode placement. Higher skin impedance can degrade signal quality, leading to weak or unreliable readings. Muscle fatigue and the involvement of multiple muscles during a contraction can further complicate the signal, as these factors introduce variability and can cause signal degradation over time. Proper amplification of these weak signals is necessary, but it can also amplify the noise, making the filtering process crucial. However, overly aggressive filtering can lead to the loss of important signal characteristics. Moreover, different users have distinct muscle patterns, which means that a system must adapt to individual variations, adding another layer of complexity. To address these challenges, basic filtering and smoothing techniques can be applied to EMG signals for simpler and more direct control of prosthetics.

1.3 MOTIVATION

Current prosthetic options often lack the necessary adaptability and ease of use which leads to underutilization. The project emphasizes the integration of advanced materials and 3D printing technology, thereby making the arm not only cost-effective but also customizable to meet individual needs and preferences. Utilizing EMG allows for a more intuitive interface that enables users to control the arm and mimic natural movement patterns.

1.4 OBJECTIVES

The objectives of this work are:

- Develop an affordable Prosthetic Solution using 3D printing technology.
- Implement EMG Signal Control to accurately interpret muscle signals from the forearm.
- Ensure a lightweight, ergonomic design for daily use.
- Allow adjustments to fit individual user needs and preferences.
- Help users regain independence and confidence.

CHAPTER 2

LITERATURE SURVEY

The paper "*A 3D Printed, Bionic Hand Powered by EMG Signals*" by Karla Avilés-Mendoza, Neil George Gaibor-León, and Víctor Asanza presents a bionic hand controlled by electromyography (EMG) signals generated by muscle contractions. Using 3D printing, the prosthetic is customizable, affordable, and ergonomic, enabling intuitive control and mimicking natural hand movements. The research highlights the potential of this technology to improve prosthetics' functionality and accessibility while addressing challenges like signal accuracy and system reliability. Future directions include enhancing human-prosthetic interfaces and improving materials. [3]

The study "*Comparison of Electromyography and Force as Interfaces for Prosthetic Control*" by Corbett, Perreault, and Kuiken compares EMG-based and force-based interfaces for controlling prosthetics. While EMG measures electrical activity indirectly related to muscle force, force interfaces directly measure exerted force. Both performed similarly in position-tracking tasks, with EMG excelling at higher tracking frequencies. These findings inform the design of prosthetics that align with user needs. [4]

The study "*Clinical Implementation of a Bionic Hand Controlled with Kinetic myographic Signals*" by Cipriani et al. introduces a novel approach to bionic hand control using kinetic myographic (KMG) signals generated by implanted magnetic tags in forearm muscles. Through tendon transfer surgery, magnets are embedded in synergistic muscles responsible for hand movements, creating magnetic signals that are processed by neural algorithms to control the prosthetic hand. This method bypasses traditional magnet localization, directly mapping muscle contractions to hand gestures. Clinical trials demonstrated the feasibility of this system, enabling intuitive control and gesture recognition. The research highlights advancements in magnetic-based prostheses and suggests potential for improved functionality and user experience in bionic hands. [5]

CHAPTER 3

METHODOLOGY

3.1 PROPOSED SYSTEM

The prosthetic bionic arm system is designed to provide real-time control using electromyography (EMG) signals processed by an ESP32 microcontroller. The system integrates multiple hardware components, including EMG sensors, servo motors, an OLED display, and additional sensors such as an NTC thermocouple sensor for enhanced functionality. The design aims to ensure precise control of finger movements, real-time feedback, and safe operation for the user. The system begins with data acquisition from the EMG sensor, which detects muscle activity through electrodes placed on the user's forearm. These signals are amplified using the onboard AD8226 instrumentation amplifier and transmitted to the ESP32 microcontroller. The ESP32 processes the raw EMG signals, applies filtering techniques to eliminate noise, and interprets the data to generate PWM signals that control three servo motors responsible for finger movements. The servos operate based on predefined thresholds, enabling accurate and proportional control of the prosthetic fingers. An NTC thermocouple sensor is integrated to monitor temperature variations, ensuring safe operation of the system by preventing overheating. The OLED display provides real-time visualization of EMG signal strength and motor states, offering intuitive feedback to the user. Power is supplied through two 9V batteries, ensuring stable voltage delivery to all components. The ESP32 also facilitates wireless communication for remote monitoring and control.

3.2 HARDWARE REQUIREMENTS

ESP32 Microcontroller

The ESP32 microcontroller serves as the central processing unit for this system. It features a dual-core processor with integrated Wi-Fi and Bluetooth capabilities, making it suitable for IoT applications. The ESP32 processes EMG signals and controls actuators using PWM outputs while managing wireless communication for remote monitoring.



Fig. 3.1: ESP32 Microcontroller

EMG Sensor Functionality

The EMG sensor plays a critical role in capturing muscle activity from the user's forearm. It measures electrical impulses generated during muscle contractions and outputs analog signals that are amplified and processed by the AD8226 amplifier. These signals are then sent to the ESP32 microcontroller for further processing. The EMG data is filtered to remove noise and interpreted using predefined thresholds to control servo motors proportionally. This ensures smooth and responsive finger movements based on muscle activation.

Fig. 3.2: EMG sensor with electrodes



Servo Motors

Three servo motors are used in this system to control the individual finger movements of the prosthetic arm. These motors receive PWM signals generated by the ESP32 based on processed EMG data. The servos allow precise control of finger positions, enabling tasks such as grasping or releasing objects. Their compact design makes them ideal for integration into a wearable prosthetic device.



Fig. 3.3: Servo Motor

NTC Thermocouple Sensor

The NTC thermocouple sensor monitors temperature variations within the system or the surrounding environment. This sensor ensures safe operation by detecting overheating conditions that could damage components or cause discomfort to the user. It provides analog temperature readings that are processed by the ESP32 for safety alerts.



Fig. 3.4: NTC Thermocouple temperature sensor

OLED Display

The OLED display offers a user-friendly interface by showing real-time updates on EMG signal strength, servo motor states, and temperature readings from the NTC sensor.

Fig. 3.5: OLED Display



Power Management

Power is supplied through two 9V batteries connected in parallel to provide stable voltage delivery to all components of the system. Efficient energy utilization ensures prolonged operation without frequent replacements or recharging.



Fig. 3.6: 9V Battery

3.3 SOFTWARE REQUIREMENTS

The software requirements for the prosthetic bionic arm project encompass various tools, programming languages, libraries, and development environments essential for interfacing, signal processing, debugging, and real-time control of hardware components. The primary programming language used in this project is C++, specifically within the Arduino framework, which provides a structured yet flexible approach for programming the ESP32 microcontroller and managing connected sensors and actuators. The Arduino IDE serves as the main development platform for writing, compiling, and uploading code to the ESP32. It offers a user-friendly interface and built-in libraries that simplify sensor communication, data processing, and motor control.

To ensure the seamless integration of multiple components, the project relies on various Arduino libraries. The Wire library is utilized for I2C communication with peripherals such as the OLED display. The Adafruit GFX and SSD1306 libraries are employed for controlling the OLED screen, enabling real-time visualization of EMG signal strength and motor states. For EMG signal processing, customized algorithms are implemented to filter noise and extract meaningful data from muscle activity. These algorithms are designed to interpret EMG signals based on predefined thresholds, ensuring proportional control of servo motors.

The servo motors are controlled using PWM signals generated by the ESP32 microcontroller. Libraries such as Servo.h are used to manage precise motor movements required for finger articulation. Additionally, real-time temperature monitoring is facilitated by interfacing the NTC thermocouple sensor with analog-to-digital conversion routines in the ESP32. This ensures safe operation by detecting overheating conditions.

For debugging and real-time monitoring during development, the Serial Monitor in Arduino IDE is extensively used to print sensor values, actuator states, and anomalies in system behavior. This allows developers to identify issues quickly and optimize performance. Timing operations critical for signal synchronization are handled using non-blocking functions like `Millis()`, ensuring efficient multitasking within the system.

Beyond development, simulation tools such as Proteus or similar circuit design software may be utilized to validate circuit functionality before hardware implementation. The final deployment involves flashing the ESP32 with compiled firmware and testing compatibility with power supply modules and communication protocols.

CHAPTER 4

DESIGN AND METHODOLOGY

4.1 ARCHITECTURE DIAGRAM




Fig. 4.1: Architecture diagram

The Bionic Palm System is an advanced prosthetics controlled using bioelectrical signals and environmental data. The system comprises multiple components, including an EMG sensor, a temperature sensor, an ESP-32 microcontroller, a motor controller, a voltage regulator, and a battery. The EMG sensor detects electrical activity from the user's muscles and sends signals to the ESP-32, which interprets the data to generate movement commands. Additionally, the temperature sensor monitors environmental conditions, ensuring safe and adaptive operation.

The ESP-32 serves as the core processing unit, receiving inputs from both sensors, processing them, and sending control signals to the motor controller. The motor controller then translates these signals into appropriate electrical outputs to drive the Bionic Palm, enabling precise and natural hand movements. The system is powered by a battery, which supplies energy through a voltage regulator, ensuring stable power distribution to all components. This design enables real-time muscle-based control, making the bionic palm highly functional for prosthetic users or robotic applications.

4.2 PROCESS FLOW DIAGRAM

The process flow begins with the user's muscle contractions, which are detected by the EMG sensors that capture the electrical signals produced by the muscles. These signals are then amplified and converted from analog to digital form, ensuring they are suitable for processing. The ESP32 microcontroller takes over by processing both the EMG data and temperature data gathered from the temperature sensor, which monitors both the environmental and body temperatures. Within the ESP32, a condition check determines whether the temperature is safe or unsafe: if safe, the process continues to activate the actuators, which control the motors responsible for moving the prosthetic palm (fingers and wrist); if unsafe, an alert is generated or adjustments are made to ensure user safety and comfort. The system is supported by a power supply that feeds both the ESP32 and the actuators, and it continuously loops back to the beginning, maintaining real-time monitoring and control.




Fig. 4.2: Process flow diagram

4.1 FUNCTIONAL BLOCK DIAGRAM




Fig. 4.3: Block diagram

The prosthetic hand control system utilizes EMG signals and an ESP32 micro- controller to enable intuitive movement. EMG leads are placed on the Flexor Carpi Ulnaris and Flexor Carpi Radialis muscles to detect electrical activity generated by muscle contractions. These signals are then sent to a signal amplification and ADC module, where they are amplified and converted into digital data for processing. The ESP32 microcontroller interprets these signals to determine the corresponding hand movements. Additionally, a temperature sensor provides environmental or safety feedback to the microcontroller. The processed signals from the ESP32 control four servo actuators, each responsible for a specific part of the prosthetic hand. One servo motor moves the index finger, another controls the remaining fingers, a third operates the thumb, and the fourth enables wrist movement. A dedicated power supply ensures the proper functioning of the actuators, while another power source supports the ESP32 and signal processing unit. By converting muscle activity into precise motorized actions, the system allows users to control the prosthetic hand naturally. The integration of sensors (NTC temperature sensor) and microcontroller-based processing enhances responsiveness, providing a functional, user-friendly assistive solution.

CHAPTER 5

RESULTS AND DISCUSSIONS

The prosthetic bionic arm system embedded with EMG sensors, an NTC thermocouple sensor, and an OLED display, with the ESP32 microcontroller efficiently processing real-time data. When muscle activity was detected, the ESP32 triggered three servo motors to control finger and wrist movements, helping users perform tasks with precision. The system's data was displayed on the OLED screen, showing real-time EMG signal strength, actuator states, and temperature readings. Users could monitor system performance and receive feedback on muscle activity and motor states. The combined functionality of the hardware and software resulted in a responsive, wearable solution capable of controlling critical prosthetic arm parameters and providing intuitive feedback to users.

5.1 HARDWARE DEVELOPED

The hardware system integrated multiple components into a compact and wearable form. EMG sensors were used to capture muscle activity, converting electrical signals into digital data accurately read by the ESP32 microcontroller. The NTC thermocouple sensor provided real-time temperature monitoring, ensuring a comprehensive view of the device's conditions, especially important for user safety. The ESP32 microcontroller played a central role in managing sensor data, executing control algorithms, and controlling the servo motors. These motors acted as actuators, allowing precise finger movements based on muscle signals. The use of a dual-core ESP32 allowed simultaneous data acquisition and real-time feedback without noticeable latency, ensuring responsive and reliable performance.

To ensure portability and usability, a power management system was implemented using two 9V batteries. This setup ensured a stable power supply to the ESP32 and peripheral devices while allowing for extended operation. The compact layout made the system practical for integration into a prosthetic arm, demonstrating the feasibility of wearable prosthetic control without compromising user comfort.

Overall, the hardware system performed effectively in test scenarios. It responded promptly to changes in muscle activity, detected temperature variations, and activated precise motor movements. The compactness, energy efficiency, and wireless capability of the ESP32 made the hardware system suitable for continuous monitoring in real-world conditions.

5.2 SOFTWARE DEVELOPED

The software component of the prosthetic bionic arm was developed using the Arduino framework, providing a seamless bridge between the ESP32 hardware and the end user. A custom program was designed to visualize real-time EMG data and control servo motors based on muscle activity. Users could monitor system performance and receive feedback on muscle signals and motor states. A key functionality of the software was the implementation of a control system that translated muscle signals into proportional motor movements. When muscle activity exceeded predefined thresholds, the software triggered the servo motors to execute precise movements.

The software system not only offered real-time insights and control but also served as a digital interface for users. Its integration with the ESP32's capabilities enabled smooth and reliable data transfer. Together with the hardware, the software formed a robust, user-friendly prosthetic arm ecosystem aimed at intuitive control and feedback.




Fig. 5.1: Final Assembly




Fig. 5.2: Temperature sensing

CHAPTER 6

CONCLUSION AND FUTURE SCOPE

6.1 FUTURE SCOPE

The future development of the Bionic Palm System presents numerous opportunities for enhancement, focusing on advanced sensory integration, artificial intelligence, user-centric design, and cost optimization. Incorporating pressure and temperature sensors will enable real-time haptic feedback, improving grip precision and user interaction. Additionally, AI-driven machine learning algorithms will enhance responsiveness by analyzing muscle signals, predicting user intent, and adapting dynamically to different environments.[6] Future iterations will prioritize ergonomics, material optimization, and weight reduction, ensuring greater comfort and usability for extended wear. Customizable and modular designs will further enhance the user experience. Moreover, reducing production costs through innovative manufacturing techniques and scalable production strategies will make the prosthetic more affordable and accessible to a wider audience, improving the quality of life for individuals with limb loss. By integrating these advancements, the Bionic Palm will become a more intelligent, efficient, and user-friendly assistive device, seamlessly bridging the gap between human capability and technological innovation.

6.2 CONCLUSIONS

A bionic prosthetic palm based on EMG sensors and the ESP32 microcontroller offers a promising solution for enhancing prosthetic functionality. The integration of EMG sensors allows precise detection of muscle signals, enabling the prosthetic to respond to the user's intent with minimal latency, creating a near-natural experience. The ESP32 microcontroller processes these signals efficiently, ensuring real-time control and responsiveness while maintaining low power consumption, which is essential for extended use. Additionally, the system's flexibility allows for easy customization to suit individual needs, while its Wi-Fi and Bluetooth capabil-

ities offer potential for remote monitoring and feedback. Although challenges like signal noise and interference can affect performance, the cost-effectiveness of the system makes it more accessible, improving the viability of bionic prosthetics for a wider population. With further advancements in signal processing and machine learning, this system could evolve to offer even greater accuracy and adaptability, providing a valuable solution for those with limb loss.

REFERENCES

- [1] M., et al. Mind-controlled arm prostheses that 'feel' are now a part of everyday life. *New England Journal of Medicine*, 2020.
- [2] M., Zbinden, J., Millenaar, J., et al. A highly integrated bionic hand with neural control and feedback for use in daily life. *Science Robotics*, 2023.
- [3] Corbett, E. A., Perreault, E. J., & Kuiken, T. A. Comparison of electromyography and force as interfaces for prosthetic control. *Journal of Rehabilitation Research & Development*, 2011.
- [4] Li, G., Schultz, A. E., Bravo-Esteban, E., Popovic, M. R., & Pacelli, M. C. Proportional and simultaneous control of a hand prosthesis using surface electromyography. *IEEE Transactions on Biomedical Engineering*, 2010.
- [5] Ortiz-Catalan, M., et al. A highly integrated bionic hand with neural control and feedback for use in daily life. *Science Robotics*, 2023.
- [6] Jiang, H. Research on Multiple Control Methods for Bionic Hands. *SHS Web of Conferences*, 2022.