AANDERAA
a xylem brand

CALIBRATION CERTIFICATE

Form No 857, Feb 15th 2017

Certificate no: Oxygen Optode 4831_287_161140
FoilID: 1824

Product: **Oxygen Optode 4831**
Serial no: 287
Calibration date: 10.09.2019

2-POINT RECALIBRATION, OXYGEN OPTODE

This is to certify that this product has been calibrated and verified using the following reference equipment:

Reference Equipment	Serial
Calibration Bath Model Toho TM-005-P-A	12-K
Fluke 5615 PRT	Serial No. 802054 and Serial No. 849155
Fluke CHUB E-4	Serial No. A7C677
Reference sensor 4330	338

Specification

	O2 Concentration	Air Saturation	Temperature
Range	0-500 µM	0 - 120%	-5 to +40°C
Accuracy	< ±8µM or ±5% 1)	±5% 2)	±0.03°C 3)
Resolution	< 1 µM	< 0.4%	0.01°C
Settling Time (63%)	< 25 seconds	< 25 seconds	< 2 seconds

1) Requires salinity compensation for salinity variations > 1mS/cm, and pressure compensation for pressure > 100meter
2) Within calibrated range 0 - 120% / 0 - 30°C
3) Within calibrated range 0 - 36°C for standard version and 0 - 30°C for multipoint version

Calibration points and readings:

	Air Saturated Water	Zero Solution (Na2SO3)
Phase reading (°)	32.50	57.57
Temperature reading (°C)	10.03	21.42
Air Pressure (hPa)	959.00	

Giving these coefficients

Index	0 (Offset)	1 (Slope)
ConcCoef	-7.789490E+00	1.143122E+00

With following settings

Index	0	1	2	3	4	5	6
SVUFoilCoef	2.79000E-03	1.24362E-04	2.24170E-06	1.89435E+02	-1.74843E-01	-4.08575E+01	3.79875E+00
PhaseCoef	0.000000E+00	1.000000E+00	0.000000E+00	0.000000E+00			
TempCoef	2.263728E+01	-3.056573E-02	2.771066E-06	-4.144054E-09	0.000000E+00	0.000000E+00	
Salinity (mS/cm)	0.00						

Date:10.09.2019

Sr. Repair Technician