CALIBRATION CERTIFICATE

a xylem brand

Form No. 710, Dec 2005

Sensing Foil Batch No: 1206
Certificate No:

Product: Oxygen Optode 4831
Serial No: 287
Calibration Date: 07 Nov 2013

This is to certify that this product has been calibrated using the following instruments:

Parameter: Internal Temperature:

Calibration points and readings:

Temperature (°C)	1.00	11.97	24.00	35.97	
Reading (mV)	705.50	354.39	-44.33	-411.42	

Giving these coefficients

Index	0	1	2	3	4	5
TempCoef	2.26373E01	-3.05657E-02	2.77107E-06	-4.14405E-09	0.00000E00	0.00000E00

Parameter: Oxygen:

	O2 Concentration	Air Saturation
Range:	0-500 µM [1)]	0 - 120%
Accuracy[1)]:	< ±8µM or ±5% (whichever is greater)	±5%
Resolution:	< 1 µM	< 0.4%
Settling Time (63%):	< 25 seconds	

Calibration points and readings[2)]:

	Air Saturated Water	Zero Solution (Na_2SO_3)
Phase reading (°)	3.19589E+01	6.25805E+01
Temperature reading (°C)	9.90155E+00	2.24921E+01
Air Pressure (hPa)	9.48880E+02	

Giving these coefficients

Index	0	1	2	3
PhaseCoef	0.00000E00	1.00000E00	0.00000E00	0.00000E00

[1)] *Valid for 0 to 2000m (6562ft) depth, salinity 33 - 37ppt*
[2)] *The calibration is performed in fresh water and the salinity setting is set to:* 0

Date: 08 Nov 2013

Sign:

Tor-Ove Kvalvaag

Tor-Ove Kvalvaag, Calibration Engineer

AANDERAA DATA INSTRUMENTS AS

5851 BERGEN, NORWAY Tel. +47 55 60 48 00 Fax. +47 55 60 48 01 E-mail: info@aadi.no Web: http://www.aadi.no