PO Box 518
620 Applegate St.
Philomath, OR 97370

WETLabs

(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

Scattering Meter Calibration Sheet

4/16/2014
Wavelength: 470 S/N BB2FLSLC-1193

Use the following equation to obtain "scaled" output values:

$$\beta(\theta_c)\ m^{-1}\ sr^{-1} = \text{Scale Factor} \times (\text{Output} - \text{Dark Counts})$$

• **Scale Factor for 470 nm**	=	1.105E-05	$(m^{-1}sr^{-1})$/counts	
• **Output**	=	meter reading	counts	
• **Dark Counts**	=	45	counts	
Instrument Resolution	=	1.2	counts	1.38E-05 $(m^{-1}sr^{-1})$

Definitions:

- **Scale Factor**: Calibration scale factor, $\beta(\theta_c)$/counts. Refer to User's Guide for derivation.
- **Output**: Measured signal output of the scattering meter.
- **Dark Counts**: Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.

620 Applegate St.
Philomath, OR 97370

WET Labs

Fax (541) 929-5277
www.wetlabs.com

Scattering Meter Calibration Sheet

4/16/2014
Wavelength: 700 S/N BB2FLSLC-1193

Use the following equation to obtain "scaled" output values:

$$\beta(\theta_c)\ m^{-1}\ sr^{-1} = \text{Scale Factor} \times (\text{Output} - \text{Dark Counts})$$

• **Scale Factor for 700 nm**	=	3.035E-06	$(m^{-1}sr^{-1})$/counts	
• **Output**	=	meter reading	counts	
• **Dark Counts**	=	46	counts	
Instrument Resolution	=	1.0	counts	2.87E-06 $(m^{-1}sr^{-1})$

Definitions:

- **Scale Factor**: Calibration scale factor, $\beta(\theta_c)$/counts. Refer to User's Guide for derivation.
- **Output**: Measured signal output of the scattering meter.
- **Dark Counts**: Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.

PO Box 518
620 Applegate St.
Philomath, OR 97370

WETLabs

(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ECO Chlorophyll Fluorometer Characterization Sheet

Date: 4/16/2014 S/N: BB2FLSLC-1193

Chlorophyll concentration expressed in µg/l can be derived using the equation:

CHL (µg/l) = Scale Factor * (Output - Dark counts)

	Digital
Dark counts	46 counts
Scale Factor (SF)	0.0121 µg/l/count
Maximum Output	4130 counts
Resolution	1.3 counts
Ambient temperature during characterization	22.5 °C

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: SF = x ÷ (output - dark counts), where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyll-a concentrations in-situ is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (Thalassiosira weissflogii). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

PO Box 518
620 Applegate St.
Philomath, OR 97370

WET Labs®

(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

SLC Testing Certification

Date 4/17/2014 S/N# BB2FLSLC-1193

Low temperature test #1

Chill 2.5 hr at -20 °C

High temperature test #1

Heat 2.5 hr at 50 °C

Low temperature test #2 same protocol as #1

High temperature test #2 same protocol as #1

Pressure test

5 cycles, 0–1250 m with 10-sec. soaks
Held at 1250 m for 2 hrs. on last cycle

Electrical isolation

Resistance between copper faceplate and grounding wire is > 1 mΩ

Calibration verification

Verify calibration and dark counts in bb, chl, and CDOM channels
Verify 5% of single point check for chl and bb
Verify 10% of single point check for CDOM

Signature William Jason Rowe / Amy Cardos Moz

NOTES: