



# SOMERSET THERAPEUTICS LIMITED

## DATA REPORT - SVP1 AND AMPOULE LINE EMS

### WASHING AND STERILIZATION AREA

Date Time	Temperature	RH	Diffrential Pressure	Diffrential Pressure	Diffrential Pressure
DD-MM-YYYY HH:MM:SS	Washing and Sterilization Area (NMT 30°C)	Washing and Sterilization Area (NMT 80%)	Washing and Sterilization Area w.r.t Unit Preparation Room (NLT 6 Pa)	Washing and Sterilization Area w.r.t Ampoule Decartoning Room (NLT 12 Pa)	Washing and Sterilization Area w.r.t Injectable main corridor (NLT 12 Pa)
10-07-2022 11:43:00	17.6	79.4	48	20	45
10-07-2022 11:44:00	17.6	79.4	49	21	46
10-07-2022 11:45:00	17.6	79.1	47	22	45
10-07-2022 11:46:00	17.6	78.9	49	21	45
10-07-2022 11:47:00	17.6	78.8	48	22	46
10-07-2022 11:48:00	17.6	78.7	19	19	43
10-07-2022 11:49:00	17.6	78.5	49	21	45
10-07-2022 11:50:00	17.6	78.4	50	22	45
10-07-2022 11:51:00	17.6	78.1	49	22	45
10-07-2022 11:52:00	17.6	78	50	21	46
10-07-2022 11:53:00	17.6	77.8	49	21	46
10-07-2022 11:54:00	17.6	77.8	51	21	46
10-07-2022 11:55:00	17.6	77.8	48	21	46
10-07-2022 11:56:00	17.6	77.7	49	21	46
10-07-2022 11:57:00	17.6	77.7	50	22	45
10-07-2022 11:58:00	17.6	77.7	49	22	47
10-07-2022 11:59:00	17.6	77.7	49	22	45
10-07-2022 12:00:00	17.6	77.7	49	21	45
10-07-2022 12:01:00	17.6	77.6	49	21	46
10-07-2022 12:02:00	17.6	77.6	49	21	46
10-07-2022 12:03:00	17.6	77.5	48	22	46
10-07-2022 12:04:00	17.6	77.4	49	21	46
10-07-2022 12:05:00	17.6	77.6	49	22	46
10-07-2022 12:06:00	17.6	77.5	49	21	46

ACTIVITY	NAME	SIGNATURE AND DATE
REVIEWED BY USER		
REVIEWED BY QA		



# SOMERSET THERAPEUTICS LIMITED

## DATA REPORT - SVP1 AND AMPOULE LINE EMS

### WASHING AND STERILIZATION AREA

Date Time	Temperature	RH	Diffrential Pressure	Diffrential Pressure	Diffrential Pressure
DD-MM-YYYY HH:MM:SS	Washing and Sterilization Area (NMT 30°C)	Washing and Sterilization Area (NMT 80%)	Washing and Sterilization Area w.r.t Unit Preparation Room (NLT 6 Pa)	Washing and Sterilization Area w.r.t Ampoule Decartoning Room (NLT 12 Pa)	Washing and Sterilization Area w.r.t Injectable main corridor (NLT 12 Pa)
10-07-2022 12:07:00	17.6	77.5	49	22	46
10-07-2022 12:08:00	17.6	77.5	50	22	47
10-07-2022 12:09:00	17.6	77.6	49	22	46
10-07-2022 12:10:00	17.5	77.5	49	22	46
10-07-2022 12:11:00	17.5	77.5	48	22	46
10-07-2022 12:12:00	17.5	77.6	50	22	46
10-07-2022 12:13:00	17.5	77.5	50	23	47
10-07-2022 12:14:00	17.5	77.6	50	21	46
10-07-2022 12:15:00	17.5	77.6	49	22	46
10-07-2022 12:16:00	17.5	77.6	48	21	46
10-07-2022 12:17:00	17.5	77.6	48	13	23
10-07-2022 12:18:00	17.5	77.7	50	5	11
10-07-2022 12:19:00	17.5	77.9	52	0	0
10-07-2022 12:20:00	17.5	78.1	53	0	0
10-07-2022 12:21:00	17.5	78.2	51	0	0
10-07-2022 12:22:00	17.5	78.3	51	0	0
10-07-2022 12:23:00	17.4	78.5	50	1	0
10-07-2022 12:24:00	17.4	78.6	52	0	0
10-07-2022 12:25:00	17.4	78.7	50	0	0
10-07-2022 12:26:00	17.4	78.8	52	0	0
10-07-2022 12:27:00	17.3	78.9	51	0	0
10-07-2022 12:28:00	17.3	79	52	0	0
10-07-2022 12:29:00	17.4	78.8	52	0	0
10-07-2022 12:30:00	17.4	78.8	52	0	0

ACTIVITY	NAME	SIGNATURE AND DATE
REVIEWED BY USER		
REVIEWED BY QA		



# SOMERSET THERAPEUTICS LIMITED

## DATA REPORT - SVP1 AND AMPOULE LINE EMS

### WASHING AND STERILIZATION AREA

Date Time	Temperature	RH	Diffrential Pressure	Diffrential Pressure	Diffrential Pressure
DD-MM-YYYY HH:MM:SS	Washing and Sterilization Area (NMT 30°C)	Washing and Sterilization Area (NMT 80%)	Washing and Sterilization Area w.r.t Unit Preparation Room (NLT 6 Pa)	Washing and Sterilization Area w.r.t Ampoule Decartoning Room (NLT 12 Pa)	Washing and Sterilization Area w.r.t Injectable main corridor (NLT 12 Pa)
10-07-2022 12:31:00	17.4	78.8	52	0	0
10-07-2022 12:32:00	17.4	78.8	51	0	0
10-07-2022 12:33:00	17.3	78.7	50	0	0
10-07-2022 12:34:00	17.3	78.6	51	1	0
10-07-2022 12:35:00	17.3	78.7	52	0	0
10-07-2022 12:36:00	17.3	78.8	51	0	0
10-07-2022 12:37:00	17.3	78.7	52	0	0
10-07-2022 12:38:00	17.3	78.7	51	0	0
10-07-2022 12:39:00	17.3	78.7	50	0	0
10-07-2022 12:40:00	17.3	78.9	51	1	0
10-07-2022 12:41:00	17.4	78.9	52	0	0
10-07-2022 12:42:00	17.4	79	51	0	0
10-07-2022 12:43:00	17.3	79.1	52	0	0
10-07-2022 12:44:00	17.3	79.2	51	1	0
10-07-2022 12:45:00	17.3	79.3	52	0	0
10-07-2022 12:46:00	17.3	79.2	52	0	0
10-07-2022 12:47:00	17.3	79.3	51	0	0
10-07-2022 12:48:00	17.3	79.3	51	0	0
10-07-2022 12:49:00	17.3	79.3	50	0	0
10-07-2022 12:50:00	17.4	79.2	20	0	-1
10-07-2022 12:51:00	17.3	79.3	51	0	0
10-07-2022 12:52:00	17.3	79.3	51	0	0
10-07-2022 12:53:00	17.4	79.1	52	0	0
10-07-2022 12:54:00	17.4	79.2	51	0	0

ACTIVITY	NAME	SIGNATURE AND DATE
REVIEWED BY USER		
REVIEWED BY QA		



# SOMERSET THERAPEUTICS LIMITED

## DATA REPORT - SVP1 AND AMPOULE LINE EMS

### WASHING AND STERILIZATION AREA

Date Time	Temperature	RH	Diffrential Pressure	Diffrential Pressure	Diffrential Pressure
DD-MM-YYYY HH:MM:SS	Washing and Sterilization Area (NMT 30°C)	Washing and Sterilization Area (NMT 80%)	Washing and Sterilization Area w.r.t Unit Preparation Room (NLT 6 Pa)	Washing and Sterilization Area w.r.t Ampoule Decartoning Room (NLT 12 Pa)	Washing and Sterilization Area w.r.t Injectable main corridor (NLT 12 Pa)
10-07-2022 12:55:00	17.4	79.2	52	0	1
10-07-2022 12:56:00	17.4	79.6	51	0	0
10-07-2022 12:57:00	17.4	80.1	52	0	0
10-07-2022 12:58:00	17.4	80.4	51	0	0
10-07-2022 12:59:00	17.4	80.5	51	0	0
10-07-2022 13:00:00	17.4	80.6	51	0	0
10-07-2022 13:01:00	17.4	80.8	52	0	0
10-07-2022 13:02:00	17.4	80.8	52	0	0
10-07-2022 13:03:00	17.4	81.2	51	0	0
10-07-2022 13:04:00	17.4	81.5	51	0	0
10-07-2022 13:05:00	17.4	81.8	50	0	0
10-07-2022 13:06:00	17.5	82	50	0	0
10-07-2022 13:07:00	17.5	82	52	0	0
10-07-2022 13:08:00	17.5	82	52	0	0
10-07-2022 13:09:00	17.5	81.9	50	0	0
10-07-2022 13:10:00	17.6	81.7	50	0	0
10-07-2022 13:11:00	17.6	81.5	51	0	0
10-07-2022 13:12:00	17.6	81.6	51	0	0
10-07-2022 13:13:00	17.6	81.7	49	1	0
10-07-2022 13:14:00	17.6	81.8	51	0	0
10-07-2022 13:15:00	17.6	81.8	50	0	0
10-07-2022 13:16:00	17.7	81.8	47	0	0
10-07-2022 13:17:00	17.6	81.8	50	0	0
10-07-2022 13:18:00	17.7	81.4	49	0	0

ACTIVITY	NAME	SIGNATURE AND DATE
REVIEWED BY USER		
REVIEWED BY QA		



# SOMERSET THERAPEUTICS LIMITED

## DATA REPORT - SVP1 AND AMPOULE LINE EMS

### WASHING AND STERILIZATION AREA

Date Time	Temperature	RH	Diffrential Pressure	Diffrential Pressure	Diffrential Pressure
DD-MM-YYYY HH:MM:SS	Washing and Sterilization Area (NMT 30°C)	Washing and Sterilization Area (NMT 80%)	Washing and Sterilization Area w.r.t Unit Preparation Room (NLT 6 Pa)	Washing and Sterilization Area w.r.t Ampoule Decartoning Room (NLT 12 Pa)	Washing and Sterilization Area w.r.t Injectable main corridor (NLT 12 Pa)
10-07-2022 13:19:00	17.6	81.1	50	0	0
10-07-2022 13:20:00	17.6	81.1	51	0	0
10-07-2022 13:21:00	17.6	80.9	50	0	0
10-07-2022 13:22:00	17.6	80.7	52	0	0
10-07-2022 13:23:00	17.6	80.5	51	0	0
10-07-2022 13:24:00	17.6	80.5	50	0	0
10-07-2022 13:25:00	17.6	80.3	51	0	0
10-07-2022 13:26:00	17.6	80.1	50	0	0
10-07-2022 13:27:00	17.6	80.1	50	0	0
10-07-2022 13:28:00	17.5	80	52	0	0
10-07-2022 13:29:00	17.5	79.9	49	0	0
10-07-2022 13:30:00	17.5	79.9	30	0	0
10-07-2022 13:31:00	17.6	80	50	0	0
10-07-2022 13:32:00	17.6	79.9	50	0	0
10-07-2022 13:33:00	17.5	80	50	0	0
10-07-2022 13:34:00	17.6	79.9	50	0	0
10-07-2022 13:35:00	17.6	80	50	0	0
10-07-2022 13:36:00	17.5	80	51	0	0
10-07-2022 13:37:00	17.6	79.9	50	1	0
10-07-2022 13:38:00	17.6	79.8	50	0	0
10-07-2022 13:39:00	17.6	79.6	51	0	0
10-07-2022 13:40:00	17.6	79.9	51	0	0
10-07-2022 13:41:00	17.6	79.9	52	0	0
10-07-2022 13:42:00	17.6	80	51	0	0

ACTIVITY	NAME	SIGNATURE AND DATE
REVIEWED BY USER		
REVIEWED BY QA		



## SOMERSET THERAPEUTICS LIMITED

### DATA REPORT - SVP1 AND AMPOULE LINE EMS

#### WASHING AND STERILIZATION AREA

Date Time	Temperature	RH	Diffrential Pressure	Diffrential Pressure	Diffrential Pressure
DD-MM-YYYY HH:MM:SS	Washing and Sterilization Area (NMT 30°C)	Washing and Sterilization Area (NMT 80%)	Washing and Sterilization Area w.r.t Unit Preparation Room (NLT 6 Pa)	Washing and Sterilization Area w.r.t Ampoule Decartoning Room (NLT 12 Pa)	Washing and Sterilization Area w.r.t Injectable main corridor (NLT 12 Pa)
10-07-2022 13:43:00	17.6	80	52	1	0
Min Value	17.3	77.4	19	0	-1
Max Value	17.7	82	53	23	47

End Of Report : From 10-07-2022 11:43:11 To 10-07-2022 13:43:11

ACTIVITY	NAME	SIGNATURE AND DATE
REVIEWED BY USER		
REVIEWED BY QA		