



SOMERSET THERAPEUTICS LIMITED

DATA REPORT - SVP1 AND AMPOULE LINE EMS

WASHING AND STERILIZATION AREA

Date Time	Temperature	RH	Diffrential Pressure	Diffrential Pressure	Diffrential Pressure
DD-MM-YYYY HH:MM:SS	Washing and Sterilization Area (NMT 30°C)	Washing and Sterilization Area (NMT 80%)	Washing and Sterilization Area w.r.t Unit Preparation Room (NLT 6 Pa)	Washing and Sterilization Area w.r.t Ampoule Decartoning Room (NLT 12 Pa)	Washing and Sterilization Area w.r.t Injectable main corridor (NLT 12 Pa)
10-07-2022 11:13:00	17.7	77.3	48	22	46
10-07-2022 11:28:00	17.6	77.4	50	22	42
10-07-2022 11:43:00	17.6	79.4	48	20	45
10-07-2022 11:58:00	17.6	77.7	49	22	47
10-07-2022 12:13:00	17.5	77.5	50	23	47
10-07-2022 12:28:00	17.3	79	52	0	0
10-07-2022 12:43:00	17.3	79.1	52	0	0
10-07-2022 12:58:00	17.4	80.4	51	0	0
10-07-2022 13:13:00	17.6	81.7	49	1	0
10-07-2022 13:28:00	17.5	80	52	0	0
10-07-2022 13:43:00	17.6	80	52	1	0
10-07-2022 13:58:00	17.5	79.9	51	0	0
10-07-2022 14:13:00	17.5	79.5	50	0	0
10-07-2022 14:28:00	17.6	79.6	51	0	0
10-07-2022 14:43:00	17.6	79.4	50	0	0
10-07-2022 14:58:00	17.6	80.2	50	1	0
10-07-2022 15:13:00	19.5	78.1	0	0	1
10-07-2022 15:28:00	20.3	76.5	0	0	1
10-07-2022 15:43:00	20.8	74.9	0	0	0
10-07-2022 15:58:00	21	74.2	0	0	0
10-07-2022 16:13:00	21	75.6	0	0	0
10-07-2022 16:28:00	21	85.4	0	0	0
10-07-2022 16:43:00	21	95	0	0	0
10-07-2022 16:58:00	21.5	96.5	0	0	0

ACTIVITY	NAME	SIGNATURE AND DATE
REVIEWED BY USER		
REVIEWED BY QA		



SOMERSET THERAPEUTICS LIMITED

DATA REPORT - SVP1 AND AMPOULE LINE EMS

WASHING AND STERILIZATION AREA

Date Time	Temperature	RH	Diffrential Pressure	Diffrential Pressure	Diffrential Pressure
DD-MM-YYYY HH:MM:SS	Washing and Sterilization Area (NMT 30°C)	Washing and Sterilization Area (NMT 80%)	Washing and Sterilization Area w.r.t Unit Preparation Room (NLT 6 Pa)	Washing and Sterilization Area w.r.t Ampoule Decartoning Room (NLT 12 Pa)	Washing and Sterilization Area w.r.t Injectable main corridor (NLT 12 Pa)
10-07-2022 17:13:00	21.9	96.7	0	0	0
10-07-2022 17:28:00	22	95	0	0	0
10-07-2022 17:43:00	21.8	94.6	0	0	0
10-07-2022 17:58:00	21.7	94.4	0	0	1
10-07-2022 18:13:00	21.7	94.2	0	0	0
10-07-2022 18:28:00	21.7	94	0	0	0
10-07-2022 18:43:00	21	82.7	45	21	44
10-07-2022 18:58:00	19.4	78.4	47	21	44
10-07-2022 19:13:00	18.7	77.1	47	16	35
10-07-2022 19:28:00	18.1	77	47	21	44
10-07-2022 19:43:00	17.9	77.3	46	20	43
10-07-2022 19:58:00	17.8	77.6	46	21	44
10-07-2022 20:13:00	17.7	77.1	46	21	43
10-07-2022 20:28:00	17.6	77.2	46	20	42
10-07-2022 20:43:00	17.8	77.8	18	19	41
10-07-2022 20:58:00	17.9	78.1	47	21	44
10-07-2022 21:13:00	17.9	77.7	48	21	45
10-07-2022 21:28:00	18.1	79.8	46	21	45
10-07-2022 21:43:00	18	79.3	29	20	42
10-07-2022 21:58:00	18	79.5	28	21	42
10-07-2022 22:13:00	17.9	77.6	47	22	46
10-07-2022 22:28:00	17.8	77.1	43	19	40
10-07-2022 22:43:00	17.7	76.6	47	21	43
10-07-2022 22:58:00	17.6	76.7	47	21	45

ACTIVITY	NAME	SIGNATURE AND DATE
REVIEWED BY USER		
REVIEWED BY QA		



SOMERSET THERAPEUTICS LIMITED

DATA REPORT - SVP1 AND AMPOULE LINE EMS

WASHING AND STERILIZATION AREA

Date Time	Temperature	RH	Diffrential Pressure	Diffrential Pressure	Diffrential Pressure
DD-MM-YYYY HH:MM:SS	Washing and Sterilization Area (NMT 30°C)	Washing and Sterilization Area (NMT 80%)	Washing and Sterilization Area w.r.t Unit Preparation Room (NLT 6 Pa)	Washing and Sterilization Area w.r.t Ampoule Decartoning Room (NLT 12 Pa)	Washing and Sterilization Area w.r.t Injectable main corridor (NLT 12 Pa)
10-07-2022 23:13:00	17.6	77.4	46	19	41
10-07-2022 23:28:00	17.6	77.9	46	19	40
10-07-2022 23:43:00	18.2	75.7	47	20	41
10-07-2022 23:58:00	17.9	75.6	46	20	42
11-07-2022 00:13:00	17.8	77.8	48	27	50
11-07-2022 00:28:00	17.9	77.4	46	21	41
11-07-2022 00:43:00	17.9	77.7	46	20	39
11-07-2022 00:58:00	17.8	76.8	47	18	37
11-07-2022 01:13:00	17.6	76.2	44	18	37
11-07-2022 01:28:00	17.4	76.5	46	19	41
11-07-2022 01:43:00	17.3	77.1	47	20	41
11-07-2022 01:58:00	17.2	76.9	48	20	42
11-07-2022 02:13:00	17.2	76.8	48	20	42
11-07-2022 02:28:00	17.3	76.6	47	19	41
11-07-2022 02:43:00	17.2	77.1	47	20	41
11-07-2022 02:58:00	17.3	76.7	49	20	41
11-07-2022 03:13:00	17.3	76.4	47	17	39
11-07-2022 03:28:00	17.6	76	46	19	39
11-07-2022 03:43:00	17.4	76.9	46	19	38
11-07-2022 03:58:00	17.8	79.8	44	19	38
11-07-2022 04:13:00	17.9	76.6	46	19	39
11-07-2022 04:28:00	17.8	76	46	19	39
11-07-2022 04:43:00	17.7	76	45	19	38
11-07-2022 04:58:00	17.6	75.9	46	18	39

ACTIVITY	NAME	SIGNATURE AND DATE
REVIEWED BY USER		
REVIEWED BY QA		



SOMERSET THERAPEUTICS LIMITED

DATA REPORT - SVP1 AND AMPOULE LINE EMS

WASHING AND STERILIZATION AREA

Date Time	Temperature	RH	Diffrential Pressure	Diffrential Pressure	Diffrential Pressure
DD-MM-YYYY HH:MM:SS	Washing and Sterilization Area (NMT 30°C)	Washing and Sterilization Area (NMT 80%)	Washing and Sterilization Area w.r.t Unit Preparation Room (NLT 6 Pa)	Washing and Sterilization Area w.r.t Ampoule Decartoning Room (NLT 12 Pa)	Washing and Sterilization Area w.r.t Injectable main corridor (NLT 12 Pa)
11-07-2022 05:13:00	17.5	76	45	17	34
11-07-2022 05:28:00	17.4	76.5	47	18	38
11-07-2022 05:43:00	17.4	76.3	47	18	38
11-07-2022 05:58:00	17.5	76	46	20	41
11-07-2022 06:13:00	17.6	75.4	47	20	42
11-07-2022 06:28:00	17.6	77.2	45	20	42
11-07-2022 06:43:00	17.9	78.3	47	20	42
11-07-2022 06:58:00	18	78.7	46	20	43
11-07-2022 07:13:00	17.9	78.6	47	20	41
11-07-2022 07:28:00	18	79.5	45	20	41
11-07-2022 07:43:00	17.8	77.5	46	20	42
11-07-2022 07:58:00	17.7	76.9	45	20	42
11-07-2022 08:13:00	17.6	78	43	18	40
11-07-2022 08:28:00	17.6	78.6	47	19	43
11-07-2022 08:43:00	17.6	78.8	47	20	41
11-07-2022 08:58:00	17.7	79.8	47	19	42
11-07-2022 09:13:00	17.7	80.7	47	20	43
11-07-2022 09:28:00	17.7	80.5	46	19	40
11-07-2022 09:43:00	17.8	79.8	47	18	39
11-07-2022 09:58:00	18	83.7	45	18	40
11-07-2022 10:13:00	17.8	81	46	18	35
11-07-2022 10:28:00	17.8	79.3	17	16	33
11-07-2022 10:43:00	18	78.8	44	16	36
11-07-2022 10:58:00	18.1	78.5	46	17	34

ACTIVITY	NAME	SIGNATURE AND DATE
REVIEWED BY USER		
REVIEWED BY QA		



SOMERSET THERAPEUTICS LIMITED

DATA REPORT - SVP1 AND AMPOULE LINE EMS

WASHING AND STERILIZATION AREA

Date Time	Temperature	RH	Diffrential Pressure	Diffrential Pressure	Diffrential Pressure
DD-MM-YYYY HH:MM:SS	Washing and Sterilization Area (NMT 30°C)	Washing and Sterilization Area (NMT 80%)	Washing and Sterilization Area w.r.t Unit Preparation Room (NLT 6 Pa)	Washing and Sterilization Area w.r.t Ampoule Decartoning Room (NLT 12 Pa)	Washing and Sterilization Area w.r.t Injectable main corridor (NLT 12 Pa)
11-07-2022 11:13:00	18.2	77	44	17	37
Min Value	17.2	74.2	0	0	0
Max Value	22	96.7	52	27	50

End Of Report : From 10-07-2022 11:13:31 To 11-07-2022 11:13:31

ACTIVITY	NAME	SIGNATURE AND DATE
REVIEWED BY USER		
REVIEWED BY QA		