



Calibration Certificate

Certificate NO.	{cno}	Certificate Date :	{cedate}
-----------------	-------	--------------------	----------

Reference Instrumentation

Device	Manufacturer	Equipment	Accuracy	Calibration Date
Temperature & Humidity Chamber	ESPEC	SH-241	$\pm 0.2^{\circ}\text{C}$	2021/07/01

Optilo Co., Ltd. certifies that the devices mentioned on this certificate have been thoroughly tested, validated and reached performance accuracy specifications over the specific ranges. These device were assembled, tested and calibrated in accordance with its specifications. All reference instruments are calibrated by Optilo Corporate-affiliated research institutes.

Product Information

Product Name	Model	Serial No.	Calibration Date	Expiry Date	Quantity
OPT A-10	ES-200	{sno}	{pdate}	{edate}	1PCS

Calibration Results (Environment Conditions: Air Temperature 20.5°C, Relative Humidity 53.0%)

Validation Points	Tolerance	Result
-20°C	± 0.5	PASS
0°C	± 0.5	PASS
40°C	± 0.5	PASS

Test Result Pass/Fail	Pass	Signature of Calibration Technician	
-----------------------	------	-------------------------------------	---

OPTILO CO.,LTD.
주식회사 옵티로

Optilo Co., Ltd

1703, Smart Valley Building D, 30, Songdomirae-Ro, Yeonsu-gu, Incheon, Republic of Korea, 21990

Tel: 070-5143-8585 Fax: 050-4886-5157 Email: info@optilol.net Web: www.optilo.net