

## Certificate of Analysis

## **Custom Standard**

Product Number: PST-560M1000 Page: 1 of 1

Lot Number: CP-1357Z Lot Issue Date: 03-Apr-2017 Expiration Date: 31-May-2018

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte CAS# Analyte Lot True Value azinphos-methyl 000086-50-0 RM06810  $1003 \pm 5 \mu g/mL$ 

Matrix: methanol (methyl alcohol)

**Storage:** Store at Room Temperature (15° to 30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

ACCREDITED

SOUTHINGS

TESTING LABORATORY

TESTING LABORATORY

TOV USA, Inc.

John Russo President Monica Bourgeois Director of QA/RA