



# Certificate of Analysis ISO Guide 34

## Phosalone Standard

**Product Number:** PST-2500K1000

**Page:** 1 of 1

**Lot Number:** CT-1189

**Lot Issue Date:** 08-Mar-2019

**Expiration Date:** 30-Apr-2021

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent Technologies ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
phosalone	002310-17-0	RM14709	1001 ± 5 µg/mL

**Matrix:** acetone

**Storage:** Store at Room Temperature (15° to 30°C), Do Not Refrigerate.

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

  
Monica Bourgeois  
QMS Representative



ISO Guide 34 Cert No.  
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026  
registered ISO 9001 Quality Management System



ISO17025 Cert No.  
AT-1937