

# Certificate of Analysis

## ISO Guide 34 Reference Material

**Product Number:** ICP-126  
**Lot Number:** CS-1067

**Lot Issue Date:** 08-Mar 2018  
**Expiration Date:** 30-Apr 2025

**Product Name:** Iron ICP Standard

### Description:

This Reference Material (RM) was gravimetrically prepared in accordance with ISO Guide 34 and under ULTRA Scientific's ISO 9001 registered quality system. The neat materials used for this product have been verified by ULTRA's ISO 17025 laboratory and under ULTRA's ISO Guide 34 accreditation. The analyte concentrations were verified by ULTRA's ISO 17025 accredited laboratory. For each analyte, the true value, with its uncertainty value calculated at the 95% confidence level, is reported below.

Analyte	Starting Material	Lot Number	Purity (%)	Analyte Concentration	Traceability & Method
iron	iron(III) nitrate, nonahydrate	RM11958	99.999%	10004 ± 20 µg/mL	NIST SRM 3126a; ICP-OES

**Solvent:** 2% nitric acid in low TOC water (< 50 ppb)

### Non-Certified Values:

**Density:** 1.0429 g/mL @ 20.00 ± 0.05°C

### Trace Metallic Impurities in Solution Standard in µg/mL

*__	Al	<0.005 D	*__	Ga	<0.005 ND	n__	Nb	n__	S	
*__	Sb	<0.005 ND	n__	Ge		n__	Os	n__	Ta	
*__	As	<0.005 ND	n__	Au		*__	Pd	<0.005 ND	n__	Te
*__	Ba	<0.005 ND	n__	Hf		*__	P	<0.005 ND	n__	Tb
*__	Be	<0.005 ND	n__	Ho		*__	Pt	<0.005 ND	*__	Tl
*__	Bi	<0.005 ND	*__	In	<0.005 ND	*__	K	<0.005 ND	n__	Th
*__	B	<0.005 ND	n__	Ir		n__	Pr		n__	Tm
*__	Cd	<0.005 ND	s__	Fe		n__	Re	*__	Sn	<0.005 ND
n__	Cs		*__	La	<0.005 ND	n__	Rh	*__	Ti	<0.005 ND
*__	Ca	<0.005 D	*__	Pb	<0.005 ND	n__	Rb	n__	W	
n__	Ce		*__	Li	<0.005 ND	n__	Ru	n__	U	
*__	Cr	<0.005 D	n__	Lu		n__	Sm	*__	V	<0.005 ND
*__	Co	<0.005 D	*__	Mg	<0.005 ND	n__	Sc	n__	Yb	
*__	Cu	<0.005 ND	*__	Mn	<0.005 D	*__	Se	<0.005 ND	n__	Y
n__	Dy		*__	Hg	<0.005 ND	*__	Si	<0.005 ND	*__	Zn
*__	Er	<0.005 ND	*__	Mo	<0.005 ND	*__	Ag	<0.005 ND	n__	Zr
*__	Eu	<0.005 ND	n__	Nd		*__	Na	<0.005 D		
*	Gd	<0.005 ND	*	Ni	<0.005 ND	*	Sr	<0.005 ND		

\* - element checked; i - spectral interference; n - element not checked; D - element detected; ND - element not detected;  
s - standard element

## ISO Guide 34 Reference Material

**Product Number:** ICP-126  
**Lot Number:** CS-1067

**Lot Issue Date:** 08-Mar 2018  
**Expiration Date:** 30-Apr 2025

**Storage:** Store at Room Temperature (15° to 30°C).

### **Traceability:**

Traceability has been established through an unbroken chain of comparisons, each having stated uncertainties. Comparisons are based on appropriate physical or chemical measurements, including gravimetric or volumetric dilution, where the mass or volume of a solution before and after dilution is measured. The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1, ISO 9001, ISO 17025, and ISO Guide 34. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 819.

### **Estimation of Uncertainties:**

The true value is reported, with its uncertainty value calculated at the 95% confidence level.

### **Homogeneity:**

This RM was formulated and unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

### **Intended Use:**

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods and continuing calibration verification.

### **Instructions for Use:**

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening and should be processed without delay for the true value to be valid within the stated uncertainties. Do not pipet from the bottle. Do not return any material removed for pipetting to the bottle. Tightly cap the bottle after removing any material and store according to the instructions noted above.

### **Hazards:**



Refer to the Safety Data Sheet for information regarding this RM.

### **Expiration of Certification:**

The certification of this RM is valid, within the measurement uncertainty specified, until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

### **Maintenance of Certification:**

The real-time, long term stability of the RM may be monitored over the lifetime of the certification. If substantive changes occur that affect the certification before the expiration of this certificate, ULTRA Scientific will notify the purchaser.

  
John Russo  
President  
Monica Bourgeois  
Director of QA/RA