



# Certificate of Analysis

**Chlordane Sample**

**Catalog Number:**

**QCM-230**

**Code Number:**

**30275**

This ULTRAcheck™ sample was gravimetrically prepared, and the analyte concentrations were confirmed using high resolution gas chromatography. All operations were carried out under ULTRA Scientific's ISO 9001 registered quality system. The true values represent the gravimetrically determined values when the sample has been prepared according to the accompanying instructions. All gravimetric values are traceable to NIST.

Analyte	CAS #	True Value (µg/L)	Advisory Range*
chlordane	57-74-9	3.79	2.08 - 5.50

\* Calculated from the NELAC Drinking Water Fields of Testing Document, effective 1/1/09.



ISO 17025:2005  
Accredited  
A2LA  
Cert. No. 0851.01

ISO 9001:2000  
Registered  
TUV USA, Inc.  
Cert. No. 06-1004

  
William J. Leary  
Quality Assurance Manager