



Certificate of Analysis ISO Guide 34

2,3,3',4,4',5'-Hexachlorobiphenyl Standard (BZ#157)

Product Number: RPC-164AS **Page:** 1 of 1
Lot Number: CS-5381 **Lot Issue Date:** 06-Nov-2018 **Expiration Date:** 31-Dec-2026

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
2,3,3',4,4',5'-hexachlorobiphenyl (BZ # 157)	069782-90-7	RM00721	99.9 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage: Store at Room Temperature (15° to 30°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.


Monica Bourgeois
QMS Representative



Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management System



ORG ver 1.1