

# Certificate of Analysis

**Regulated Volatiles Sample**
**Catalog Number: QCM-111**
**Code Number: 34573**

This ULTRAcheck™ sample was gravimetrically prepared, and the analyte concentrations were confirmed using high resolution gas chromatography. All operations were carried out under ULTRA Scientific's ISO 9001 registered quality system. The assigned values represent the gravimetrically determined values when the sample has been prepared according to the accompanying instructions. All gravimetric values are traceable to NIST.


Analyte	CAS #	Assigned Value (µg/L)	Advisory Range*
benzene	71-43-2	9.64	5.78 - 13.50
carbon tetrachloride	56-23-5	18.4	14.7 - 22.1
chlorobenzene	108-90-7	9.42	5.65 - 13.19
1,2-dichlorobenzene	95-50-1	17.7	14.2 - 21.3
1,4-dichlorobenzene	106-46-7	8.76	5.26 - 12.26
1,2-dichloroethane	107-06-2	15.3	12.2 - 18.4
1,1-dichloroethene	75-35-4	13.8	11.1 - 16.6
cis-1,2-dichloroethene	156-59-4	10.1	8.1 - 12.1
trans-1,2-dichloroethene	156-60-5	19.6	15.7 - 23.5
1,2-dichloropropane	78-87-5	18.0	14.4 - 21.5
ethylbenzene	100-41-4	7.01	4.21 - 9.81
methylene chloride	75-09-2	15.8	12.7 - 19.0
styrene	100-42-5	1.10	0.66 - 1.54
tetrachloroethene	127-18-4	13.3	10.6 - 15.9
toluene	108-88-3	17.2	13.7 - 20.6
1,2,4-trichlorobenzene	120-82-1	13.8	11.0 - 16.5
1,1,1-trichloroethane	71-55-6	15.6	12.5 - 18.7
1,1,2-trichloroethane	79-00-5	4.77	2.86 - 6.68
trichloroethene	79-01-6	19.1	15.3 - 22.9
vinyl chloride	75-01-4	6.03	3.62 - 8.44
o-xylene	95-47-6	17.0	13.6 - 20.4
m-xylene	108-38-3	17.5	14.0 - 21.0
p-xylene	106-42-3	9.43	5.66 - 13.20

\* Calculated from the NELAP Drinking Water Fields of Testing Document, effective 1/1/09.



ISO 17025:2005  
Accredited  
A2LA  
Cert. No. 0851.01

ISO 9001:2000  
Registered  
TUV USA, Inc.  
Cert. No. 06-1004



William J. Leary  
Quality Assurance Manager