



# Certificate of Analysis ISO Guide 34

## Monocrotophos Standard

**Product Number:** PST-040A1000

**Page:** 1 of 1

**Lot Number:** CT-1023

**Lot Issue Date:** 04-Mar-2019

**Expiration Date:** 30-Apr-2021

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent Technologies ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
monocrotophos	006923-22-4	RM12452	1003 ± 5 µg/mL

**Matrix:** acetonitrile

**Storage:** Store Refrigerated (2° to 8°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

  
Monica Bourgeois  
QMS Representative



ISO Guide 34 Cert No.  
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026  
registered ISO 9001 Quality Management System



ISO17025 Cert No.  
AT-1937