

Certificate of Analysis

Product Name: Custom Standard

Product Number: CUS-18071

Lot Issue Date: 10-Dec-2019

Lot Number: 0006507486

Expiration Date: 31-Jan-2022

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system, and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration \pm Uncertainty
trichlorofluoromethane	000075-69-4	RM00017	1002 \pm 5 μ g/mL
acetone	000067-64-1	RM13102	1001 \pm 5 μ g/mL
methylene chloride	000075-09-2	RM09575	1004 \pm 5 μ g/mL
ethyl acetate	000141-78-6	RM12930	1002 \pm 5 μ g/mL
benzene	000071-43-2	RM03830	1001 \pm 5 μ g/mL
trichloroethene	000079-01-6	RM14232	1002 \pm 5 μ g/mL
2-nitropropane	000079-46-9	RM12249	1002 \pm 5 μ g/mL
pyridine	000110-86-1	RM06750	1001 \pm 5 μ g/mL
toluene	000108-88-3	RM15383	1001 \pm 5 μ g/mL
1,1,2-trichloroethane	000079-00-5	RM01175	1001 \pm 5 μ g/mL
n-butyl acetate	000123-86-4	RM07291	1002 \pm 5 μ g/mL
chlorobenzene	000108-90-7	RM01874	1002 \pm 5 μ g/mL
o-cresol	000095-48-7	RM12877	1002 \pm 5 μ g/mL
p-cresol	000106-44-5	RM01988	1001 \pm 5 μ g/mL
o-xylene	000095-47-6	RM15639	1002 \pm 5 μ g/mL
p-xylene	000106-42-3	RM02647	1002 \pm 5 μ g/mL

Matrix: tetradecane



Certificate of Analysis

Product Number: CUS-18071

Lot Number: 0006507486

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert No.
AR-1936

RM was produced in accordance with TUV USA Inc registered ISO
9001 Quality Management System. Cert # 56 100 18560026

Page: 2 of 2

[www.agilent.com/quality/
CSD-QA-015.1](http://www.agilent.com/quality/CSD-QA-015.1)



ISO 17025 Cert
No. AT-1937