



Sam Parker <samuel.parker5@gmail.com>

EMI shielding for non-invasive BCI systems

David Grayden <grayden@unimelb.edu.au>
To: Sam Parker <samuel.parker5@gmail.com>

Mon, Jun 15, 2020 at 7:57 PM

Hi Sam,

It is good to hear from you and of your success so far. I remember meeting you back in 2018.

I reckon that EMI will have something to do with your performance but I think it is most likely to be related to the quality of equipment and number of participants. Just using dry electrodes could easily lead to this much reduction in performance compared to wet electrodes.

Regards,

David

Professor David B. Grayden

Clifford Chair of Neural Engineering

Director, MSE Medical Technologies Research Platform

Dept of Biomedical Engineering

Melbourne School of Engineering (MSE)

The University of Melbourne, Victoria 3010 Australia

<https://biomedical.eng.unimelb.edu.au/> | <https://www.linkedin.com/in/dgrayden>

[Quoted text hidden]