

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: DAVAO DOCTORS HOSPITAL ADDRESS: DAVAO CITY
SERVICE REQUESTED BY: DEPARTMENT: ANNEX 4
EQUIPMENT/MODEL: MOBILE RAD. SYSTEM / MUX-10 DR SERIAL NO.: 0262P84109

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: SCHEDULED PMS - 2ND VISIT (YEAR 2026)

ACTIONS TAKEN

-TESTED UNIT BEFORE PMS, OK
-CHECKED AND CLEANED THE UNIT
-CHECKED ALL MECHANICAL MOVEMENTS, OK
-CHECKED BOARDS AND FUSES
-APPLIED LUBRICANT(WD-40) TO MOVING PARTS
-CHECKED FAULT LOGS, RECORDED F14 ERRORS
-PERFORMED TUBE AGING, NO ERROR
-CHECKED TUBE CALIBRATION, OK
-CHECKED AND CLEANED DR SYSTEM, PC, CRADLE, SCU
-PERFORMED SELF DIAGNOSIS FOR FPD, OK
-CHECKED STORAGE AND ENDORSED TO END USER TO DELETE OLD PATIENT DATA
-TESTED IMAGE ACQUISITION AFTER PMS, OK
-TESTED IMAGE SENDING, OK
-ENDORSED TO END USER TO PERFORM TUBE AGING/WARM-UP

EXPOSURE COUNT: 49,197

REMARKS/RECOMMENDATIONS

-UNIT IN GOOD WORKING CONDITION

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☐ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
05/05/2026	10:00AM	12:00PM	SALES INVOICE
			PURCHASE ORDER
			SC

SERVICED BY:

JOSEPH REY B. DUMASIG
NAME & SIGNATURE

ACKNOWLEDGED BY:

JONIEL JANSAN TESARA
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting