

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: UNIVERSITY OF CEBU MEDICAL CENTER ADDRESS: MADNAUE CITY, CEBU
SERVICE REQUESTED BY: BIOMED DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: RADSPED FIT W/ CXDI SERIAL NO.: MP41EBA6A004

SERVICE CATEGORY

☐ WARRANTY ☒ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: COLUMN PROBLEM

ACTIONS TAKEN

-COLUMN HAS BEEN REINFORCED BY FABRICATION OF HOSPITAL
-VISITED SITE
-INSTALLED WEIGHTS OF COLUMN
-INSTALLED CARRIAGE ASSY
-INSTALLED WIRE ROPE; ADJUSTED TENSION OF WIRE ROPE
-INSTALLED X-RAY TUBE AND REGREASED HIGH VOLTAGE CABLE
-INSTALLED COLLIMATOR
-INSTALLED CABLE CONNECTIONS
-PERFORMED X-RAY TUBE AGING/WARM UP
-PERFORMED X-RAY TUBE ALIGNMENT WITH RESPECT TO TABLE
-PERFORMED X-RAY TUBE ALIGNMENT WITH RESPECT TO STAND; LIMITED ALIGNMENT DUE TO THE WELDING DONE IN ARM AND ALSO THE CARRIAGE ASSY MOVES WHEN X-RAY TUBE IS EXTENDED
-CHECKED IMAGE OUTPUT; FOR CHEST IMAGES, IMAGES HAS NO ARTIFACT AND IMAGE QUALITY IS NOT THE BEST BECAUSE OF THE X-RAY TUBE BEING GASSY; WHEN SHOOTING THICK PATIENTS, EXPECT IMAGE NOT TO BE OF BEST QUALITY
-XRAY TUBE IS RECOMMENDED FOR REPLACEMENT
-ARM ASSY IS ALSO RECOMMENDED FOR REPLACEMENT DUE TO THE WELDED PART WHICH MAKES ALIGNMENT DIFFICULT
-ENDORSED TO BIOMED

REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING CONDITION TEMPORARILY WHILE WAITING FOR THE COLUMN/STAND ASSY AND X-RAY TUBE. OBSERVE IMAGES TAKEN AND OBSERVE CAUTION WHEN ROTATING ARM TO PERFORM STAND PROCEDURES

SUBMITTED ORIGINAL DOCUMENTS


☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT
1		ARM ASSY	1	

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 04, 2026	12:00 NN	7:00 PM	SALES INVOICE
			PURCHASE ORDER

SERVICED BY:


JASPER B. REDOBLE
NAME & SIGNATURE

ACKNOWLEDGED BY:


KENNETH S. LUCERO
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting