

- ☐ Analytical  
☒ Medical  
☐ Testing Machine/Big Range  
☐ Balance

**TECHNICAL SERVICE REPORT**

CUSTOMER NAME: SIMBULAN STO. NINO HOSPITAL ADDRESS: DON CARLOS, BUKIDNON  
SERVICE REQUESTED BY: DEPARTMENT: X-RAY ROOM  
EQUIPMENT/MODEL: SHIMADZU DR SYSTEM - VIEWWORKS SERIAL NO.: VEBBEW004

## SERVICE CATEGORY

☒ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: SCHEDULED PMS - 1ST VISIT

## ACTIONS TAKEN

-TESTED THE UNIT BEFORE PMS, OK  
-ASKED FOR END-USER CONCERNS  
-CHECKED AND CLEANED THE PC AND OTHER PARTS  
-CHECKED DETECTOR CRADLE FUNCTIONALITY, CHARGING OK ON BOTH CHARGING SLOTS  
-CHECKED SCU AND DETECTOR, SELF DIAGNOSIS OK  
-CHECKED NETWORK SETTINGS, OK  
-SET DICOM TAGGING FOR END USER REQUEST OF OPERATOR NAME TO BE SEEN ON PACS  
-CHECKED IMPORT FUNCTIONS, TICK CREATE NEW STUDY WITH IMAGE TIME  
-TESTED IMAGE ACQUISITION AFTER PMS, OK  
-TESTED IMAGE SENDING, OK  
-ENDORSED TO END-USER

## REMARKS/RECOMMENDATIONS

-UNIT IN GOOD WORKING CONDITION

## SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT  
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS  
☐ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

## PARTS SUPPLIED/RECOMMENDED

| ITEM NO. | PART NUMBER | DESCRIPTION | QTY | DELIVERY RECEIPT |
|----------|-------------|-------------|-----|------------------|
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |

| DATE OF SERVICE | TIME STARTED | TIME FINISHED | BILLING STATEMENT |
|-----------------|--------------|---------------|-------------------|
| 05/09/2026      | 9:30AM       | 11:30AM       | SALES INVOICE     |
|                 |              |               | PURCHASE ORDER    |

SERVICED BY:

JOSEPH REY B. DUMASIG

NAME &amp; SIGNATURE

ACKNOWLEDGED BY:

CLARENCE MIE CABALLERO

NAME &amp; SIGNATURE

CONTACT NO:

09525898017

EMAIL ADD:

simbulanradiology@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting