

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: VRP MEDICAL CENTER ADDRESS: MANDALUYONG
SERVICE REQUESTED BY: MA'AM MIMI DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: MOBILE DART EVOLUTION - MX8C SERIAL NO.: MQ00005A5041

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☒ INSTALLATION ☒ OTHERS DELIVERY

COMPLAINT: DELIVERY AND INSTALLATION OF PARTS

ACTIONS TAKEN

-DELIVERED 20PCS. MAIN BATTERY AND WAS RECEIVED BY ENDUSER AND BIOMED
-DISMANTLED DEFECTIVE BATTERY AND REPLACED WITH NEW ONES
-WASTE MATERIALS (DEFECTIVE BATTERIES) WAS ENDORSED TO BIOMED
-CHECKED INDIVIDUAL BATTERY RATING, OK
-CHECKED AND ADJUSTED BATTERY CHARGING DROP VOLTAGE
-CHECKED BATTERY CODE, 6
-RESETTED BATTERY CHARGING USAGE
-TESTED IMAGE ACQUISITION, OK
-UNIT IS NORMALLY CHARGING
-ENDORSED TO END-USER 3-CYCLES OF CHARGING AND DISCHARGING
-UNIT IS IN WORKING CONDITION
-ENDORSED TO END-USER.

RECOMMENDATION:

-FOR REPLACEMENT CABLE COVER

REMARKS/RECOMMENDATIONS

-UNIT IS IN WORKING CONDITION

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☒ INVOICE/BILLING ☒ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT
1	074-71014-11	BATTERY, PXL12090F2B0/Z	20	35643

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT	
MAY 7, 2026	11AM	2PM	SALES INVOICE	
			PURCHASE ORDER	P.O.S.T.S.

SERVICED BY:


ENGR. RENZ ROBIN R. CANO / 0917-847-7457
NAME & SIGNATURE

ACKNOWLEDGED BY:


HERALDA JAY BUENA
NAME & SIGNATURE
09178455839
habuena@mtgracehospitals.com.ph

CONTACT NO:

EMAIL ADD:

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting