



SHIMADZU


Philippines Corporation

19F Marajo Tower, 26th St. cor. 4th Ave., BGC, Taguig City 1634 Metro Manila
(632) 8 869-9563 / (0917) 897-2057 (632) 8 519-9285 spcmanila@shimadzu.com.ph

No. 20260513 01 RRC

- ☐ Analytical
- ☒ Medical
- ☐ Testing Machine/Big Range
- ☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: FAR DIAGNOSTIC CENTER		ADDRESS: NAIC, CAVITE		
SERVICE REQUESTED BY: P.O.S.T.S.		DEPARTMENT: RADIOLOGY		
EQUIPMENT/MODEL: EzyRAD Pro/X's Anesis		SERIAL NO.: MQC8328FC001		
SERVICE CATEGORY				
☐ WARRANTY ☐ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☒ INSTALLATION ☐ OTHERS _____				
COMPLAINT: INSTALLATION OF EzyRAD Pro/X's Anesis				
ACTIONS TAKEN				
INSTALLED THE FOLLOWING SYSTEM		PERFORMED X RAY TUBE AGING		
*MAIN UNIT (EZYRAD PRO) MQC8328FC001		CHECKED AND ADJUSTED FVR CALIBRATION		
*GENERATOR (UD125P CH2 PRO) MFC0B34FC001		TESTED IMAGE ACQUISITION USING CLIENTS DR SYSTEM		
*X RAY TUBE (1.2U161CS 31) RM69598FC003		CONDUCTED OPERATIONAL TRAINING		
*X RAY TUBE SUPPORT (HR 110M) MQCAA95FC001		UNIT IS IN GOOD WORKING CONDITION		
*COLLIMATOR (R 20CA) LZ61EAEFC001		ENDORSED TO END USER		
*CONSOLE (GSC XSY) MFC0B34FC001				
*GRID (J1040 100) MPA54C4G1002				
*HAND SWITCH LM1CF85FC005				
TERMINATE CABLE CONNECTIONS				
MOUNTED GSC CONSOLE				
CHANGE POWER SUPPLY TAPPING				
CHECKED DIPSW AND JUMPER SETTINGS				
ENERGIZED THE SYSTEM				
SETUP VERTICAL CASSETTE HOLDER				
BOLTED X RAY TABLE AND RAILINGS				
REMARKS/RECOMMENDATIONS				
UNIT IS IN GOOD WORKING CONDITION				
SUBMITTED ORIGINAL DOCUMENTS				
☐ STICKER	☐ CALIBRATION CERTIFICATE	☐ PM CERTIFICATE	☐ COMPLETION REPORT	
☐ SYSTEM CHECK	☐ IQ DOCUMENTS	☐ OQ DOCUMENTS	☐ PM DOCUMENTS	
☒ SERVICE REPORT	☐ CALIBRATION DATA (MEDICAL)	☐ INVOICE/BILLING	☒ DELIVERY RECEIPT	
PARTS SUPPLIED/RECOMMENDED				
ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT
DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT	
MAY 11, 2026	8AM	5PM	SALES INVOICE	
MAY 12, 2026	10AM	4PM	PURCHASE ORDER	
MAY 13, 2026	8AM	1PM	P.O.S.T.S.	
SERVICED BY:			ACKNOWLEDGED BY:	
 ENGR. RENZ ROBIN R. CANO / 0917 847 7457 NAME & SIGNATURE			 MA. SUNSHINE B. POBLETE NAME & SIGNATURE	
CONTACT NO:			09562077224	
EMAIL ADD:			fardiagnostic2019@gmail.com	
WHITE: Customer Accounting		PINK: SPC Service Dept	GREEN: Customer End User	YELLOW: SPC Accounting